

CARDIAC RHYTHM MANAGEMENT

Product Performance Report

Important Patient Management Information for Physicians

2025

1st Edition – Issue 92

Medtronic

CRM Product Performance Report

Introduction	3
Method for Estimating CRT, ICD, and IPG Device Performance	7
CRT-D	12
CRT-P	45
ICD	54
IPG	95
Methods for Estimating TPS Performance (Micra)	119
TPS	123
Method for Estimating Lead Performance	127
Pacing Leads	133
Defibrillation Leads	147
Left Heart Leads	161
Epi/Myocardial Leads	171
VDD Single Pass Lead	174
Method for Estimating ICM Performance	175
ICM	177
ICD and CRT-D Charge Time Performance	178
Charge Time	179
Customer Communications	184
Education Briefs	225
Legacy Models	234

2025

1st Edition

Issue 92

Publish Date:

24 May 2025

Cutoff date for this edition is 21 January 2025 for Lead Study data and 01 May 2025 for all other data, unless otherwise stated.

Our Commitment to Quality

Medtronic was founded in 1949 and has grown to become a global leader in medical technology. Seeing what a difference medical technology could make in the lives of patients inspired our founder to develop the Medtronic Mission.

The third tenet of the mission is all about quality:

"To strive without reserve for the greatest possible reliability and quality in our products, to be the unsurpassed standard of comparison, and to be recognized as a company of dedication, honesty, integrity, and service."

Regardless of function, all CRM employees play a role in product quality. Whether designing new therapies, sourcing components, manufacturing products, hiring talented people, assigning financial resources to project teams, or serving in one of the hundreds of other roles, every employee has an influence on product quality.

Product performance information is received from many sources through various channels. Medtronic monitors information from many sources from Research and Development through Manufacturing and Field Performance Vigilance.

When a device is returned to Medtronic, laboratory technicians and engineers assess overall device function. Analysis of returned product is performed according to written procedures. These procedures determine the minimum analysis required. The analysis required varies depending on the type of device, age of the device, the associated information received with the device, actual experience with models of similar design, and other factors. Additional analysis is performed as necessary to investigate a performance concern from a customer, or to collect specific reliability data.

When a malfunction is identified, failure analysis is performed to provide the detailed information necessary to investigate possible causes and actions. Medtronic CRM maintains in-house expertise and performs its failure analysis using facilities it owns and supports. This capability permits detailed failure analysis.

Analysis results are compared to original manufacturing records and design intent. Clinical observations are added to laboratory findings to help determine root cause. Each event is then compared to other events. If a pattern is detected, actions are taken to identify a common root cause, assess patient risk and an appropriate course of action.

Medtronic instituted the industry's first product performance reports in 1983 by publishing data on our chronic lead studies. Pacemakers and other devices followed as our performance reporting has constantly evolved based on customer needs and feedback. One thing has been a constant. It is our sincere commitment to communicate clearly, offering timely and appropriate product performance data and reliability information. This has always been and will continue to be our goal.

Contact Information

We invite our customers to use these telephone numbers to call with suggestions, inquiries, or specific problems related to our products.

US Technical Services Department

tshelp@medtronic.com

Phone:

1 (800) 723-4636 (Tachy)

1 (800) 505-4636 (Brady)

Fax:

1 (800) 824-2362

US Instrumental Technical Services

1 (800) 638-1991

Editorial Staff

Editor

Carrie Schleis, *Vice President, CRM Quality*

International Technical Centers

Europe, the Middle East and Africa (Heerlen NL)

Please contact local Medtronic Representative.

Japan (Tokyo)

Please contact local Medtronic Representative.

Australia-New Zealand

au.crdmtechservices@medtronic.com

For questions related to returning explanted product or returning product that shows signs of malfunction, please contact:

Outside the United States:

Your Medtronic representative or international technical center at the number above.

Within the United States:

Your Medtronic representative or

CRM Returned Product Analysis Laboratory

Phone: 1 (800) 328-2518, ext. 44800

crdm.returnedproduct@medtronic.com

Trademarks of Medtronic

Adapta®	Attesta™	CareLink Express™	LINQ II™	Revo MRI®	SureScan™
AdaptivCRT™	Aurora EV-ICD™	Claria MRI™	Maximo®	Secura®	Syncra®
Advisa DR MRI®	Azure™	Cobalt™	Micra™	Secure®	Transvene
Advisa®	Brava™	Compia MRI™	Mirro MRI™	SelectSecure®	TruRhythm™
Amplia MRI™	CapSure Sense®	Consulta®	MVP®	Sensia®	Versa®
Astra™	CapSure®	Crome™	MyCareLink Heart™	Sensing Assurance	Virtuoso®
Atrial Lead Position Check™	CapSureFix Novus™	EffectivCRT™	Percepta™	Serena™	Visia AF MRI™
Attain Ability®	CapSureFix®	Egida™	Primo MRI™	SmartSync™	Viva™
Attain Performa™	Capture Management®	Encore™	Protecta®	Solara™	Wavelet™
Attain Prevail®	Cardia™	Ensura MRI™	Quick Look™	Sphera™	
Attain Stability™	CareAlert™	Epsilon EV™	Relia™	Sprint Fidelis®	
Attain StarFix®	CareLink™	Evera™	Reveal LINQ™	Sprint Quattro®	

Introduction

For 41 years, Medtronic has monitored performance via both returned product analysis and multicenter clinical studies.

This Product Performance Report (PPR) presents device survival estimates, advisory summaries, performance notes, and other information pertinent to assessing the performance of Medtronic implantable pulse generators (IPGs), implantable cardioverter defibrillators (ICDs), cardiac resynchronization therapy (CRT) devices, insertable cardiac monitors (ICMs), and implantable pacing and defibrillation leads.

This Product Performance Report has been prepared in accordance with International Standard ISO 5841-2:2000(E). As Transcatheter Pacing Systems (TPS) combine the pacing functions of an IPG with the therapy delivery functions of an implantable lead into a single device implanted inside the heart, TPS is subject to complications similar to pacing leads and malfunctions or battery depletion events similar to an implanted pulse generator (IPG). The TPS performance report has been developed to align with these guidelines to the extent possible due to the unique difference between TPS compared to a typical implantable device or lead.

The survival estimates provided in this report are considered to be representative of worldwide performance.

Survival Estimates

Medtronic, like other companies, monitors CRT, ICD, and IPG device performance using returned product analysis. We also monitor CRT, ICD, and IPG device performance using an active multicenter clinical study.

Returned product analysis is a passive approach to assessing product performance. This approach provides a suitable measure of product performance only when a significant number of explanted products are returned to the manufacturer. Returned product analysis provides a measure of hardware performance, but not necessarily the total clinical performance (e.g., the incidence of complications such as infection, erosion, muscle stimulation, etc. are not estimated).

The survival estimates provided in this report for CRT, ICD, and IPG devices are based on returned product analysis. This approach is suitable because a significant number of explanted generators are returned for analysis.

Lead performance is monitored differently. In contrast to CRT, ICD, and IPG devices, a very small percentage of leads are returned to the manufacturer due to the difficulty of explanting them. For leads, an active clinical study provides more accurate survival estimates compared to estimates based solely on returned product analysis.

Survival estimates for leads are based on clinical observations recorded via Medtronic's PAN Registry. This multicenter clinical study is designed to record clinical observations representative of the total clinical experience. Therefore, the lead survival estimates include both lead hardware failure and lead-related medical complications, and do not differentiate a lead hardware failure from other clinical events such as exit block, perforation, dislodgement, or concurrent pulse generator failure.

Transcatheter Pacing Systems are monitored differently. Transcatheter pacing systems (TPS) combine the pacing functions of an IPG with the therapy delivery functions of an implantable lead into a single device implanted inside the heart. To account for the shortfalls of returned product analysis due to a very small percentage of devices being returned, a study of de-identified product data on the Medtronic CareLink™ network is used. The number of devices enrolled and transmitting actively enables a population large enough to give a representative volume of normal battery depletions and to provide insight into the complications that may occur after the device was successfully implanted. TPS survival estimates include both product failures and device-related medical complications and do not differentiate product failures from these complications such as perforation, dislodgement or elevated pacing thresholds.

Introduction continued

ICD Charge Times

Since May 2000, Medtronic has provided important information on charge time performance of ICDs. The information provided in this report shows how ICD charge time can vary during the time a device is implanted. The information is presented in graphical format showing charge time as a function of implant time. The data for charge times are collected from devices enrolled in the PAN Registry.

Customer Communications - Advisory Summaries

This Product Performance Report includes summaries of all Customer Communications classified as Advisories applicable to the performance of the products included in the report. An advisory is added to the report when any product affected by the advisory remains in service and at risk of experiencing the behavior described in the advisory. The advisory will remain in the report until Medtronic estimates no product affected by the advisory remains active, or the risk of experiencing the behavior described in the advisory has passed.

For most advisories, the products subject to the advisory retain essentially the same survival probability as the products of the same model(s) not affected by the advisory. For those advisories where the survival probabilities of the affected and non-affected populations do differ significantly, Medtronic will provide separate survival data for each population. The separate survival data will remain in the report until Medtronic estimates no affected product remains in active service.

Customer Communications - Performance Notes

This report concludes with a number of Customer Communications classified as Non-Advisory Performance Notes developed by Medtronic to provide additional product performance information relevant to follow-up practice and patient management.

Customer Communications - Product Education Briefs

These communications are educational in nature and typically elaborate on information found in the Instructions for Use (IFU) or other approved labeling materials. A product education brief typically serves to clarify information found in labeling that may be misunderstood or misinterpreted by physicians or healthcare professionals. Product education briefs do not provide new patient management guidance, but may be used to reinforce existing recommendations already discussed in the IFU.

How You Can Help

Medtronic urges all physicians to return explanted products and to notify Medtronic when a product is no longer in use, regardless of the reason for explant or removal from use. The procedures for returning products vary by geographic location.

Mailer kits with prepaid US postage are available for use within the United States to send CRTs, ICDs, IPGs, ICMs, and leads to Medtronic's Cardiac Rhythm Management (CRM) Returned Product Analysis Lab. These mailers are sized to accommodate the devices and leads from a single patient or clinical event and are designed to meet US postal regulations for mailing biohazard materials.

If the product being returned is located outside the United States, please contact your local Medtronic representative for instructions.

Medtronic also requests the return of explanted products from non-clinical sources, such as funeral homes, and will assume responsibility for storage and disposal of the product once received.

Introduction continued

Mailer kits can be obtained by contacting the Returned Product Lab. For information on how to contact the Lab, refer to the Contact Information page of this report.

We continually strive to improve this CRM Product Performance Report. In keeping with this philosophy, we ask for your suggestions on the content and format of this report, as well as any information you have regarding the performance of Medtronic products. For information on how to comment on this report, see the Contact Information page.

Overview of Survival Analysis

Medtronic uses the Cutler-Ederer actuarial life table method for devices, standard actuarial method for TPS and Kaplan-Meier for leads to estimate the length of time over which they will perform within performance limits established by Medtronic. This probability to perform within performance limits over time is called the survival probability.

Devices and leads are followed until an event occurs where the device or lead ceases to operate within performance limits. The length of time from implant to the event is recorded for individual devices and leads in the population sample. The population sample for CRT, ICD, and IPG devices is made up of patients whose devices are registered as implanted in the United States. For leads, the population sample is the patients enrolled in our multicenter, international prospective Product Surveillance Registry. For TPS, the population is the de-identified devices on the Medtronic CareLink™ network.

For CRTs, IPGs and ICDs, the events can be normal battery depletion or a device malfunction. For leads, the events are complications as defined in the study protocol. For TPS, the events are complications or malfunctions as defined in the methods for estimating.

The actuarial life table method allows Medtronic to account for devices and leads removed from service for reasons unrelated to performance and for device and leads still in service. Devices and leads removed for reasons unrelated to performance or are still in service are said to be suspended. Examples of devices and leads removed from service for reasons unrelated to performance include:

- Removed to upgrade the device or lead
- No longer in service due to the death of the patient for reasons unrelated to the device or leads
- Implanted in patients who are lost to follow-up

For each suspension, the device or lead has performed within performance limits for a period of time, after which its performance is unknown.

Confidence Intervals

Since survival curves are based on a sample of the device and lead population, they are only estimates of survival. The larger the effective sample size, the more confident the estimate. A confidence interval can be calculated to assess the confidence in an estimate. In the Product Performance Report, Medtronic provides a 95% confidence interval. This can be interpreted as meaning that 95% of the time, the true survival of the device will fall somewhere in the interval.

Survival Curves in the Product Performance Report

Since the survival estimate can become very imprecise with small effective sample sizes, Medtronic truncates the survival curve when the effective sample size is less than 100 for CRTs, ICDs, and IPGs, and when the number entered is less than 50 for leads. The survival charts in the Product Performance Report show the effective sample size for

Introduction continued

each year interval where Medtronic has experience. When the effective sample size reaches 100 for CRTs, ICDs, and IPGs or when the number entered reaches 50 for leads, the next data point is added to the survival curve.

Although the report provides tabular data in one-year intervals, the device curves are actually computed and plotted using the Cutler-Ederer method and 1-month intervals (for CRT, ICD, and IPG devices), the TPS curves are actually computed and plotted using the standard actuarial method and 1-month intervals, and leads curves are computed and plotted using Kaplan-Meier, which uses individual survival times.

A number of references are available for additional information on survival analysis using the Cutler-Ederer life table method¹ and for the Kaplan-Meier method.²

¹Lee, Elisa T.(2003) Statistical Methods for Survival Data Analysis – 3rd Edition (Wiley Series in Probability and Statistics).

²Klein, John P., Moeschberger, Melvin L. Survival Analysis Techniques for Censored and Truncated Data, New York: Springer-Verlag New York, Inc., 1997.

Method for Estimating CRT, ICD, and IPG Device Performance

The performance of CRT, ICD, and IPG devices is expressed in terms of device survival estimates, where “survival” refers to the function of the device, not the survival of the patient. These survival estimates are intended to illustrate the probability that a device will survive for a given number of years without malfunction or battery depletion.

The survival estimates are determined from the analysis of Medtronic Cardiac Rhythm Management (CRM's) United States device registration data and United States returned product analysis data. These data are presented graphically and numerically.

Because this analysis is based on returned product analysis, the performance data does not reflect any device-related medical complications such as erosion, infection, muscle stimulation, or muscle inhibition.

Categorization of Depleted and Malfunctioning Devices for Survival Analysis

For survival estimation, every device analyzed in the CRM Returned Product Analysis laboratory is assigned to one of three categories. The device 1) is functioning normally, 2) has reached normal battery depletion, or 3) has malfunctioned. This categorization is combined with data from our device registry for the total number of implants and the implant durations to create the survival curves presented on the following pages.

Definition of Malfunction

In alignment with industry guidance, Medtronic CRM considers a device as having malfunctioned when the device is removed from service and the analysis shows that any parameter was outside the performance limits established by Medtronic while in service.

Devices damaged after explant, damaged due to failure to heed warnings or contraindications in the labeling, or damaged due to interaction with other implanted devices (including leads) are not considered device malfunctions.

A device subject to a safety advisory is not considered to have malfunctioned unless it has been found, through analysis, to have performed outside the performance limits established by Medtronic.

Not all malfunctions expose the patient to a loss of therapy. Some malfunctions included in the following survival estimates may not have been detected at all by the physician or the patient. All malfunctions, however, are included in the survival estimates and provide important feedback to our product development organization.

To provide insight into the nature of malfunctions, each malfunction is categorized as Malfunction with Compromised Therapy Function or Malfunction without Compromised Therapy Function.

For this report, Normal Battery Depletion, Malfunction with Compromised Therapy Function, and Malfunction without Compromised Therapy Function are defined as follows:

Normal Battery Depletion – The condition when:

- (a) a device is returned with no associated complaint and the device has reached its elective replacement indicator(s) with implant time that meets or exceeds the nominal (50 percentile) predicted longevity at default (labeled) settings

Or

- (b) a device is returned and the device has reached its elective replacement indicator(s) with implant time exceeding 80% of the expected longevity calculated using the available device setting information

Method for Estimating CRT, ICD, and IPG Device Performance continued

Or

- (c) a device is taken out of service without an associated complaint and with evidence the battery reached its elective replacement indicator(s).

Medtronic CRM establishes expected longevity by statistically characterizing the power consumed by the device and the power available from the device battery. This characterization is applied to a number of parameter configurations. The statistical mean value minus three standard deviations is used as the expected longevity for determining if a battery depleted normally. The actual longevity achieved for any device while implanted will depend on the actual programmed parameters and patient factors, and may differ significantly from these estimates.

Malfunction with Compromised Therapy Function

The condition when a device is found to have malfunctioned in a manner that compromised pacing or defibrillation therapy (including complete loss or partial degradation), while implanted and in service, as confirmed by returned product analysis.

Examples: Sudden loss of battery voltage; accelerated current drain such that low battery was not detected before loss of therapy; sudden malfunction during defibrillation therapy resulting in aborted delivery of therapy, intermittent malfunction where therapy is or may be compromised while in the malfunction state.

Malfunction without Compromised Therapy Function

The condition when a device is found to have malfunctioned in a manner that did not compromise pacing or defibrillation therapy, while implanted and in service, as confirmed by returned product analysis.

Examples: Error affecting diagnostic functions, telemetry function, data storage; malfunction of a component that causes battery to lose power quickly enough to cause premature battery depletion, but slowly enough that the condition is detected through normal follow-up before therapy is lost; mechanical problems with connector header that do not affect therapy.

Expanded Malfunction Detail

The malfunctions are further divided into categories that identify the subject area of the malfunction. The malfunctions are divided into the following subject areas:

Battery – Findings linked to the battery and its components

Device-Related Current Pathway – Findings confirmed through device or device memory returns linked to reduced or no energy shocks and other effects (e.g., 50% drop in daily pacing lead impedance measurements not due to a lead issue)

Electrical Component – Findings linked to electrical components such as integrated circuits, resistors, capacitors, diodes, etc.

Electrical Interconnect – Findings linked to the connections between electrical components such as wires, solder joints, wire bonds, etc.

Possible Early Battery Depletion – Findings where the actual reported implant time is less than 80% of the expected longevity calculated using the available device setting information with no device malfunction observed. There may not be sufficient device setting information to determine conclusively if battery depletion was normal or premature in the absence of a specific root cause finding. However, returned devices meeting the above criteria are conservatively classified as Possible Early Battery Depletion malfunctions.

Method for Estimating CRT, ICD, and IPG Device Performance continued

Software/Firmware – Findings linked to software or firmware function

Other – Findings related to other components such as insulators, grommets, setscrews, and packaging, and findings where analysis is inconclusive.

Returned Product Analysis Process

Analysis of returned product or device memory data is performed according to written procedures. These procedures determine the minimum analysis required. The analysis required varies depending on the type of device, age of the device, the associated information received with the device, actual experience with models of similar design, and other factors. Additional analysis is performed as necessary to investigate a performance concern from a customer, or to collect specific reliability data.

When a device is returned with a performance concern from a customer, the general analysis process includes a preliminary analysis of the device in its as-received condition, followed by an automated functional test using test equipment equivalent to the equipment used in manufacturing.

When a malfunction is identified, failure analysis is performed to provide the detailed information necessary to investigate possible causes and actions. Medtronic CRM maintains in-house expertise and performs its failure analysis using facilities it owns and supports. This capability permits detailed failure analysis.

Statistical Methods for Survival Analysis

Of the several different statistical methods available for survival analysis, the Standard Actuarial Method, with suspensions assumed distributed evenly within the intervals (Cutler-Ederer Method), is used to determine survival estimates for CRT, IPG and ICD devices. Implant times are calculated from the implant date to the earlier of the explant date or the cutoff date of the report. From this data an estimate of the probability of device survival is calculated at each monthly interval.

On the following pages, each graph includes a survival curve where events include malfunctions and normal battery depletions (labeled as "Including Normal Battery Depletion"). This survival curve is a good representation of the probability a device will survive a period of time without malfunction and without battery depletion. For example, if a device survival probability is 95% after 5 years of service, then the device has a 5% chance of being removed due to battery depletion or malfunction in the first 5 years following implant.

In addition, a second curve is included to show survival excluding normal battery depletion (labeled as "Excluding Normal Battery Depletion"). This curve is a good representation of the probability for a device to survive without malfunction. This curve includes only malfunctions as events and excludes normal battery depletion.

Since the survival estimate can become very imprecise with small effective sample sizes, Medtronic truncates the survival curve when the effective sample size is less than 100 for CRT, ICD, and IPG devices. The survival charts in the Product Performance Report show the effective sample size for each year interval where we have experience. When the effective sample size reaches 100, the next data point is added to the survival curve.

Although the report provides tabular data in one-year intervals, the curves are actually computed and plotted using one-month intervals.

The data in the tables are rounded to the nearest tenth of one percent. Occasionally, a graph may show 100% survival, but have one or more malfunctions or battery depletions. This occurs because, even with the malfunctions or battery depletions, the data rounds to 100%.

Method for Estimating CRT, ICD, and IPG Device Performance continued

Sample Size and How the Population and Population Samples Are Defined

The population sample from which the survival estimates are derived is comprised of the devices registered as implanted in the United States as of the report cutoff date. The number of registered implants, as well as an estimate of the number that remain in active service, is listed for each model. To be included in the population, the device must have been registered with Medtronic's registration system and implanted for at least one day.

This sample based on US implants is considered to be representative of the worldwide population, and therefore the survival estimates shown in this report should be representative of the performance worldwide of these models.

A CRT, ICD, or IPG model or model family will be included in this report when it has accumulated at least 10,000 implant months and will remain in the report as long as at least 500 devices remain active.

Methods Used to Adjust for Underreporting of Malfunction and Battery Depletion

The tables on the following pages show the actual number of malfunctions and battery depletions recorded by the analysis lab for US registered devices. Since not all devices are returned to Medtronic CRM for analysis, these numbers underestimate the true number of malfunctions and battery depletions. To more accurately estimate the device survival probabilities, the number of malfunctions and battery depletions used to plot each interval of the "Including Normal Battery Depletion" survival curves is adjusted (multiplied) by a factor that is based on an estimate of the magnitude of underreporting. The magnitude of underreporting is estimated by comparing data in Medtronic's Device Registration Tracking Application (DTrak) with data from Returned Product Analysis.

The DTrak system is an important element of Medtronic's Quality System. The DTrak system is designed to meet or exceed the US FDA's device tracking requirements set forth by the Safe Medical Devices Act. In the United States, over 98% of Medtronic's CRT, ICD, and IPG implants become registered in the DTrak system.

Because pacemakers do not cure the patient's underlying health problem, when a pacemaker stops functioning (due to either normal battery replacement or malfunction) it is replaced with a new pacemaker. Therefore, the replacement recorded in the DTrak system is a good indication that the previous pacemaker experienced either battery depletion or malfunction. The fraction of replaced devices that are subsequently returned can be used to estimate the correction factor for the under reporting of the combination of battery depletion and malfunction.

Note that devices of patients who have expired do not factor into the calculation of the correction. It is possible some proportion of these devices experienced battery depletion or malfunction. Since these are not counted into the correction factor based on the return rate of replaced devices, a correction factor based only on the return rate of replaced devices may still underestimate the true rate of battery depletion and malfunction. However, devices that are replaced because the patient is receiving a system upgrade or are removed because the patient no longer needs it (e.g., due to heart transplant) do contribute to the calculation of the correction factor and therefore impart an opposite bias.

Also note that this method of calculating the correction factor cannot distinguish between devices that are removed due to malfunction and those due to normal battery depletion. It might seem intuitive that devices that unexpectedly malfunction should be much more likely to be returned to the manufacturer than a device with ordinary normal battery depletion. But this has not been conclusively demonstrated. Therefore, this method only provides a correction factor reflecting the combination of battery depletion and malfunction.

No adjustment for underreporting is applied to the malfunction-free survival curve because a method for estimating malfunction-only underreporting has not been developed.

Method for Estimating CRT, ICD, and IPG Device Performance continued

Adjustments to Registered Implants to Compensate for Unreported Devices Removed from Service


Devices are at times removed from service for reasons other than device malfunction or battery depletion. Examples are devices removed from service due to non-device related patient mortality and devices removed due to changes in the patient's medical condition. Because an accurate estimate of device survival depends on an accurate estimate of the number of devices in service, it is important not to overstate the number of devices in service.

Medtronic addresses this under reporting to ensure the number of devices in service is not overstated. Regular updates obtained from third party sources such as the Social Security Administration are used to update Medtronic's DTrak data about patients who have died but whose deaths had not been reported to Medtronic. In addition, the patient mortality rate derived from our DTrak system is monitored and compared to published mortality rates for comparable patient populations. If, during calculation of the survival curves, the patient mortality indicated by the data in DTrak is significantly different from published rates, an adjustment is applied to correct the difference. The correction factor is also applied to account for devices that were removed and not reported or returned.

D284TRK

Maximo II CRT-D

US Market Release	17Sep2008	Total Malfunctions (USA)	135
CE Approval Date	14Mar2008	Therapy Function Not Compromised	129
Registered USA Implants	14,990	Electrical Component	5
Estimated Active USA Implants	1,319	Possible Early Battery Depletion	124
Normal Battery Depletions	4,085	Therapy Function Compromised	6
		Electrical Component	6




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 170 mo
Excluding NBD	100.0%	99.7%	99.3%	98.7%	98.6%	98.5%	98.5%	98.5%	98.5%	98.5%	98.5%	98.5%	98.5%	98.5%	98.5%
Including NBD	99.7%	98.2%	93.7%	82.1%	58.2%	32.5%	25.3%	23.3%	22.1%	21.7%	21.3%	21.0%	20.8%	20.4%	20.4%
Effective Sample Size	12496	11082	9495	7252	3989	1659	1091	915	801	745	677	588	420	161	114

D314TRG

Protecta XT CRT-D

US Market Release	25Mar2011	Total Malfunctions (USA)	94
CE Approval Date		Therapy Function Not Compromised	75
Registered USA Implants	41,864	Battery	8
Estimated Active USA Implants	4,627	Electrical Component	40
Normal Battery Depletions	10,525	Possible Early Battery Depletion	25
		Other	2
		Therapy Function Compromised	19
		Battery	11
		Electrical Component	8



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 151 mo
Excluding NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%
Including NBD	99.8%	98.9%	95.1%	86.5%	66.7%	35.3%	23.8%	21.8%	21.1%	20.9%	20.7%	20.4%	20.2%
Effective Sample Size	54150	48922	42272	33503	21052	8905	4901	4058	3643	3406	3096	1984	177

D354TRG

Protecta XT CRT-D

US Market Release

Total Malfunctions (USA)

CE Approval Date

25Mar2010

Therapy Function Not Compromised


Registered USA Implants

1

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 151 mo
Excluding NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%
Including NBD	99.8%	98.9%	95.1%	86.5%	66.7%	35.3%	23.8%	21.8%	21.1%	20.9%	20.7%	20.4%	20.2%
Effective Sample Size	54150	48922	42272	33503	21052	8905	4901	4058	3643	3406	3096	1984	177

D354TRM

Protecta XT CRT-D

US Market Release

Total Malfunctions (USA)

CE Approval Date

15Jul2010


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 151 mo
Excluding NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%
Including NBD	99.8%	98.9%	95.1%	86.5%	66.7%	35.3%	23.8%	21.8%	21.1%	20.9%	20.7%	20.4%	20.2%
Effective Sample Size	54150	48922	42272	33503	21052	8905	4901	4058	3643	3406	3096	1984	177

D364TRG

Protecta CRT-D

US Market Release

Total Malfunctions (USA)

CE Approval Date

25Mar2010


Therapy Function Not Compromised

Registered USA Implants

Estimated Active USA Implants

Therapy Function Compromised

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 151 mo
Excluding NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%
Including NBD	99.8%	98.9%	95.1%	86.5%	66.7%	35.3%	23.8%	21.8%	21.1%	20.9%	20.7%	20.4%	20.2%
Effective Sample Size	54150	48922	42272	33503	21052	8905	4901	4058	3643	3406	3096	1984	177

D364TRM

Protecta CRT-D

US Market Release

Total Malfunctions (USA)

CE Approval Date

15Jul2010


Therapy Function Not Compromised

Registered USA Implants

Estimated Active USA Implants

Therapy Function Compromised

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 151 mo
Excluding NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%
Including NBD	99.8%	98.9%	95.1%	86.5%	66.7%	35.3%	23.8%	21.8%	21.1%	20.9%	20.7%	20.4%	20.2%
Effective Sample Size	54150	48922	42272	33503	21052	8905	4901	4058	3643	3406	3096	1984	177

D394TRG

Egida CRT-D

US Market Release

12Jan2011

Total Malfunctions (USA)


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 151 mo
Excluding NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%
Including NBD	99.8%	98.9%	95.1%	86.5%	66.7%	35.3%	23.8%	21.8%	21.1%	20.9%	20.7%	20.4%	20.2%
Effective Sample Size	54150	48922	42272	33503	21052	8905	4901	4058	3643	3406	3096	1984	177

DTBA1D1

Viva XT

US Market Release

29Jan2013

Total Malfunctions (USA)

72

CE Approval Date

Therapy Function Not Compromised

47

Registered USA Implants

56,948

Battery

10

Estimated Active USA Implants

13,376

Electrical Component

33

Normal Battery Depletions

16,562

Possible Early Battery Depletion

1

Other

3

Therapy Function Compromised

25

Battery


19

Device-Related Current Pathway

2

Electrical Component


4



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 117 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.7%	98.8%	96.4%	91.5%	82.3%	67.7%	49.0%	34.0%	20.4%
Effective Sample Size	86431	78571	71239	63014	53121	41442	28359	14878	4871	182


US Market Release	29Jan2013	Total Malfunctions (USA)	33
CE Approval Date		Therapy Function Not Compromised	24
Registered USA Implants	19,628	Battery	6
Estimated Active USA Implants	4,527	Electrical Component	15
Normal Battery Depletions	6,913	Possible Early Battery Depletion	3
		Therapy Function Compromised	9
		Battery	6
		Electrical Component	3



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 117 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.7%	98.8%	96.4%	91.5%	82.3%	67.7%	49.0%	34.0%	20.4%
Effective Sample Size	86431	78571	71239	63014	53121	41442	28359	14878	4871	182

US Market Release	03Jul2014	Total Malfunctions (USA)	14
CE Approval Date		Therapy Function Not Compromised	9
Registered USA Implants	11,065	Battery	3
Estimated Active USA Implants	2,885	Electrical Component	4
Normal Battery Depletions	3,154	Possible Early Battery Depletion	1
		Other	1
		Therapy Function Compromised	5
		Battery	4
		Electrical Component	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 110 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.8%	99.3%	98.1%	94.4%	85.1%	69.1%	45.8%	22.6%	14.7%
Effective Sample Size	33763	31338	28908	25969	22290	17379	11688	5634	850	198

DTBA1QQ Viva Quad XT

US Market Release	03Jul2014	Total Malfunctions (USA)	49
CE Approval Date		Therapy Function Not Compromised	37
Registered USA Implants	27,415	Battery	12
Estimated Active USA Implants	7,090	Electrical Component	20
Normal Battery Depletions	9,873	Electrical Interconnect	1
		Possible Early Battery Depletion	3
		Other	1
		Therapy Function Compromised	12
		Battery	9
		Electrical Component	3




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 110 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.8%	99.3%	98.1%	94.4%	85.1%	69.1%	45.8%	22.6%	14.7%
Effective Sample Size	33763	31338	28908	25969	22290	17379	11688	5634	850	198

DTBA2D1 Viva XT

US Market Release		Total Malfunctions (USA)	
CE Approval Date	29Aug2016	Therapy Function Not Compromised	
Registered USA Implants	1	Therapy Function Compromised	
Estimated Active USA Implants	1		
Normal Battery Depletions			



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 117 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.7%	98.8%	96.4%	91.5%	82.3%	67.7%	49.0%	34.0%	20.4%
Effective Sample Size	86431	78571	71239	63014	53121	41442	28359	14878	4871	182

DTBA2D4

Viva XT

US Market Release

Total Malfunctions (USA)

CE Approval Date

08Aug2012


Therapy Function Not Compromised

Registered USA Implants

Estimated Active USA Implants

Therapy Function Compromised

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 117 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.7%	98.8%	96.4%	91.5%	82.3%	67.7%	49.0%	34.0%	20.4%
Effective Sample Size	86431	78571	71239	63014	53121	41442	28359	14878	4871	182

DTBA2Q1

Viva Quad XT

US Market Release

Total Malfunctions (USA)

CE Approval Date

12Sep2013


Therapy Function Not Compromised

Registered USA Implants

Estimated Active USA Implants

Therapy Function Compromised

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 110 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.8%	99.3%	98.1%	94.4%	85.1%	69.1%	45.8%	22.6%	14.7%
Effective Sample Size	33763	31338	28908	25969	22290	17379	11688	5634	850	198

US Market Release

08Aug2012

Total Malfunctions (USA)


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 110 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.8%	99.3%	98.1%	94.4%	85.1%	69.1%	45.8%	22.6%	14.7%
Effective Sample Size	33763	31338	28908	25969	22290	17379	11688	5634	850	198

US Market Release

29Jan2013

Total Malfunctions (USA)

25

CE Approval Date

Therapy Function Not Compromised

20

Registered USA Implants

14,104

Battery

9

Estimated Active USA Implants

2,791

Electrical Component

8

Normal Battery Depletions

4,951

Possible Early Battery Depletion

2

Other

1

Therapy Function Compromised


5

Battery

4

Electrical Component

1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 117 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.7%	98.8%	96.4%	91.5%	82.3%	67.7%	49.0%	34.0%	20.4%
Effective Sample Size	86431	78571	71239	63014	53121	41442	28359	14878	4871	182

DTBB1D4 Viva S

US Market Release	29Jan2013	Total Malfunctions (USA)	9
CE Approval Date		Therapy Function Not Compromised	6
Registered USA Implants	4,637	Battery	3
Estimated Active USA Implants	1,054	Electrical Component	2
Normal Battery Depletions	1,647	Other	1
		Therapy Function Compromised	3
		Battery	3




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 117 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.7%	98.8%	96.4%	91.5%	82.3%	67.7%	49.0%	34.0%	20.4%
Effective Sample Size	86431	78571	71239	63014	53121	41442	28359	14878	4871	182

DTBB1Q1 Viva Quad S

US Market Release	03Jul2014	Total Malfunctions (USA)	2
CE Approval Date		Therapy Function Not Compromised	2
Registered USA Implants	2,362	Battery	1
Estimated Active USA Implants	580	Electrical Component	1
Normal Battery Depletions	833	Therapy Function Compromised	0




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 117 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.7%	98.8%	96.4%	91.5%	82.3%	67.7%	49.0%	34.0%	20.4%
Effective Sample Size	86431	78571	71239	63014	53121	41442	28359	14878	4871	182

DTBB1QQ Viva Quad S

US Market Release	03Jul2014	Total Malfunctions (USA)	10
CE Approval Date		Therapy Function Not Compromised	8
Registered USA Implants	5,114	Battery	1
Estimated Active USA Implants	1,274	Electrical Component	4
Normal Battery Depletions	2,142	Possible Early Battery Depletion	2
		Other	1
		Therapy Function Compromised	2
		Electrical Component	2




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 117 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.7%	98.8%	96.4%	91.5%	82.3%	67.7%	49.0%	34.0%	20.4%
Effective Sample Size	86431	78571	71239	63014	53121	41442	28359	14878	4871	182

DTBB2D1 Viva S

US Market Release		Total Malfunctions (USA)	
CE Approval Date	08Aug2012	Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 117 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.7%	98.8%	96.4%	91.5%	82.3%	67.7%	49.0%	34.0%	20.4%
Effective Sample Size	86431	78571	71239	63014	53121	41442	28359	14878	4871	182

US Market Release

Total Malfunctions (USA)

CE Approval Date

08Aug2012


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 117 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.7%	98.8%	96.4%	91.5%	82.3%	67.7%	49.0%	34.0%	20.4%
Effective Sample Size	86431	78571	71239	63014	53121	41442	28359	14878	4871	182

US Market Release

Total Malfunctions (USA)

CE Approval Date

08Aug2012


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 117 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.7%	98.8%	96.4%	91.5%	82.3%	67.7%	49.0%	34.0%	20.4%
Effective Sample Size	86431	78571	71239	63014	53121	41442	28359	14878	4871	182

DTBC2D1 **Brava**

US Market Release
 CE Approval Date 08Aug2012
 Registered USA Implants
 Estimated Active USA Implants
 Normal Battery Depletions

Total Malfunctions (USA)
 Therapy Function Not Compromised
 Therapy Function Compromised




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 117 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.7%	98.8%	96.4%	91.5%	82.3%	67.7%	49.0%	34.0%	20.4%
Effective Sample Size	86431	78571	71239	63014	53121	41442	28359	14878	4871	182

DTBC2D4 **Brava**

US Market Release
 CE Approval Date 08Aug2012
 Registered USA Implants
 Estimated Active USA Implants
 Normal Battery Depletions

Total Malfunctions (USA)
 Therapy Function Not Compromised
 Therapy Function Compromised




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 117 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.7%	98.8%	96.4%	91.5%	82.3%	67.7%	49.0%	34.0%	20.4%
Effective Sample Size	86431	78571	71239	63014	53121	41442	28359	14878	4871	182

DTBC2Q1 Brava Quad

US Market Release
CE Approval Date 12Sep2013
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions (USA)
Therapy Function Not Compromised
Therapy Function Compromised




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 117 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.7%	98.8%	96.4%	91.5%	82.3%	67.7%	49.0%	34.0%	20.4%
Effective Sample Size	86431	78571	71239	63014	53121	41442	28359	14878	4871	182

DTBC2QQ Brava Quad

US Market Release
CE Approval Date 08Aug2012
Registered USA Implants 1
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions (USA)
Therapy Function Not Compromised
Therapy Function Compromised




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 117 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.7%	98.8%	96.4%	91.5%	82.3%	67.7%	49.0%	34.0%	20.4%
Effective Sample Size	86431	78571	71239	63014	53121	41442	28359	14878	4871	182

DTBX1QQ

Viva Quad C

US Market Release	03Jul2014	Total Malfunctions (USA)	1
CE Approval Date		Therapy Function Not Compromised	1
Registered USA Implants	638	Electrical Component	1
Estimated Active USA Implants	72	Therapy Function Compromised	0
Normal Battery Depletions	382		




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 110 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.8%	99.3%	98.1%	94.4%	85.1%	69.1%	45.8%	22.6%	14.7%
Effective Sample Size	33763	31338	28908	25969	22290	17379	11688	5634	850	198

DTBX2QQ

Viva Quad C


US Market Release	03Jul2014	Total Malfunctions (USA)	
CE Approval Date		Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 110 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.8%	99.8%
Including NBD	99.9%	99.8%	99.3%	98.1%	94.4%	85.1%	69.1%	45.8%	22.6%	14.7%
Effective Sample Size	33763	31338	28908	25969	22290	17379	11688	5634	850	198


US Market Release	05Dec2016	Total Malfunctions (USA)	9
CE Approval Date		Therapy Function Not Compromised	7
Registered USA Implants	17,540	Battery	4
Estimated Active USA Implants	12,266	Electrical Component	1
Normal Battery Depletions	909	Electrical Interconnect	1
		Other	1
		Therapy Function Compromised	2
		Battery	1
		Electrical Component	1



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 92 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.8%	98.9%	96.2%	92.1%	85.7%	73.6%	53.6%
Effective Sample Size	43047	38094	30057	22977	16038	10065	3832	387


US Market Release	05Dec2016	Total Malfunctions (USA)	12
CE Approval Date		Therapy Function Not Compromised	6
Registered USA Implants	16,844	Battery	1
Estimated Active USA Implants	12,429	Electrical Component	5
Normal Battery Depletions	711	Therapy Function Compromised	6
		Battery	1
		Device-Related Current Pathway	2
		Electrical Component	2
		Electrical Interconnect	1



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 92 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.8%	98.9%	96.2%	92.1%	85.7%	73.6%	53.6%
Effective Sample Size	43047	38094	30057	22977	16038	10065	3832	387

US Market Release	05Dec2016	Total Malfunctions (USA)	4
CE Approval Date		Therapy Function Not Compromised	4
Registered USA Implants	12,311	Battery	1
Estimated Active USA Implants	9,029	Electrical Interconnect	1
Normal Battery Depletions	512	Possible Early Battery Depletion	1
		Other	1
		Therapy Function Compromised	0




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 96 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.4%	95.7%	89.8%	77.4%	34.7%
Effective Sample Size	115199	104647	88133	72046	53159	32609	14152	580

DTMA1QQ Claria MRI

US Market Release	05Dec2016	Total Malfunctions (USA)	35
CE Approval Date		Therapy Function Not Compromised	23
Registered USA Implants	76,558	Battery	2
Estimated Active USA Implants	58,994	Electrical Component	15
Normal Battery Depletions	2,975	Electrical Interconnect	1
		Possible Early Battery Depletion	1
		Software/Firmware	1
		Other	3
		Therapy Function Compromised	12
		Device-Related Current Pathway	5
		Electrical Component	7




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 96 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.4%	95.7%	89.8%	77.4%	34.7%
Effective Sample Size	115199	104647	88133	72046	53159	32609	14152	580

DTMA2D1 Claria MRI

US Market Release		Total Malfunctions (USA)	
CE Approval Date	29Aug2016	Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 92 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.8%	98.9%	96.2%	92.1%	85.7%	73.6%	53.6%
Effective Sample Size	43047	38094	30057	22977	16038	10065	3832	387

DTMA2D4

Claria MRI

US Market Release

Total Malfunctions (USA)

CE Approval Date

19Feb2016


Therapy Function Not Compromised

Registered USA Implants

Estimated Active USA Implants

Therapy Function Compromised

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 92 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.8%	98.9%	96.2%	92.1%	85.7%	73.6%	53.6%
Effective Sample Size	43047	38094	30057	22977	16038	10065	3832	387

DTMA2Q1

Claria MRI

US Market Release

Total Malfunctions (USA)

CE Approval Date

29Aug2016


Therapy Function Not Compromised

Registered USA Implants

Estimated Active USA Implants

Therapy Function Compromised

Normal Battery Depletions




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 96 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.4%	95.7%	89.8%	77.4%	34.7%
Effective Sample Size	115199	104647	88133	72046	53159	32609	14152	580

DTMA2QQ Claria MRI

US Market Release
CE Approval Date 19Feb2016
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions (USA)
Therapy Function Not Compromised
Therapy Function Compromised




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 96 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.4%	95.7%	89.8%	77.4%	34.7%
Effective Sample Size	115199	104647	88133	72046	53159	32609	14152	580

DTMB1D1 Amplia MRI

US Market Release 05Dec2016
CE Approval Date
Registered USA Implants 6,765
Estimated Active USA Implants 4,139
Normal Battery Depletions 607

Total Malfunctions (USA) 6
Therapy Function Not Compromised 5
 Battery 2
 Electrical Component 2
 Other 1
Therapy Function Compromised 1
 Battery 1




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 92 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.8%	98.9%	96.2%	92.1%	85.7%	73.6%	53.6%
Effective Sample Size	43047	38094	30057	22977	16038	10065	3832	387

DTMB1D4 Amplia MRI

US Market Release	01Feb2016	Total Malfunctions (USA)	4
CE Approval Date		Therapy Function Not Compromised	3
Registered USA Implants	7,007	Battery	1
Estimated Active USA Implants	4,000	Electrical Component	2
Normal Battery Depletions	770	Therapy Function Compromised	1
		Possible Early Battery Depletion	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 92 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.8%	98.9%	96.2%	92.1%	85.7%	73.6%	53.6%
Effective Sample Size	43047	38094	30057	22977	16038	10065	3832	387

DTMB1Q1 Amplia MRI

US Market Release	05Dec2016	Total Malfunctions (USA)	2
CE Approval Date		Therapy Function Not Compromised	1
Registered USA Implants	4,471	Battery	1
Estimated Active USA Implants	2,927	Therapy Function Compromised	1
Normal Battery Depletions	352	Battery	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 96 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.4%	95.7%	89.8%	77.4%	34.7%
Effective Sample Size	115199	104647	88133	72046	53159	32609	14152	580

DTMB1QQ Amplia MRI

US Market Release	01Feb2016	Total Malfunctions (USA)	40
CE Approval Date		Therapy Function Not Compromised	31
Registered USA Implants	31,999	Battery	13
Estimated Active USA Implants	17,724	Electrical Component	12
Normal Battery Depletions	5,024	Possible Early Battery Depletion	3
		Other	3
		Therapy Function Compromised	9
		Battery	8
		Device-Related Current Pathway	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 96 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.4%	95.7%	89.8%	77.4%	34.7%
Effective Sample Size	115199	104647	88133	72046	53159	32609	14152	580

DTMB2D1 Amplia MRI

US Market Release		Total Malfunctions (USA)	
CE Approval Date	29Aug2016	Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 92 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.8%	98.9%	96.2%	92.1%	85.7%	73.6%	53.6%
Effective Sample Size	43047	38094	30057	22977	16038	10065	3832	387

DTMB2D4 Amplia MRI

US Market Release
CE Approval Date 19Feb2016
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions (USA)
Therapy Function Not Compromised
Therapy Function Compromised




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 92 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.8%	98.9%	96.2%	92.1%	85.7%	73.6%	53.6%
Effective Sample Size	43047	38094	30057	22977	16038	10065	3832	387

DTMB2Q1 Amplia MRI

US Market Release
CE Approval Date 29Aug2016
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions (USA)
Therapy Function Not Compromised
Therapy Function Compromised




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 96 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.4%	95.7%	89.8%	77.4%	34.7%
Effective Sample Size	115199	104647	88133	72046	53159	32609	14152	580

DTMB2QQ **Amplia MRI**

US Market Release
CE Approval Date 19Feb2016
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions (USA)
Therapy Function Not Compromised
Therapy Function Compromised




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 96 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.4%	95.7%	89.8%	77.4%	34.7%
Effective Sample Size	115199	104647	88133	72046	53159	32609	14152	580

DTMC1D1 **Compia MRI**

US Market Release 05Dec2016
CE Approval Date
Registered USA Implants 1,142
Estimated Active USA Implants 769
Normal Battery Depletions 88

Total Malfunctions (USA) 1
Therapy Function Not Compromised 0
Therapy Function Compromised 1
Device-Related Current Pathway 1




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 92 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.8%	98.9%	96.2%	92.1%	85.7%	73.6%	53.6%
Effective Sample Size	43047	38094	30057	22977	16038	10065	3832	387

DTMC1QQ Compia MRI

US Market Release	01Feb2016	Total Malfunctions (USA)	5
CE Approval Date		Therapy Function Not Compromised	5
Registered USA Implants	5,017	Battery	3
Estimated Active USA Implants	3,210	Electrical Component	2
Normal Battery Depletions	706	Therapy Function Compromised	0




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 96 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.4%	95.7%	89.8%	77.4%	34.7%
Effective Sample Size	115199	104647	88133	72046	53159	32609	14152	580

DTMC2D1 Compia MRI

US Market Release		Total Malfunctions (USA)	
CE Approval Date	29Aug2016	Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 92 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.8%	98.9%	96.2%	92.1%	85.7%	73.6%	53.6%
Effective Sample Size	43047	38094	30057	22977	16038	10065	3832	387

US Market Release

Total Malfunctions (USA)

CE Approval Date

19Feb2016


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 92 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.8%	98.9%	96.2%	92.1%	85.7%	73.6%	53.6%
Effective Sample Size	43047	38094	30057	22977	16038	10065	3832	387

US Market Release

Total Malfunctions (USA)

CE Approval Date

19Feb2016

Therapy Function Not Compromised


Registered USA Implants

1

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 96 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.4%	95.7%	89.8%	77.4%	34.7%
Effective Sample Size	115199	104647	88133	72046	53159	32609	14152	580

DTPA2D1

Cobalt XT HF

US Market Release	23Apr2020	Total Malfunctions (USA)	1
CE Approval Date	18Dec2019	Therapy Function Not Compromised	1
Registered USA Implants	7,495	Other	1
Estimated Active USA Implants	7,067	Therapy Function Compromised	0
Normal Battery Depletions	22		




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 58 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.6%	97.8%
Effective Sample Size	73512	41429	23006	7620	110

DTPA2D4

Cobalt XT HF

US Market Release	23Apr2020	Total Malfunctions (USA)	4
CE Approval Date	18Dec2019	Therapy Function Not Compromised	2
Registered USA Implants	12,266	Electrical Component	1
Estimated Active USA Implants	11,675	Electrical Interconnect	1
Normal Battery Depletions	20	Therapy Function Compromised	2
		Electrical Component	1
		Electrical Interconnect	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 58 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.6%	97.8%
Effective Sample Size	73512	41429	23006	7620	110

DTPA2Q1

Cobalt XT HF Quad

US Market Release	23Apr2020	Total Malfunctions (USA)	1
CE Approval Date	18Dec2019	Therapy Function Not Compromised	1
Registered USA Implants	5,266	Software/Firmware	1
Estimated Active USA Implants	4,968	Therapy Function Compromised	0
Normal Battery Depletions	4		




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 58 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.6%	97.8%
Effective Sample Size	73512	41429	23006	7620	110

DTPA2QQ

Cobalt XT HF Quad

US Market Release	23Apr2020	Total Malfunctions (USA)	6
CE Approval Date	18Dec2019	Therapy Function Not Compromised	4
Registered USA Implants	50,576	Electrical Component	3
Estimated Active USA Implants	48,379	Software/Firmware	1
Normal Battery Depletions	42	Therapy Function Compromised	2
		Electrical Component	1
		Electrical Interconnect	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 58 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.6%	97.8%
Effective Sample Size	73512	41429	23006	7620	110

DTPB2D1 Cobalt HF

US Market Release	23Apr2020	Total Malfunctions (USA)	3
CE Approval Date	18Dec2019	Therapy Function Not Compromised	1
Registered USA Implants	4,407	Electrical Component	1
Estimated Active USA Implants	3,968	Therapy Function Compromised	2
Normal Battery Depletions	48	Electrical Component	1
		Electrical Interconnect	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 58 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.6%	97.8%
Effective Sample Size	73512	41429	23006	7620	110

DTPB2D4 Cobalt HF

US Market Release	23Apr2020	Total Malfunctions (USA)	6
CE Approval Date	18Dec2019	Therapy Function Not Compromised	5
Registered USA Implants	4,970	Electrical Component	1
Estimated Active USA Implants	4,580	Electrical Interconnect	3
Normal Battery Depletions	28	Software/Firmware	1
		Therapy Function Compromised	1
		Electrical Component	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 58 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.6%	97.8%
Effective Sample Size	73512	41429	23006	7620	110

DTPB2Q1

Cobalt HF Quad

US Market Release	23Apr2020	Total Malfunctions (USA)	
CE Approval Date	18Dec2019	Therapy Function Not Compromised	
Registered USA Implants	2,966		
Estimated Active USA Implants	2,705	Therapy Function Compromised	
Normal Battery Depletions	12		




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 58 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.6%	97.8%
Effective Sample Size	73512	41429	23006	7620	110

DTPB2QQ

Cobalt HF Quad

US Market Release	23Apr2020	Total Malfunctions (USA)	14
CE Approval Date	18Dec2019	Therapy Function Not Compromised	8
Registered USA Implants	23,117	Electrical Component	6
Estimated Active USA Implants	21,535	Electrical Interconnect	1
Normal Battery Depletions	116	Other	1
		Therapy Function Compromised	6
		Electrical Component	3
		Electrical Interconnect	3




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 58 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.6%	97.8%
Effective Sample Size	73512	41429	23006	7620	110

DTPC2D1 Crome HF

US Market Release	23Apr2020	Total Malfunctions (USA)
CE Approval Date	18Dec2019	Therapy Function Not Compromised
Registered USA Implants	508	
Estimated Active USA Implants	467	Therapy Function Compromised
Normal Battery Depletions	4	




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 58 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.6%	97.8%
Effective Sample Size	73512	41429	23006	7620	110

DTPC2D4 Crome HF

US Market Release	23Apr2020	Total Malfunctions (USA)
CE Approval Date	18Dec2019	Therapy Function Not Compromised
Registered USA Implants	601	
Estimated Active USA Implants	562	Therapy Function Compromised
Normal Battery Depletions	4	




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 58 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.6%	97.8%
Effective Sample Size	73512	41429	23006	7620	110

DTPC2Q1

Crome HF Quad

US Market Release	23Apr2020	Total Malfunctions (USA)	
CE Approval Date	18Dec2019	Therapy Function Not Compromised	
Registered USA Implants	250		
Estimated Active USA Implants	235	Therapy Function Compromised	
Normal Battery Depletions	6		




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 58 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.6%	97.8%
Effective Sample Size	73512	41429	23006	7620	110

DTPC2QQ

Crome HF Quad

US Market Release	23Apr2020	Total Malfunctions (USA)	1
CE Approval Date	18Dec2019	Therapy Function Not Compromised	1
Registered USA Implants	2,318	Electrical Component	1
Estimated Active USA Implants	2,184	Therapy Function Compromised	0
Normal Battery Depletions	10		




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 58 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.6%	98.6%	97.8%
Effective Sample Size	73512	41429	23006	7620	110

C2TR01

Syncra CRT-P

US Market Release	22Mar2011	Total Malfunctions (USA)	7
CE Approval Date	11May2010	Therapy Function Not Compromised	6
Registered USA Implants	10,237	Possible Early Battery Depletion	5
Estimated Active USA Implants	2,052	Other	1
Normal Battery Depletions	1,031	Therapy Function Compromised	1
		Possible Early Battery Depletion	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 132 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.8%	99.5%	98.0%	95.5%	91.1%	84.4%	74.1%	60.9%	46.8%	27.5%
Effective Sample Size	26187	23392	20952	18301	15672	13095	10457	7571	4653	2093	100

C3TR01

Consulta CRT-P

US Market Release		Total Malfunctions (USA)	
CE Approval Date	11May2010	Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 132 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.8%	99.5%	98.0%	95.5%	91.1%	84.4%	74.1%	60.9%	46.8%	27.5%
Effective Sample Size	26187	23392	20952	18301	15672	13095	10457	7571	4653	2093	100

C4TR01

Consulta CRT-P

US Market Release	22Mar2011	Total Malfunctions (USA)	8
CE Approval Date		Therapy Function Not Compromised	5
Registered USA Implants	23,406	Possible Early Battery Depletion	5
Estimated Active USA Implants	5,201	Therapy Function Compromised	3
Normal Battery Depletions	2,452	Electrical Component	2
		Possible Early Battery Depletion	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 132 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.8%	99.5%	98.0%	95.5%	91.1%	84.4%	74.1%	60.9%	46.8%	27.5%
Effective Sample Size	26187	23392	20952	18301	15672	13095	10457	7571	4653	2093	100

C5TR01

Viva CRT-P

US Market Release		Total Malfunctions (USA)	
CE Approval Date	04Apr2014	Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 105 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.8%	99.8%
Including NBD	99.9%	99.7%	99.3%	97.5%	95.1%	90.7%	83.8%	71.2%	47.2%
Effective Sample Size	7363	6603	5919	5150	4411	3636	2873	1723	192

C6TR01

Viva CRT-P

US Market Release	09Jul2014	Total Malfunctions (USA)	8
CE Approval Date		Therapy Function Not Compromised	8
Registered USA Implants	9,202	Electrical Component	2
Estimated Active USA Implants	3,360	Possible Early Battery Depletion	6
Normal Battery Depletions	789	Therapy Function Compromised	0




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 105 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%	99.8%	99.8%
Including NBD	99.9%	99.7%	99.3%	97.5%	95.1%	90.7%	83.8%	71.2%	47.2%
Effective Sample Size	7363	6603	5919	5150	4411	3636	2873	1723	192

W1TR01

Percepta CRTP MRI

US Market Release	06May2017	Total Malfunctions (USA)	6
CE Approval Date		Therapy Function Not Compromised	4
Registered USA Implants	19,875	Electrical Component	2
Estimated Active USA Implants	16,559	Possible Early Battery Depletion	1
Normal Battery Depletions	228	Other	1
		Therapy Function Compromised	2
		Electrical Component	2




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 87 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.3%	97.7%	93.8%	89.0%	84.4%	82.5%
Effective Sample Size	21321	16263	12071	8315	4920	2330	458	160

W1TR02

Serena CRTP MRI

US Market Release	06May2017	Total Malfunctions (USA)	3
CE Approval Date		Therapy Function Not Compromised	3
Registered USA Implants	3,449	Electrical Component	2
Estimated Active USA Implants	2,719	Other	1
Normal Battery Depletions	65	Therapy Function Compromised	0




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 87 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.3%	97.7%	93.8%	89.0%	84.4%	82.5%
Effective Sample Size	21321	16263	12071	8315	4920	2330	458	160

W1TR03

Solara CRTP MRI

US Market Release	06May2017	Total Malfunctions (USA)	3
CE Approval Date		Therapy Function Not Compromised	3
Registered USA Implants	4,231	Electrical Component	2
Estimated Active USA Implants	3,124	Possible Early Battery Depletion	1
Normal Battery Depletions	112	Therapy Function Compromised	0



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 87 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.3%	97.7%	93.8%	89.0%	84.4%	82.5%
Effective Sample Size	21321	16263	12071	8315	4920	2330	458	160

W1TR04

Percepta CRTP MRI

US Market Release

Total Malfunctions (USA)

CE Approval Date

10Feb2017


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 87 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.3%	97.7%	93.8%	89.0%	84.4%	82.5%
Effective Sample Size	21321	16263	12071	8315	4920	2330	458	160

W1TR05

Serena CRTP MRI

US Market Release

Total Malfunctions (USA)

CE Approval Date

10Feb2017

Therapy Function Not Compromised


Registered USA Implants

1

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion


Years	1	2	3	4	5	6	7	at 87 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.3%	97.7%	93.8%	89.0%	84.4%	82.5%
Effective Sample Size	21321	16263	12071	8315	4920	2330	458	160

W1TR06

Solara CRTP MRI

US Market Release
 CE Approval Date 10Feb2017
 Registered USA Implants
 Estimated Active USA Implants
 Normal Battery Depletions

Total Malfunctions (USA)
Therapy Function Not Compromised
Therapy Function Compromised



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion


Years	1	2	3	4	5	6	7	at 87 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.3%	97.7%	93.8%	89.0%	84.4%	82.5%
Effective Sample Size	21321	16263	12071	8315	4920	2330	458	160

W4TR01

Percepta Quad CRTP MRI SureScan

US Market Release 06May2017
 CE Approval Date
 Registered USA Implants 66,613
 Estimated Active USA Implants 56,230
 Normal Battery Depletions 444

Total Malfunctions (USA) 14
Therapy Function Not Compromised 13
 Electrical Component 11
 Possible Early Battery Depletion 1
 Other 1
Therapy Function Compromised 1
 Electrical Component 1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 92 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	100.0%	99.8%	99.2%	98.0%	95.9%	92.4%	87.9%
Effective Sample Size	68655	54026	40953	29145	18394	10062	3352	260

W4TR02

Serena Quad CRTP MRI SureScan

US Market Release	06May2017	Total Malfunctions (USA)	3
CE Approval Date		Therapy Function Not Compromised	3
Registered USA Implants	9,124	Electrical Component	3
Estimated Active USA Implants	7,353	Therapy Function Compromised	0
Normal Battery Depletions	67		




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 92 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	100.0%	99.8%	99.2%	98.0%	95.9%	92.4%	87.9%
Effective Sample Size	68655	54026	40953	29145	18394	10062	3352	260

W4TR03

Solara Quad CRTP MRI SureScan

US Market Release	06May2017	Total Malfunctions (USA)	7
CE Approval Date		Therapy Function Not Compromised	4
Registered USA Implants	10,862	Electrical Component	4
Estimated Active USA Implants	8,325	Therapy Function Compromised	3
Normal Battery Depletions	114	Electrical Component	2
		Possible Early Battery Depletion	1



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 92 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	100.0%	99.8%	99.2%	98.0%	95.9%	92.4%	87.9%
Effective Sample Size	68655	54026	40953	29145	18394	10062	3352	260

W4TR04

Percepta Quad CRTP MRI SureScan

US Market Release

Total Malfunctions (USA)

CE Approval Date

10Feb2017


Therapy Function Not Compromised

Registered USA Implants

Estimated Active USA Implants

Therapy Function Compromised

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 92 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	100.0%	99.8%	99.2%	98.0%	95.9%	92.4%	87.9%
Effective Sample Size	68655	54026	40953	29145	18394	10062	3352	260

W4TR05

Serena Quad CRTP MRI SureScan

US Market Release

Total Malfunctions (USA)

CE Approval Date

10Feb2017


Therapy Function Not Compromised

Registered USA Implants

Estimated Active USA Implants

Therapy Function Compromised

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 92 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	100.0%	99.8%	99.2%	98.0%	95.9%	92.4%	87.9%
Effective Sample Size	68655	54026	40953	29145	18394	10062	3352	260

US Market Release

Total Malfunctions (USA)

CE Approval Date

10Feb2017


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 92 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	100.0%	99.8%	99.2%	98.0%	95.9%	92.4%	87.9%
Effective Sample Size	68655	54026	40953	29145	18394	10062	3352	260

D214VRM

Secura VR

US Market Release

Total Malfunctions (USA)

CE Approval Date

17Dec2010


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 176 mo
Excluding NBD	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.7%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%
Including NBD	99.8%	99.6%	99.3%	99.0%	98.6%	98.0%	96.2%	91.1%	77.1%	54.7%	46.7%	44.9%	43.9%	43.6%	43.3%
Effective Sample Size	17638	16328	15175	14070	12958	11844	10615	8588	5571	2906	2092	1656	1235	747	150

D234VRC

Secura VR

US Market Release

Total Malfunctions (USA)

CE Approval Date

14Mar2008


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 176 mo
Excluding NBD	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.7%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%
Including NBD	99.8%	99.6%	99.3%	99.0%	98.6%	98.0%	96.2%	91.1%	77.1%	54.7%	46.7%	44.9%	43.9%	43.6%	43.3%
Effective Sample Size	17638	16328	15175	14070	12958	11844	10615	8588	5571	2906	2092	1656	1235	747	150

D264DRM

Maximo II DR

US Market Release	09Jan2012	Total Malfunctions (USA)
CE Approval Date	22Jul2010	Therapy Function Not Compromised
Registered USA Implants	6	
Estimated Active USA Implants		Therapy Function Compromised
Normal Battery Depletions	2	




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 179 mo
Excluding NBD	100.0%	100.0%	99.9%	99.8%	99.7%	99.7%	99.5%	99.3%	99.3%	99.3%	99.3%	99.2%	99.2%	99.2%	99.2%
Including NBD	99.9%	99.6%	99.3%	98.4%	95.6%	87.7%	71.7%	48.5%	35.1%	32.9%	32.2%	31.9%	31.5%	31.3%	30.9%
Effective Sample Size	17236	15934	14783	13616	12097	9584	5993	2812	1730	1494	1374	1215	979	590	111

D264VRM

Maximo II VR


US Market Release	02May2012	Total Malfunctions (USA)
CE Approval Date	17Dec2010	Therapy Function Not Compromised
Registered USA Implants		
Estimated Active USA Implants		Therapy Function Compromised
Normal Battery Depletions		



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 178 mo
Excluding NBD	100.0%	99.9%	99.9%	99.9%	99.9%	99.8%	99.7%	99.7%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%
Including NBD	99.8%	99.6%	99.4%	99.2%	98.8%	98.0%	95.3%	89.3%	75.3%	54.6%	46.3%	45.2%	44.7%	44.4%	44.4%
Effective Sample Size	10871	10123	9421	8722	8029	7336	6493	5260	3413	1865	1355	1106	851	507	130

US Market Release	17Sep2008	Total Malfunctions (USA)	71
CE Approval Date	14Mar2008	Therapy Function Not Compromised	54
Registered USA Implants	19,956	Battery	7
Estimated Active USA Implants	2,330	Electrical Component	15
Normal Battery Depletions	3,648	Possible Early Battery Depletion	30
		Other	2
		Therapy Function Compromised	17
		Battery	11
		Electrical Component	5
		Possible Early Battery Depletion	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 179 mo
Excluding NBD	100.0%	100.0%	99.9%	99.8%	99.7%	99.7%	99.5%	99.3%	99.3%	99.3%	99.3%	99.2%	99.2%	99.2%	99.2%
Including NBD	99.9%	99.6%	99.3%	98.4%	95.6%	87.7%	71.7%	48.5%	35.1%	32.9%	32.2%	31.9%	31.5%	31.3%	30.9%
Effective Sample Size	17236	15934	14783	13616	12097	9584	5993	2812	1730	1494	1374	1215	979	590	111

D284VRC

Maximo II VR

US Market Release	17Sep2008	Total Malfunctions (USA)	32
CE Approval Date	14Mar2008	Therapy Function Not Compromised	22
Registered USA Implants	12,861	Battery	10
Estimated Active USA Implants	2,057	Electrical Component	5
Normal Battery Depletions	1,624	Possible Early Battery Depletion	4
		Software/Firmware	3
		Therapy Function Compromised	10
		Battery	6
		Electrical Component	3
		Software/Firmware	1



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 178 mo
Excluding NBD	100.0%	99.9%	99.9%	99.9%	99.9%	99.8%	99.7%	99.7%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%	99.6%
Including NBD	99.8%	99.6%	99.4%	99.2%	98.8%	98.0%	95.3%	89.3%	75.3%	54.6%	46.3%	45.2%	44.7%	44.4%	44.4%
Effective Sample Size	10871	10123	9421	8722	8029	7336	6493	5260	3413	1865	1355	1106	851	507	130

D294VRC

Virtuoso II VR


US Market Release		Total Malfunctions (USA)	
CE Approval Date	20Aug2008	Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 171 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.7%	99.7%	99.7%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%
Including NBD	99.7%	99.7%	99.4%	99.2%	98.8%	98.3%	96.8%	92.1%	80.9%	56.8%	47.5%	46.2%	45.3%	44.7%	44.7%
Effective Sample Size	7677	7159	6652	6136	5663	5130	4571	3748	2499	1295	887	714	561	278	127

US Market Release	25Mar2011	Total Malfunctions (USA)	77
CE Approval Date		Therapy Function Not Compromised	39
Registered USA Implants	34,746	Battery	8
Estimated Active USA Implants	4,576	Electrical Component	25
Normal Battery Depletions	4,557	Electrical Interconnect	1
		Possible Early Battery Depletion	4
		Other	1
		Therapy Function Compromised	38
		Battery	30
		Electrical Component	8




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 153 mo
Excluding NBD	100.0%	99.9%	99.9%	99.9%	99.9%	99.8%	99.7%	99.7%	99.6%	99.6%	99.6%	99.6%	99.6%
Including NBD	99.8%	99.7%	99.5%	99.0%	97.2%	91.3%	77.7%	53.0%	39.0%	37.3%	36.9%	36.7%	36.6%
Effective Sample Size	54185	50301	46252	42262	37892	31014	20481	9415	5336	4639	4168	2749	136

D314DRM Protecta XT DR

US Market Release	09Nov2011	Total Malfunctions (USA)	25
CE Approval Date		Therapy Function Not Compromised	17
Registered USA Implants	13,914	Battery	3
Estimated Active USA Implants	2,162	Electrical Component	12
Normal Battery Depletions	1,929	Other	2
		Therapy Function Compromised	8
		Battery	7
		Electrical Component	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 153 mo
Excluding NBD	100.0%	99.9%	99.9%	99.9%	99.9%	99.8%	99.7%	99.7%	99.6%	99.6%	99.6%	99.6%	99.6%
Including NBD	99.8%	99.7%	99.5%	99.0%	97.2%	91.3%	77.7%	53.0%	39.0%	37.3%	36.9%	36.7%	36.6%
Effective Sample Size	54185	50301	46252	42262	37892	31014	20481	9415	5336	4639	4168	2749	136

D354DRG Protecta XT DR

US Market Release		Total Malfunctions (USA)	
CE Approval Date	25Mar2010	Therapy Function Not Compromised	
Registered USA Implants	2	Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions	1		



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 153 mo
Excluding NBD	100.0%	99.9%	99.9%	99.9%	99.9%	99.8%	99.7%	99.7%	99.6%	99.6%	99.6%	99.6%	99.6%
Including NBD	99.8%	99.7%	99.5%	99.0%	97.2%	91.3%	77.7%	53.0%	39.0%	37.3%	36.9%	36.7%	36.6%
Effective Sample Size	54185	50301	46252	42262	37892	31014	20481	9415	5336	4639	4168	2749	136

D354DRM

Protecta XT DR

US Market Release

Total Malfunctions (USA)

CE Approval Date

15Jul2010

Therapy Function Not Compromised


Registered USA Implants

1

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 153 mo
Excluding NBD	100.0%	99.9%	99.9%	99.9%	99.9%	99.8%	99.7%	99.7%	99.6%	99.6%	99.6%	99.6%	99.6%
Including NBD	99.8%	99.7%	99.5%	99.0%	97.2%	91.3%	77.7%	53.0%	39.0%	37.3%	36.9%	36.7%	36.6%
Effective Sample Size	54185	50301	46252	42262	37892	31014	20481	9415	5336	4639	4168	2749	136

D354VRG

Protecta XT VR

US Market Release

Total Malfunctions (USA)

CE Approval Date

25Mar2010


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 151 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.7%	99.5%	99.2%	98.8%	97.4%	92.9%	79.9%	57.1%	49.0%	47.5%	46.5%
Effective Sample Size	25806	23987	22247	20577	19028	17457	15706	13118	8608	4368	3064	1727	111

D354VRM

Protecta XT VR

US Market Release

Total Malfunctions (USA)

CE Approval Date

17Dec2010

Therapy Function Not Compromised


Registered USA Implants

1

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 151 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.7%	99.5%	99.2%	98.8%	97.4%	92.9%	79.9%	57.1%	49.0%	47.5%	46.5%
Effective Sample Size	25806	23987	22247	20577	19028	17457	15706	13118	8608	4368	3064	1727	111

D364DRG

Protecta DR

US Market Release

Total Malfunctions (USA)

CE Approval Date

25Mar2010

Therapy Function Not Compromised


Registered USA Implants

1

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 153 mo
Excluding NBD	100.0%	99.9%	99.9%	99.9%	99.9%	99.8%	99.7%	99.7%	99.6%	99.6%	99.6%	99.6%	99.6%
Including NBD	99.8%	99.7%	99.5%	99.0%	97.2%	91.3%	77.7%	53.0%	39.0%	37.3%	36.9%	36.7%	36.6%
Effective Sample Size	54185	50301	46252	42262	37892	31014	20481	9415	5336	4639	4168	2749	136

D364DRM

Protecta DR

US Market Release

Total Malfunctions (USA)

CE Approval Date

15Jul2010


Therapy Function Not Compromised

Registered USA Implants

Estimated Active USA Implants

Therapy Function Compromised

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 153 mo
Excluding NBD	100.0%	99.9%	99.9%	99.9%	99.9%	99.8%	99.7%	99.7%	99.6%	99.6%	99.6%	99.6%	99.6%
Including NBD	99.8%	99.7%	99.5%	99.0%	97.2%	91.3%	77.7%	53.0%	39.0%	37.3%	36.9%	36.7%	36.6%
Effective Sample Size	54185	50301	46252	42262	37892	31014	20481	9415	5336	4639	4168	2749	136

D364VRG

Protecta VR

US Market Release

Total Malfunctions (USA)

CE Approval Date

25Mar2010


Therapy Function Not Compromised

Registered USA Implants

Estimated Active USA Implants

Therapy Function Compromised

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 151 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.7%	99.5%	99.2%	98.8%	97.4%	92.9%	79.9%	57.1%	49.0%	47.5%	46.5%
Effective Sample Size	25806	23987	22247	20577	19028	17457	15706	13118	8608	4368	3064	1727	111

D364VRM

Protecta VR

US Market Release

Total Malfunctions (USA)

CE Approval Date

17Dec2010


Therapy Function Not Compromised

Registered USA Implants

Estimated Active USA Implants

Therapy Function Compromised

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 151 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.7%	99.5%	99.2%	98.8%	97.4%	92.9%	79.9%	57.1%	49.0%	47.5%	46.5%
Effective Sample Size	25806	23987	22247	20577	19028	17457	15706	13118	8608	4368	3064	1727	111

D384DRG

Cardia DR

US Market Release

Total Malfunctions (USA)

CE Approval Date

12Jan2011


Therapy Function Not Compromised

Registered USA Implants

Estimated Active USA Implants

Therapy Function Compromised

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion


Years	1	2	3	4	5	6	7	8	9	10	11	12	at 153 mo
Excluding NBD	100.0%	99.9%	99.9%	99.9%	99.9%	99.8%	99.7%	99.7%	99.6%	99.6%	99.6%	99.6%	99.6%
Including NBD	99.8%	99.7%	99.5%	99.0%	97.2%	91.3%	77.7%	53.0%	39.0%	37.3%	36.9%	36.7%	36.6%
Effective Sample Size	54185	50301	46252	42262	37892	31014	20481	9415	5336	4639	4168	2749	136

D384VRG

Cardia VR

US Market Release
CE Approval Date 12Jan2011
Registered USA Implants 1
Estimated Active USA Implants
Normal Battery Depletions 1

Total Malfunctions (USA)
Therapy Function Not Compromised
Therapy Function Compromised



■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion


Years	1	2	3	4	5	6	7	8	9	10	11	12	at 151 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.7%	99.5%	99.2%	98.8%	97.4%	92.9%	79.9%	57.1%	49.0%	47.5%	46.5%
Effective Sample Size	25806	23987	22247	20577	19028	17457	15706	13118	8608	4368	3064	1727	111

D394DRG

Egida DR

US Market Release
CE Approval Date 12Jan2011
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions (USA)
Therapy Function Not Compromised
Therapy Function Compromised



■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 153 mo
Excluding NBD	100.0%	99.9%	99.9%	99.9%	99.9%	99.8%	99.7%	99.7%	99.6%	99.6%	99.6%	99.6%	99.6%
Including NBD	99.8%	99.7%	99.5%	99.0%	97.2%	91.3%	77.7%	53.0%	39.0%	37.3%	36.9%	36.7%	36.6%
Effective Sample Size	54185	50301	46252	42262	37892	31014	20481	9415	5336	4639	4168	2749	136

US Market Release

Total Malfunctions (USA)

CE Approval Date

12Jan2011


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants


Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 151 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.7%	99.7%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.7%	99.5%	99.2%	98.8%	97.4%	92.9%	79.9%	57.1%	49.0%	47.5%	46.5%
Effective Sample Size	25806	23987	22247	20577	19028	17457	15706	13118	8608	4368	3064	1727	111

US Market Release	03Apr2013	Total Malfunctions (USA)	89
CE Approval Date		Therapy Function Not Compromised	53
Registered USA Implants	43,052	Battery	34
Estimated Active USA Implants	18,078	Electrical Component	16
Normal Battery Depletions	4,495	Software/Firmware	1
		Other	2
		Therapy Function Compromised	36
		Battery	31
		Device-Related Current Pathway	1
		Electrical Component	2
		Electrical Interconnect	1
		Other	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 130 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.5%	99.1%	97.9%	94.9%	89.6%	76.4%	31.7%
Effective Sample Size	218772	201953	181288	161981	138338	110959	82722	55553	31324	11984	445

DDBB1D4 Evera XT

US Market Release	03Apr2013	Total Malfunctions (USA)	78
CE Approval Date		Therapy Function Not Compromised	48
Registered USA Implants	30,219	Battery	34
Estimated Active USA Implants	11,578	Electrical Component	10
Normal Battery Depletions	4,007	Electrical Interconnect	2
		Possible Early Battery Depletion	1
		Other	1
		Therapy Function Compromised	30
		Battery	23
		Device-Related Current Pathway	4
		Electrical Component	3




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 130 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.5%	99.1%	97.9%	94.9%	89.6%	76.4%	31.7%
Effective Sample Size	218772	201953	181288	161981	138338	110959	82722	55553	31324	11984	445

DDBB2D1 Evera XT

US Market Release		Total Malfunctions (USA)	
CE Approval Date	17Dec2012	Therapy Function Not Compromised	
Registered USA Implants	1	Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 130 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.5%	99.1%	97.9%	94.9%	89.6%	76.4%	31.7%
Effective Sample Size	218772	201953	181288	161981	138338	110959	82722	55553	31324	11984	445

DDBB2D4 Evera XT

US Market Release
CE Approval Date 17Dec2012
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions (USA)
Therapy Function Not Compromised
Therapy Function Compromised



■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion


Years	1	2	3	4	5	6	7	8	9	10	at 130 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.5%	99.1%	97.9%	94.9%	89.6%	76.4%	31.7%
Effective Sample Size	218772	201953	181288	161981	138338	110959	82722	55553	31324	11984	445

DDBC3D1 Evera S

US Market Release 03Apr2013
CE Approval Date 17Dec2012
Registered USA Implants 8,434
Estimated Active USA Implants 3,437
Normal Battery Depletions 1,047

Total Malfunctions (USA) 18
Therapy Function Not Compromised 9
Therapy Function Compromised 9


Battery 7
 Electrical Component 2
 Battery 6
 Device-Related Current Pathway 1
 Electrical Component 2



■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 130 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.5%	99.1%	97.9%	94.9%	89.6%	76.4%	31.7%
Effective Sample Size	218772	201953	181288	161981	138338	110959	82722	55553	31324	11984	445


US Market Release	03Apr2013	Total Malfunctions (USA)	14
CE Approval Date	17Dec2013	Therapy Function Not Compromised	5
Registered USA Implants	6,064	Battery	3
Estimated Active USA Implants	2,366	Electrical Component	2
Normal Battery Depletions	919	Therapy Function Compromised	9
		Battery	5
		Device-Related Current Pathway	2
		Electrical Component	1
		Possible Early Battery Depletion	1



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 130 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.5%	99.1%	97.9%	94.9%	89.6%	76.4%	31.7%
Effective Sample Size	218772	201953	181288	161981	138338	110959	82722	55553	31324	11984	445

US Market Release	12Oct2016	Total Malfunctions (USA)	47
CE Approval Date		Therapy Function Not Compromised	28
Registered USA Implants	37,350	Battery	15
Estimated Active USA Implants	28,367	Electrical Component	11
Normal Battery Depletions	273	Electrical Interconnect	1
		Other	1
		Therapy Function Compromised	19
		Battery	6
		Device-Related Current Pathway	5
		Electrical Component	8




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 130 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.5%	99.1%	97.9%	94.9%	89.6%	76.4%	31.7%
Effective Sample Size	218772	201953	181288	161981	138338	110959	82722	55553	31324	11984	445

DDMB1D4 Evera MRI XT

US Market Release	11Sep2015	Total Malfunctions (USA)	109
CE Approval Date		Therapy Function Not Compromised	65
Registered USA Implants	107,237	Battery	34
Estimated Active USA Implants	83,275	Electrical Component	24
Normal Battery Depletions	1,172	Electrical Interconnect	4
		Other	3
		Therapy Function Compromised	44
		Battery	24
		Device-Related Current Pathway	15
		Electrical Component	5




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 130 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.5%	99.1%	97.9%	94.9%	89.6%	76.4%	31.7%
Effective Sample Size	218772	201953	181288	161981	138338	110959	82722	55553	31324	11984	445

DDMB2D1 Evera MRI XT

US Market Release		Total Malfunctions (USA)	
CE Approval Date	05Sep2016	Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 130 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.5%	99.1%	97.9%	94.9%	89.6%	76.4%	31.7%
Effective Sample Size	218772	201953	181288	161981	138338	110959	82722	55553	31324	11984	445


DDMB2D4

Evera MRI XT

US Market Release
 CE Approval Date 31Mar2014
 Registered USA Implants 1
 Estimated Active USA Implants 1

Total Malfunctions (USA)
 Therapy Function Not Compromised
 Therapy Function Compromised

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion


Years	1	2	3	4	5	6	7	8	9	10	at 130 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.5%	99.1%	97.9%	94.9%	89.6%	76.4%	31.7%
Effective Sample Size	218772	201953	181288	161981	138338	110959	82722	55553	31324	11984	445

DDMC3D1

Evera MRI S

US Market Release 12Oct2016
 CE Approval Date 05Sep2016
 Registered USA Implants 3,379
 Estimated Active USA Implants 2,542
 Normal Battery Depletions 39

Total Malfunctions (USA) 3
 Therapy Function Not Compromised 2
 Battery 1
 Electrical Component 1
 Therapy Function Compromised 1
 Device-Related Current Pathway 1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 130 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.5%	99.1%	97.9%	94.9%	89.6%	76.4%	31.7%
Effective Sample Size	218772	201953	181288	161981	138338	110959	82722	55553	31324	11984	445

DDMC3D4 Evera MRI

US Market Release	11Sep2015	Total Malfunctions (USA)	9
CE Approval Date	31Mar2014	Therapy Function Not Compromised	5
Registered USA Implants	7,184	Battery	4
Estimated Active USA Implants	5,474	Electrical Component	1
Normal Battery Depletions	103	Therapy Function Compromised	4
		Battery	1
		Device-Related Current Pathway	2
		Electrical Component	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 130 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.5%	99.1%	97.9%	94.9%	89.6%	76.4%	31.7%
Effective Sample Size	218772	201953	181288	161981	138338	110959	82722	55553	31324	11984	445

DDMD3D1 Primo

US Market Release	01Mar2018	Total Malfunctions (USA)	1
CE Approval Date	10Nov2017	Therapy Function Not Compromised	1
Registered USA Implants	451	Electrical Component	1
Estimated Active USA Implants	387	Therapy Function Compromised	0
Normal Battery Depletions			




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 130 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.5%	99.1%	97.9%	94.9%	89.6%	76.4%	31.7%
Effective Sample Size	218772	201953	181288	161981	138338	110959	82722	55553	31324	11984	445

DDMD3D4 Primo

US Market Release	01Mar2018	Total Malfunctions (USA)
CE Approval Date	10Nov2017	Therapy Function Not Compromised
Registered USA Implants	1,487	
Estimated Active USA Implants	1,318	Therapy Function Compromised
Normal Battery Depletions	1	




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 130 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.5%	99.1%	97.9%	94.9%	89.6%	76.4%	31.7%
Effective Sample Size	218772	201953	181288	161981	138338	110959	82722	55553	31324	11984	445

DDME3D1 Mirro

US Market Release	01Mar2018	Total Malfunctions (USA)
CE Approval Date	10Nov2017	Therapy Function Not Compromised
Registered USA Implants		
Estimated Active USA Implants		Therapy Function Compromised
Normal Battery Depletions		




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 130 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.5%	99.1%	97.9%	94.9%	89.6%	76.4%	31.7%
Effective Sample Size	218772	201953	181288	161981	138338	110959	82722	55553	31324	11984	445

DDME3D4 Mirro

US Market Release 01Mar2018 **Total Malfunctions (USA)**
CE Approval Date 10Nov2017 **Therapy Function Not Compromised**
Registered USA Implants
Estimated Active USA Implants **Therapy Function Compromised**
Normal Battery Depletions




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 130 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.7%	99.7%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.5%	99.1%	97.9%	94.9%	89.6%	76.4%	31.7%
Effective Sample Size	218772	201953	181288	161981	138338	110959	82722	55553	31324	11984	445

DDPA2D1 Cobalt XT

US Market Release 23Apr2020 **Total Malfunctions (USA)** 1
CE Approval Date 18Dec2019 **Therapy Function Not Compromised** 1
Registered USA Implants 6,442 **Electrical Component** 1
Estimated Active USA Implants 6,175 **Therapy Function Compromised** 0
Normal Battery Depletions




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 58 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%
Effective Sample Size	43366	24376	13933	4812	133

DDPA2D4 Cobalt XT

US Market Release	23Apr2020	Total Malfunctions (USA)	3
CE Approval Date	18Dec2019	Therapy Function Not Compromised	2
Registered USA Implants	42,996	Electrical Component	2
Estimated Active USA Implants	40,989	Therapy Function Compromised	1
Normal Battery Depletions	10	Electrical Interconnect	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 58 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%
Effective Sample Size	43366	24376	13933	4812	133

DDPB3D1 Cobalt

US Market Release	23Apr2020	Total Malfunctions (USA)	2
CE Approval Date	18Dec2019	Therapy Function Not Compromised	1
Registered USA Implants	2,907	Battery	1
Estimated Active USA Implants	2,691	Therapy Function Compromised	1
Normal Battery Depletions		Device-Related Current Pathway	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 58 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%
Effective Sample Size	43366	24376	13933	4812	133

DDPB3D4 Cobalt

US Market Release	23Apr2020	Total Malfunctions (USA)	8
CE Approval Date	18Dec2019	Therapy Function Not Compromised	5
Registered USA Implants	16,133	Battery	1
Estimated Active USA Implants	14,908	Electrical Component	2
Normal Battery Depletions	8	Other	2
		Therapy Function Compromised	3
		Electrical Component	1
		Electrical Interconnect	2




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 58 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%
Effective Sample Size	43366	24376	13933	4812	133

DDPC3D1 Crome

US Market Release	23Apr2020	Total Malfunctions (USA)	
CE Approval Date	18Dec2019	Therapy Function Not Compromised	
Registered USA Implants	393		
Estimated Active USA Implants	364	Therapy Function Compromised	
Normal Battery Depletions			




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 58 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%
Effective Sample Size	43366	24376	13933	4812	133

DDPC3D4 Crome

US Market Release	23Apr2020	Total Malfunctions (USA)	
CE Approval Date	18Dec2019	Therapy Function Not Compromised	
Registered USA Implants	1,804		
Estimated Active USA Implants	1,698	Therapy Function Compromised	

Normal Battery Depletions




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 58 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%
Effective Sample Size	43366	24376	13933	4812	133

DVAB1D1 Visia AF

US Market Release	19Jan2016	Total Malfunctions (USA)	14
CE Approval Date		Therapy Function Not Compromised	10
Registered USA Implants	3,051	Battery	10
Estimated Active USA Implants	2,032	Therapy Function Compromised	4
Normal Battery Depletions	15	Battery	2
		Device-Related Current Pathway	2




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 102 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.6%	99.6%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.5%	99.3%	99.0%	98.7%
Effective Sample Size	76469	70628	61654	52808	42238	29813	17365	5421	218

DVAB1D4 Visia AF

US Market Release	19Jan2016	Total Malfunctions (USA)	5
CE Approval Date		Therapy Function Not Compromised	2
Registered USA Implants	2,045	Battery	2
Estimated Active USA Implants	1,416	Therapy Function Compromised	3
Normal Battery Depletions		Battery	2
		Device-Related Current Pathway	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 102 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.6%	99.6%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.5%	99.3%	99.0%	98.7%
Effective Sample Size	76469	70628	61654	52808	42238	29813	17365	5421	218

DVAB2D1 Visia AF XT

US Market Release		Total Malfunctions (USA)	
CE Approval Date	19Oct2015	Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 102 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.6%	99.6%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.5%	99.3%	99.0%	98.7%
Effective Sample Size	76469	70628	61654	52808	42238	29813	17365	5421	218

DVAC3D1 Visia AF S

US Market Release 19Jan2016 **Total Malfunctions (USA)**
CE Approval Date 19Oct2015 **Therapy Function Not Compromised**
Registered USA Implants
Estimated Active USA Implants **Therapy Function Compromised**
Normal Battery Depletions




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 102 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.6%	99.6%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.5%	99.3%	99.0%	98.7%
Effective Sample Size	76469	70628	61654	52808	42238	29813	17365	5421	218

DVBB1D1 Evera XT

US Market Release 03Apr2013 **Total Malfunctions (USA)** 76
CE Approval Date **Therapy Function Not Compromised** 52
Registered USA Implants 16,113 Battery 45
Estimated Active USA Implants 8,573 Electrical Component 7
Normal Battery Depletions 128 **Therapy Function Compromised** 24
 Battery 16
 Device-Related Current Pathway 4
 Electrical Component 4




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 135 mo
Excluding NBD	100.0%	100.0%	99.9%	99.8%	99.8%	99.7%	99.6%	99.5%	99.4%	99.3%	99.3%	99.3%
Including NBD	100.0%	99.9%	99.7%	99.6%	99.4%	99.1%	98.9%	98.6%	98.3%	97.7%	96.4%	89.4%
Effective Sample Size	52434	48863	45476	42393	39297	36127	33276	30753	24872	12566	2292	175

DVBB1D4 Evera XT

US Market Release	03Apr2013	Total Malfunctions (USA)	99
CE Approval Date		Therapy Function Not Compromised	66
Registered USA Implants	21,952	Battery	51
Estimated Active USA Implants	12,682	Electrical Component	9
Normal Battery Depletions	202	Possible Early Battery Depletion	2
		Other	4
		Therapy Function Compromised	33
		Battery	26
		Device-Related Current Pathway	6
		Electrical Component	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 135 mo
Excluding NBD	100.0%	100.0%	99.9%	99.8%	99.8%	99.7%	99.6%	99.5%	99.4%	99.3%	99.3%	99.3%
Including NBD	100.0%	99.9%	99.7%	99.6%	99.4%	99.1%	98.9%	98.6%	98.3%	97.7%	96.4%	89.4%
Effective Sample Size	52434	48863	45476	42393	39297	36127	33276	30753	24872	12566	2292	175

DVBB2D1 Evera XT

US Market Release		Total Malfunctions (USA)	
CE Approval Date	17Dec2012	Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion


Years	1	2	3	4	5	6	7	8	9	10	11	at 135 mo
Excluding NBD	100.0%	100.0%	99.9%	99.8%	99.8%	99.7%	99.6%	99.5%	99.4%	99.3%	99.3%	99.3%
Including NBD	100.0%	99.9%	99.7%	99.6%	99.4%	99.1%	98.9%	98.6%	98.3%	97.7%	96.4%	89.4%
Effective Sample Size	52434	48863	45476	42393	39297	36127	33276	30753	24872	12566	2292	175

DVBB2D4 Evera XT

US Market Release
 CE Approval Date
 Registered USA Implants
 Estimated Active USA Implants
 Normal Battery Depletions

17Dec2012

Total Malfunctions (USA)
 Therapy Function Not Compromised
 Therapy Function Compromised



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 135 mo
Excluding NBD	100.0%	100.0%	99.9%	99.8%	99.8%	99.7%	99.6%	99.5%	99.4%	99.3%	99.3%	99.3%
Including NBD	100.0%	99.9%	99.7%	99.6%	99.4%	99.1%	98.9%	98.6%	98.3%	97.7%	96.4%	89.4%
Effective Sample Size	52434	48863	45476	42393	39297	36127	33276	30753	24872	12566	2292	175


DVBC3D1 Evera S

US Market Release
 CE Approval Date
 Registered USA Implants
 Estimated Active USA Implants
 Normal Battery Depletions

03Apr2013

Total Malfunctions (USA)
 Therapy Function Not Compromised
 Battery
 Electrical Component
 Therapy Function Compromised
 Battery
 Electrical Component

28
 19
 17
 2
 9
 8
 1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 135 mo
Excluding NBD	100.0%	100.0%	99.9%	99.8%	99.8%	99.7%	99.6%	99.5%	99.4%	99.3%	99.3%	99.3%
Including NBD	100.0%	99.9%	99.7%	99.6%	99.4%	99.1%	98.9%	98.6%	98.3%	97.7%	96.4%	89.4%
Effective Sample Size	52434	48863	45476	42393	39297	36127	33276	30753	24872	12566	2292	175

DVBC3D4 Evera S

US Market Release	03Apr2013	Total Malfunctions (USA)	24
CE Approval Date	17Dec2012	Therapy Function Not Compromised	17
Registered USA Implants	5,623	Battery	14
Estimated Active USA Implants	3,371	Electrical Component	3
Normal Battery Depletions	40	Therapy Function Compromised	7
		Battery	5
		Device-Related Current Pathway	2




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 135 mo
Excluding NBD	100.0%	100.0%	99.9%	99.8%	99.8%	99.7%	99.6%	99.5%	99.4%	99.3%	99.3%	99.3%
Including NBD	100.0%	99.9%	99.7%	99.6%	99.4%	99.1%	98.9%	98.6%	98.3%	97.7%	96.4%	89.4%
Effective Sample Size	52434	48863	45476	42393	39297	36127	33276	30753	24872	12566	2292	175

DVEA3E4 Aurora EV-ICD


US Market Release	20Oct2023	Total Malfunctions (USA)	
CE Approval Date	17Feb2023	Therapy Function Not Compromised	
Registered USA Implants	1,661	Therapy Function Compromised	
Estimated Active USA Implants	1,614		
Normal Battery Depletions			



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	at 15 mo
Excluding NBD	100.0%	100.0%
Including NBD	100.0%	100.0%
Effective Sample Size	286	114

US Market Release	12Oct2016	Total Malfunctions (USA)	24
CE Approval Date		Therapy Function Not Compromised	16
Registered USA Implants	17,778	Battery	10
Estimated Active USA Implants	14,260	Electrical Component	5
Normal Battery Depletions	16	Other	1
		Therapy Function Compromised	8
		Battery	2
		Device-Related Current Pathway	3
		Electrical Component	3




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 102 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.6%	99.6%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.5%	99.3%	99.0%	98.7%
Effective Sample Size	76469	70628	61654	52808	42238	29813	17365	5421	218

DVFB1D4

Visia MRI AF

US Market Release	19Jan2016	Total Malfunctions (USA)	85
CE Approval Date		Therapy Function Not Compromised	52
Registered USA Implants	57,274	Battery	41
Estimated Active USA Implants	45,719	Device-Related Current Pathway	1
Normal Battery Depletions	43	Electrical Component	9
		Other	1
		Therapy Function Compromised	33
		Battery	16
		Device-Related Current Pathway	14
		Electrical Component	3




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 102 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.6%	99.6%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.5%	99.3%	99.0%	98.7%
Effective Sample Size	76469	70628	61654	52808	42238	29813	17365	5421	218

DVFB2D1

Visia MRI AF XT

US Market Release		Total Malfunctions (USA)	
CE Approval Date	05Sep2016	Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 102 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.6%	99.6%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.5%	99.3%	99.0%	98.7%
Effective Sample Size	76469	70628	61654	52808	42238	29813	17365	5421	218


DVFB2D4

Visia MRI AF XT

US Market Release
 CE Approval Date 19Oct2015
 Registered USA Implants 2
 Estimated Active USA Implants 1

Total Malfunctions (USA)
 Therapy Function Not Compromised
 Therapy Function Compromised

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 102 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.6%	99.6%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.5%	99.3%	99.0%	98.7%
Effective Sample Size	76469	70628	61654	52808	42238	29813	17365	5421	218


DVFC3D1

Visia MRI AF S

US Market Release 12Oct2016
 CE Approval Date 05Sep2016
 Registered USA Implants 1,477
 Estimated Active USA Implants 1,208

Total Malfunctions (USA) 1
 Therapy Function Not Compromised 1
 Battery 1
 Therapy Function Compromised 0

Normal Battery Depletions




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 102 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.6%	99.6%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.5%	99.3%	99.0%	98.7%
Effective Sample Size	76469	70628	61654	52808	42238	29813	17365	5421	218

DVFC3D4

Visia MRI AF S

US Market Release	19Jan2016	Total Malfunctions (USA)	4
CE Approval Date	19Oct2015	Therapy Function Not Compromised	4
Registered USA Implants	3,592	Battery	4
Estimated Active USA Implants	2,981	Therapy Function Compromised	0
Normal Battery Depletions	7		




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 102 mo
Excluding NBD	100.0%	100.0%	100.0%	99.9%	99.9%	99.8%	99.7%	99.6%	99.6%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.5%	99.3%	99.0%	98.7%
Effective Sample Size	76469	70628	61654	52808	42238	29813	17365	5421	218

DVMB1D4

Evera MRI XT

US Market Release	11Sep2015	Total Malfunctions (USA)	38
CE Approval Date		Therapy Function Not Compromised	19
Registered USA Implants	10,270	Battery	15
Estimated Active USA Implants	6,682	Electrical Component	3
Normal Battery Depletions	26	Other	1
		Therapy Function Compromised	19
		Battery	15
		Device-Related Current Pathway	4



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 135 mo
Excluding NBD	100.0%	100.0%	99.9%	99.8%	99.8%	99.7%	99.6%	99.5%	99.4%	99.3%	99.3%	99.3%
Including NBD	100.0%	99.9%	99.7%	99.6%	99.4%	99.1%	98.9%	98.6%	98.3%	97.7%	96.4%	89.4%
Effective Sample Size	52434	48863	45476	42393	39297	36127	33276	30753	24872	12566	2292	175

DVMB2D1

Evera MRI XT

US Market Release

Total Malfunctions (USA)

CE Approval Date

05Sep2016


Therapy Function Not Compromised

Registered USA Implants

Estimated Active USA Implants

Therapy Function Compromised

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 135 mo
Excluding NBD	100.0%	100.0%	99.9%	99.8%	99.8%	99.7%	99.6%	99.5%	99.4%	99.3%	99.3%	99.3%
Including NBD	100.0%	99.9%	99.7%	99.6%	99.4%	99.1%	98.9%	98.6%	98.3%	97.7%	96.4%	89.4%
Effective Sample Size	52434	48863	45476	42393	39297	36127	33276	30753	24872	12566	2292	175

DVMB2D4

Evera MRI XT

US Market Release

Total Malfunctions (USA)

CE Approval Date

31Mar2014

Therapy Function Not Compromised


Registered USA Implants

2

Estimated Active USA Implants

Therapy Function Compromised

Normal Battery Depletions




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 135 mo
Excluding NBD	100.0%	100.0%	99.9%	99.8%	99.8%	99.7%	99.6%	99.5%	99.4%	99.3%	99.3%	99.3%
Including NBD	100.0%	99.9%	99.7%	99.6%	99.4%	99.1%	98.9%	98.6%	98.3%	97.7%	96.4%	89.4%
Effective Sample Size	52434	48863	45476	42393	39297	36127	33276	30753	24872	12566	2292	175

DVMC3D1 Evera MRI S

US Market Release 12Oct2016 **Total Malfunctions (USA)**
CE Approval Date 05Sep2016 **Therapy Function Not Compromised**
Registered USA Implants
Estimated Active USA Implants **Therapy Function Compromised**
Normal Battery Depletions




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 135 mo
Excluding NBD	100.0%	100.0%	99.9%	99.8%	99.8%	99.7%	99.6%	99.5%	99.4%	99.3%	99.3%	99.3%
Including NBD	100.0%	99.9%	99.7%	99.6%	99.4%	99.1%	98.9%	98.6%	98.3%	97.7%	96.4%	89.4%
Effective Sample Size	52434	48863	45476	42393	39297	36127	33276	30753	24872	12566	2292	175

DVMC3D4 Evera MRI S

US Market Release 11Sep2015 **Total Malfunctions (USA)**
CE Approval Date 31Mar2014 **Therapy Function Not Compromised**
Registered USA Implants
Estimated Active USA Implants **Therapy Function Compromised**
Normal Battery Depletions




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 135 mo
Excluding NBD	100.0%	100.0%	99.9%	99.8%	99.8%	99.7%	99.6%	99.5%	99.4%	99.3%	99.3%	99.3%
Including NBD	100.0%	99.9%	99.7%	99.6%	99.4%	99.1%	98.9%	98.6%	98.3%	97.7%	96.4%	89.4%
Effective Sample Size	52434	48863	45476	42393	39297	36127	33276	30753	24872	12566	2292	175

DVMD3D1 Primo

US Market Release 01Mar2018 **Total Malfunctions (USA)**
CE Approval Date 10Nov2017 **Therapy Function Not Compromised**
Registered USA Implants 274
Estimated Active USA Implants 238 **Therapy Function Compromised**

Normal Battery Depletions




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 135 mo
Excluding NBD	100.0%	100.0%	99.9%	99.8%	99.8%	99.7%	99.6%	99.5%	99.4%	99.3%	99.3%	99.3%
Including NBD	100.0%	99.9%	99.7%	99.6%	99.4%	99.1%	98.9%	98.6%	98.3%	97.7%	96.4%	89.4%
Effective Sample Size	52434	48863	45476	42393	39297	36127	33276	30753	24872	12566	2292	175

DVMD3D4 Primo

US Market Release 01Mar2018 **Total Malfunctions (USA)**
CE Approval Date 10Nov2017 **Therapy Function Not Compromised**
Registered USA Implants 627
Estimated Active USA Implants 556 **Therapy Function Compromised**

Normal Battery Depletions




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 135 mo
Excluding NBD	100.0%	100.0%	99.9%	99.8%	99.8%	99.7%	99.6%	99.5%	99.4%	99.3%	99.3%	99.3%
Including NBD	100.0%	99.9%	99.7%	99.6%	99.4%	99.1%	98.9%	98.6%	98.3%	97.7%	96.4%	89.4%
Effective Sample Size	52434	48863	45476	42393	39297	36127	33276	30753	24872	12566	2292	175

DVME3D1 Mirro

US Market Release 01Mar2018 **Total Malfunctions (USA)**
CE Approval Date 10Nov2017 **Therapy Function Not Compromised**
Registered USA Implants
Estimated Active USA Implants **Therapy Function Compromised**
Normal Battery Depletions




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 135 mo
Excluding NBD	100.0%	100.0%	99.9%	99.8%	99.8%	99.7%	99.6%	99.5%	99.4%	99.3%	99.3%	99.3%
Including NBD	100.0%	99.9%	99.7%	99.6%	99.4%	99.1%	98.9%	98.6%	98.3%	97.7%	96.4%	89.4%
Effective Sample Size	52434	48863	45476	42393	39297	36127	33276	30753	24872	12566	2292	175

DVME3D4 Mirro

US Market Release 01Mar2018 **Total Malfunctions (USA)**
CE Approval Date 10Nov2017 **Therapy Function Not Compromised**
Registered USA Implants
Estimated Active USA Implants **Therapy Function Compromised**
Normal Battery Depletions




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 135 mo
Excluding NBD	100.0%	100.0%	99.9%	99.8%	99.8%	99.7%	99.6%	99.5%	99.4%	99.3%	99.3%	99.3%
Including NBD	100.0%	99.9%	99.7%	99.6%	99.4%	99.1%	98.9%	98.6%	98.3%	97.7%	96.4%	89.4%
Effective Sample Size	52434	48863	45476	42393	39297	36127	33276	30753	24872	12566	2292	175

DVPA2D1 Cobalt XT

US Market Release	23Apr2020	Total Malfunctions (USA)
CE Approval Date	18Dec2019	Therapy Function Not Compromised
Registered USA Implants	1,941	
Estimated Active USA Implants	1,811	Therapy Function Compromised
Normal Battery Depletions	2	




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 57 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.9%	99.9%	99.9%
Effective Sample Size	18850	11908	7341	2529	164

DVPA2D4 Cobalt XT

US Market Release	23Apr2020	Total Malfunctions (USA)	2
CE Approval Date	18Dec2019	Therapy Function Not Compromised	1
Registered USA Implants	16,071	Electrical Interconnect	1
Estimated Active USA Implants	15,230	Therapy Function Compromised	1
Normal Battery Depletions		Device-Related Current Pathway	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 57 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.9%	99.9%	99.9%
Effective Sample Size	18850	11908	7341	2529	164

DVPB3D1 Cobalt

US Market Release	23Apr2020	Total Malfunctions (USA)	2
CE Approval Date	18Dec2019	Therapy Function Not Compromised	0
Registered USA Implants	1,725		
Estimated Active USA Implants	1,559	Therapy Function Compromised	2
Normal Battery Depletions		Electrical Interconnect	2




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 57 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.9%	99.9%	99.9%
Effective Sample Size	18850	11908	7341	2529	164

DVPB3D4 Cobalt

US Market Release	23Apr2020	Total Malfunctions (USA)	4
CE Approval Date	18Dec2019	Therapy Function Not Compromised	1
Registered USA Implants	6,786	Other	1
Estimated Active USA Implants	6,275	Therapy Function Compromised	3
Normal Battery Depletions	2	Device-Related Current Pathway	2
		Electrical Interconnect	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 57 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.9%	99.9%	99.9%
Effective Sample Size	18850	11908	7341	2529	164

DVPC3D1 Crome

US Market Release 23Apr2020 **Total Malfunctions (USA)**
CE Approval Date 18Dec2019 **Therapy Function Not Compromised**
Registered USA Implants 152
Estimated Active USA Implants 138 **Therapy Function Compromised**

Normal Battery Depletions




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 57 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.9%	99.9%	99.9%
Effective Sample Size	18850	11908	7341	2529	164

DVPC3D4 Crome

US Market Release 23Apr2020 **Total Malfunctions (USA)**
CE Approval Date 18Dec2019 **Therapy Function Not Compromised**
Registered USA Implants 781
Estimated Active USA Implants 729 **Therapy Function Compromised**

Normal Battery Depletions




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	at 57 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%
Including NBD	100.0%	99.9%	99.9%	99.9%	99.9%
Effective Sample Size	18850	11908	7341	2529	164

A2DR01

Advisa DR MRI

US Market Release	15Jan2013	Total Malfunctions (USA)	85
CE Approval Date		Therapy Function Not Compromised	80
Registered USA Implants	344,436	Battery	1
Estimated Active USA Implants	194,547	Electrical Component	40
Normal Battery Depletions	13,905	Electrical Interconnect	4
		Possible Early Battery Depletion	26
		Software/Firmware	6
		Other	3
		Therapy Function Compromised	5
		Electrical Component	5




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 121 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.4%	98.8%	95.9%	87.9%	53.7%	23.3%
Effective Sample Size	308287	290448	273721	257107	237694	216804	195953	133012	62188	4911	1140

A3DR01

Advisa DR MRI

US Market Release		Total Malfunctions (USA)	
CE Approval Date	02Jun2009	Therapy Function Not Compromised	
Registered USA Implants	23	Therapy Function Compromised	
Estimated Active USA Implants	3		
Normal Battery Depletions	4		




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 121 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.4%	98.8%	95.9%	87.9%	53.7%	23.3%
Effective Sample Size	308287	290448	273721	257107	237694	216804	195953	133012	62188	4911	1140

A3SR01

Advisa SR MRI

US Market Release	19Mar2015	Total Malfunctions (USA)	9
CE Approval Date	24Apr2014	Therapy Function Not Compromised	8
Registered USA Implants	28,082	Electrical Component	3
Estimated Active USA Implants	15,602	Electrical Interconnect	1
Normal Battery Depletions	62	Possible Early Battery Depletion	2
		Other	2
		Therapy Function Compromised	1
		Electrical Component	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 103 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.6%	99.5%	99.2%	99.1%	98.3%
Effective Sample Size	22016	19374	17194	15015	12894	10979	9231	3967	514

A5DR01

Advisa DR

US Market Release		Total Malfunctions (USA)	
CE Approval Date	02Jun2009	Therapy Function Not Compromised	
Registered USA Implants		Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 121 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.4%	98.8%	95.9%	87.9%	53.7%	23.3%
Effective Sample Size	308287	290448	273721	257107	237694	216804	195953	133012	62188	4911	1140

ADD01 Adapta D

US Market Release 17Jul2006 **Total Malfunctions (USA)**
CE Approval Date 20Sep2005 **Therapy Function Not Compromised**
Registered USA Implants 1
Estimated Active USA Implants **Therapy Function Compromised**
Normal Battery Depletions




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 171 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.8%	99.6%	99.3%	98.8%	97.9%	95.0%	87.7%	74.9%	61.3%	46.1%	32.9%	23.1%	8.9%
Effective Sample Size	393174	365286	338807	313172	289378	265139	241047	211890	171070	125128	82659	45004	18460	2963	128

ADDR01 Adapta DR

US Market Release 17Jul2006 **Total Malfunctions (USA)** 95
CE Approval Date 20Sep2005 **Therapy Function Not Compromised** 67
Registered USA Implants 454,887 Electrical Component 59
Estimated Active USA Implants 113,692 Electrical Interconnect 1
Normal Battery Depletions 53,749 Possible Early Battery Depletion 6
 Other 1
Therapy Function Compromised 28
 Electrical Component 23
 Electrical Interconnect 3
 Other 2




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 171 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.8%	99.6%	99.3%	98.8%	97.9%	95.0%	87.7%	74.9%	61.3%	46.1%	32.9%	23.1%	8.9%
Effective Sample Size	393174	365286	338807	313172	289378	265139	241047	211890	171070	125128	82659	45004	18460	2963	128

ADDR03 Adapta DR

US Market Release	17Jul2006	Total Malfunctions (USA)	2
CE Approval Date	20Sep2005	Therapy Function Not Compromised	1
Registered USA Implants	4,570	Electrical Component	1
Estimated Active USA Implants	1,239	Therapy Function Compromised	1
Normal Battery Depletions	651	Electrical Component	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 171 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.8%	99.6%	99.3%	98.8%	97.9%	95.0%	87.7%	74.9%	61.3%	46.1%	32.9%	23.1%	8.9%
Effective Sample Size	393174	365286	338807	313172	289378	265139	241047	211890	171070	125128	82659	45004	18460	2963	128

ADDR06 Adapta DR


US Market Release	17Jul2006	Total Malfunctions (USA)	1
CE Approval Date	20Sep2005	Therapy Function Not Compromised	1
Registered USA Implants	3,664	Electrical Component	1
Estimated Active USA Implants	841	Therapy Function Compromised	0
Normal Battery Depletions	438		



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 171 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.8%	99.6%	99.3%	98.8%	97.9%	95.0%	87.7%	74.9%	61.3%	46.1%	32.9%	23.1%	8.9%
Effective Sample Size	393174	365286	338807	313172	289378	265139	241047	211890	171070	125128	82659	45004	18460	2963	128

US Market Release	17Jul2006	Total Malfunctions (USA)	25
CE Approval Date	20Sep2005	Therapy Function Not Compromised	17
Registered USA Implants	138,613	Electrical Component	13
Estimated Active USA Implants	62,478	Electrical Interconnect	1
Normal Battery Depletions	2,911	Possible Early Battery Depletion	2
		Software/Firmware	1
		Therapy Function Compromised	8
		Electrical Component	5
		Electrical Interconnect	1
		Other	2




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 170 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.5%	99.2%	99.0%	98.5%	97.3%	94.4%	90.0%	87.8%	86.7%	85.6%
Effective Sample Size	119657	112672	106027	99505	92193	84465	77100	69472	59955	49001	36333	22983	10999	1770	559

ADDRS1 Adapta S DR

US Market Release	17Jul2006	Total Malfunctions (USA)	15
CE Approval Date	20Sep2005	Therapy Function Not Compromised	9
Registered USA Implants	49,315	Electrical Component	5
Estimated Active USA Implants	9,525	Possible Early Battery Depletion	3
Normal Battery Depletions	6,727	Other	1
		Therapy Function Compromised	6
		Electrical Component	4
		Other	2




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	at 164 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.8%	99.6%	99.2%	98.5%	96.9%	92.0%	80.2%	63.4%	47.3%	37.6%	35.1%	34.0%	32.9%	32.6%
Effective Sample Size	40118	36091	32379	29052	25765	21765	16542	11062	6745	4127	2821	1672	725	138

ADSR01 Adapta SR

US Market Release	17Jul2006	Total Malfunctions (USA)	18
CE Approval Date	20Sep2005	Therapy Function Not Compromised	12
Registered USA Implants	91,664	Electrical Component	7
Estimated Active USA Implants	18,481	Electrical Interconnect	1
Normal Battery Depletions	6,630	Possible Early Battery Depletion	4
		Therapy Function Compromised	6
		Electrical Component	5
		Electrical Interconnect	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 138 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.7%	99.4%	99.0%	98.3%	97.1%	94.6%	88.7%	72.1%	52.4%	33.1%	20.6%
Effective Sample Size	72043	62959	55132	48118	41283	34939	29008	22603	14951	7630	2024	332

ADSR03

Adapta SR

US Market Release	17Jul2006	Total Malfunctions (USA)	
CE Approval Date	20Sep2005	Therapy Function Not Compromised	
Registered USA Implants	2,129		
Estimated Active USA Implants	414	Therapy Function Compromised	
Normal Battery Depletions	210		




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 138 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.7%	99.4%	99.0%	98.3%	97.1%	94.6%	88.7%	72.1%	52.4%	33.1%	20.6%
Effective Sample Size	72043	62959	55132	48118	41283	34939	29008	22603	14951	7630	2024	332

ADSR06

Adapta SR

US Market Release	17Jul2006	Total Malfunctions (USA)	2
CE Approval Date	20Sep2005	Therapy Function Not Compromised	2
Registered USA Implants	2,919	Electrical Component	2
Estimated Active USA Implants	591	Therapy Function Compromised	0
Normal Battery Depletions	283		




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 138 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.7%	99.4%	99.0%	98.3%	97.1%	94.6%	88.7%	72.1%	52.4%	33.1%	20.6%
Effective Sample Size	72043	62959	55132	48118	41283	34939	29008	22603	14951	7630	2024	332

ADVDD01 Adapta VDD

US Market Release	17Jul2006	Total Malfunctions (USA)
CE Approval Date	20Sep2005	Therapy Function Not Compromised
Registered USA Implants	858	
Estimated Active USA Implants	214	Therapy Function Compromised
Normal Battery Depletions	95	




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 117 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	100.0%	100.0%	100.0%	97.0%	86.1%	67.3%	57.4%	52.4%
Effective Sample Size	706	650	590	536	474	412	324	192	131	101

ATDR01 Attest DR MRI

US Market Release	03Aug2017	Total Malfunctions (USA)
CE Approval Date	16Jun2017	Therapy Function Not Compromised
Registered USA Implants	2,221	
Estimated Active USA Implants	2,112	Therapy Function Compromised
Normal Battery Depletions		



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion


Years	1	2	3	at 48 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	100.0%	100.0%
Effective Sample Size	1882	1319	757	124

ATDRL1

Attestation L DR MRI

US Market Release	03Aug2017	Total Malfunctions (USA)
CE Approval Date	16Jun2017	Therapy Function Not Compromised
Registered USA Implants	320	
Estimated Active USA Implants	303	Therapy Function Compromised

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion


Years	1	2	at 35 mo
Excluding NBD	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	100.0%
Effective Sample Size	271	188	107

ATDRS1

Attestation S DR MRI

US Market Release	03Aug2017	Total Malfunctions (USA)
CE Approval Date	16Jun2017	Therapy Function Not Compromised
Registered USA Implants	1,518	
Estimated Active USA Implants	1,376	Therapy Function Compromised

Normal Battery Depletions




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	at 47 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	100.0%	100.0%
Effective Sample Size	1150	791	405	100

ATSR01

Attest SR MRI

US Market Release 03Aug2017 **Total Malfunctions (USA)**
CE Approval Date 16Jun2017 **Therapy Function Not Compromised**
Registered USA Implants 1,303
Estimated Active USA Implants 971 **Therapy Function Compromised**
Normal Battery Depletions 1




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	at 43 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%
Including NBD	99.7%	99.7%	99.7%	99.7%
Effective Sample Size	810	520	262	106

EN1DR01

Ensura MRI

US Market Release **Total Malfunctions (USA)**
CE Approval Date 23Jun2010 **Therapy Function Not Compromised**
Registered USA Implants 6
Estimated Active USA Implants 2 **Therapy Function Compromised**
Normal Battery Depletions



■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	at 121 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.4%	98.8%	95.9%	87.9%	53.7%	23.3%
Effective Sample Size	308287	290448	273721	257107	237694	216804	195953	133012	62188	4911	1140

EN1SR01

Ensura SR MRI

US Market Release

Total Malfunctions (USA)

CE Approval Date

24Apr2014


Therapy Function Not Compromised

Registered USA Implants

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	at 103 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.8%	99.7%	99.6%	99.5%	99.2%	99.1%	98.3%
Effective Sample Size	22016	19374	17194	15015	12894	10979	9231	3967	514

RED01

Relia D

US Market Release

Total Malfunctions (USA)

CE Approval Date

07May2008

Therapy Function Not Compromised

Registered USA Implants


2

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions

1



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 171 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.8%	99.6%	99.3%	98.8%	97.9%	95.0%	87.7%	74.9%	61.3%	46.1%	32.9%	23.1%	8.9%
Effective Sample Size	393174	365286	338807	313172	289378	265139	241047	211890	171070	125128	82659	45004	18460	2963	128

REDR01

Relia DR

US Market Release

Total Malfunctions (USA)

CE Approval Date

07May2008

Therapy Function Not Compromised

Registered USA Implants

11


Estimated Active USA Implants

2

Therapy Function Compromised

Normal Battery Depletions

2



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 171 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.8%	99.6%	99.3%	98.8%	97.9%	95.0%	87.7%	74.9%	61.3%	46.1%	32.9%	23.1%	8.9%
Effective Sample Size	393174	365286	338807	313172	289378	265139	241047	211890	171070	125128	82659	45004	18460	2963	128

RES01

Relia S

US Market Release

Total Malfunctions (USA)

CE Approval Date

07May2008

Therapy Function Not Compromised

Registered USA Implants


4

Estimated Active USA Implants

2

Therapy Function Compromised

Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 138 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.7%	99.4%	99.0%	98.3%	97.1%	94.6%	88.7%	72.1%	52.4%	33.1%	20.6%
Effective Sample Size	72043	62959	55132	48118	41283	34939	29008	22603	14951	7630	2024	332

RESR01

Relia SR

US Market Release

Total Malfunctions (USA)

CE Approval Date

07May2008

Therapy Function Not Compromised

Registered USA Implants


7

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions

1



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 138 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.7%	99.4%	99.0%	98.3%	97.1%	94.6%	88.7%	72.1%	52.4%	33.1%	20.6%
Effective Sample Size	72043	62959	55132	48118	41283	34939	29008	22603	14951	7630	2024	332

REVDD01

Relia VDD

US Market Release

Total Malfunctions (USA)

CE Approval Date

07May2008

Therapy Function Not Compromised


Registered USA Implants

1

Therapy Function Compromised

Estimated Active USA Implants

Normal Battery Depletions




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	at 117 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	100.0%	100.0%	100.0%	97.0%	86.1%	67.3%	57.4%	52.4%
Effective Sample Size	706	650	590	536	474	412	324	192	131	101

RVDR01

Revo MRI SureScan

US Market Release	08Feb2011	Total Malfunctions (USA)	111
CE Approval Date		Therapy Function Not Compromised	108
Registered USA Implants	69,116	Battery	1
Estimated Active USA Implants	13,901	Electrical Component	40
Normal Battery Depletions	12,077	Electrical Interconnect	1
		Possible Early Battery Depletion	61
		Software/Firmware	4
		Other	1
		Therapy Function Compromised	3
		Electrical Component	3




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	at 138 mo
Excluding NBD	100.0%	100.0%	99.9%	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.7%	99.7%	99.7%
Including NBD	100.0%	99.9%	99.8%	99.6%	99.4%	98.8%	95.8%	88.8%	74.1%	49.8%	19.6%	5.3%
Effective Sample Size	59299	56147	53134	49973	46280	42265	37438	31211	22538	11765	2933	376

SED01

Sensia D

US Market Release	17Jul2006	Total Malfunctions (USA)	
CE Approval Date	20Sep2005	Therapy Function Not Compromised	
Registered USA Implants	5	Therapy Function Compromised	
Estimated Active USA Implants	1		
Normal Battery Depletions	1		




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	at 194 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.8%	99.6%	99.2%	98.6%	97.3%	93.3%	84.6%	72.9%	60.3%	49.8%	43.9%	42.0%	41.4%	41.0%	40.3%
Effective Sample Size	120511	108984	98337	88703	79973	72181	64960	56748	46430	34770	24009	15562	9993	6338	3235	593	169

SEDR01 Sensia DR

US Market Release	17Jul2006	Total Malfunctions (USA)	33
CE Approval Date	20Sep2005	Therapy Function Not Compromised	17
Registered USA Implants	149,410	Electrical Component	15
Estimated Active USA Implants	28,126	Electrical Interconnect	1
Normal Battery Depletions	17,253	Other	1
		Therapy Function Compromised	16
		Electrical Component	6
		Electrical Interconnect	3
		Possible Early Battery Depletion	1
		Other	6




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	at 194 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.8%	99.6%	99.2%	98.6%	97.3%	93.3%	84.6%	72.9%	60.3%	49.8%	43.9%	42.0%	41.4%	41.0%	40.3%
Effective Sample Size	120511	108984	98337	88703	79973	72181	64960	56748	46430	34770	24009	15562	9993	6338	3235	593	169

SEDRL1 Sensia L DR

US Market Release	17Jul2006	Total Malfunctions (USA)	
CE Approval Date	20Sep2005	Therapy Function Not Compromised	
Registered USA Implants	5		
Estimated Active USA Implants	1	Therapy Function Compromised	
Normal Battery Depletions			




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 170 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	99.9%	99.9%	99.8%	99.7%	99.5%	99.2%	99.0%	98.5%	97.3%	94.4%	90.0%	87.8%	86.7%	85.6%
Effective Sample Size	119657	112672	106027	99505	92193	84465	77100	69472	59955	49001	36333	22983	10999	1770	559

SES01 Sensia S

US Market Release	17Jul2006	Total Malfunctions (USA)	
CE Approval Date	20Sep2005	Therapy Function Not Compromised	
Registered USA Implants	4		
Estimated Active USA Implants	1	Therapy Function Compromised	

Normal Battery Depletions




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 154 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.8%	99.6%	99.2%	98.6%	97.8%	95.9%	91.2%	78.4%	60.4%	45.8%	38.8%	33.7%
Effective Sample Size	85817	74447	64543	56004	48247	41088	34580	27806	20110	12154	5903	2349	261

SESR01 Sensia SR

US Market Release	17Jul2006	Total Malfunctions (USA)	17
CE Approval Date	20Sep2005	Therapy Function Not Compromised	13
Registered USA Implants	117,372	Electrical Component	7
Estimated Active USA Implants	20,954	Possible Early Battery Depletion	4
Normal Battery Depletions	9,122	Other	2
		Therapy Function Compromised	4
		Electrical Component	3
		Electrical Interconnect	1



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	at 154 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.8%	99.6%	99.2%	98.6%	97.8%	95.9%	91.2%	78.4%	60.4%	45.8%	38.8%	33.7%
Effective Sample Size	85817	74447	64543	56004	48247	41088	34580	27806	20110	12154	5903	2349	261

SPDR01

Sphera DR MRI

US Market Release 03Aug2017 **Total Malfunctions (USA)**
CE Approval Date 16Jun2017 **Therapy Function Not Compromised**
Registered USA Implants
Estimated Active USA Implants **Therapy Function Compromised**
Normal Battery Depletions




Years	
Excluding NBD	
Including NBD	
Effective Sample Size	

SPDR1

Sphera L DR MRI

US Market Release 03Aug2017 **Total Malfunctions (USA)**
CE Approval Date 16Jun2017 **Therapy Function Not Compromised**
Registered USA Implants 1 **Therapy Function Compromised**
Estimated Active USA Implants **Therapy Function Compromised**
Normal Battery Depletions



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	at 35 mo
Excluding NBD	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	100.0%
Effective Sample Size	271	188	107

SPSR01

Sphera SR MRI

US Market Release 03Aug2017 **Total Malfunctions (USA)**
CE Approval Date 16Jun2017 **Therapy Function Not Compromised**
Registered USA Implants 1
Estimated Active USA Implants 1 **Therapy Function Compromised**
Normal Battery Depletions




Years	
Excluding NBD	
Including NBD	
Effective Sample Size	

VEDR01

Versa DR


US Market Release 17Jul2006 **Total Malfunctions (USA)** 28
CE Approval Date 20Sep2005 **Therapy Function Not Compromised** 13
Registered USA Implants 118,957 Electrical Component 9
Estimated Active USA Implants 23,370 Electrical Interconnect 2
Normal Battery Depletions 15,041 Possible Early Battery Depletion 2
Therapy Function Compromised 15
 Electrical Component 11
 Other 4



■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	at 189 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	99.9%	99.9%	99.8%	99.6%	99.2%	98.6%	97.2%	92.6%	81.6%	67.6%	54.3%	43.6%	37.6%	34.9%	34.0%	33.6%
Effective Sample Size	98637	90146	82056	74674	67967	62034	55730	47177	35507	25260	16918	10624	6616	3749	1570	181

US Market Release	16Aug2017	Total Malfunctions (USA)	162
CE Approval Date	02Mar2017	Therapy Function Not Compromised	148
Registered USA Implants	805,165	Battery	4
Estimated Active USA Implants	724,496	Electrical Component	88
Normal Battery Depletions	868	Possible Early Battery Depletion	5
		Software/Firmware	28
		Other	23
		Therapy Function Compromised	14
		Battery	2
		Electrical Component	12




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 88 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	99.9%	99.8%	99.6%	99.2%	98.7%	98.7%
Effective Sample Size	693189	534521	395621	275289	171478	85267	13708	736

W1SR01 Azure XT SR

US Market Release	16Aug2017	Total Malfunctions (USA)	11
CE Approval Date	02Mar2017	Therapy Function Not Compromised	10
Registered USA Implants	62,539	Battery	1
Estimated Active USA Implants	51,956	Electrical Component	6
Normal Battery Depletions	36	Software/Firmware	1
		Other	2
		Therapy Function Compromised	1
		Electrical Component	1




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 86 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	99.9%	99.9%	99.8%	99.5%	99.5%	99.5%
Effective Sample Size	58407	45103	33630	23582	14118	6559	810	165

W2DR01 Azure XT DR

US Market Release		Total Malfunctions (USA)	
CE Approval Date	02Mar2017	Therapy Function Not Compromised	
Registered USA Implants	3	Therapy Function Compromised	
Estimated Active USA Implants	2		
Normal Battery Depletions			



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion


Years	1	2	3	4	5	6	7	at 88 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	99.9%	99.8%	99.6%	99.2%	98.7%	98.7%
Effective Sample Size	693189	534521	395621	275289	171478	85267	13708	736

W2SR01

Azure XT SR

US Market Release
CE Approval Date 02Mar2017
Registered USA Implants
Estimated Active USA Implants
Normal Battery Depletions

Total Malfunctions (USA)
Therapy Function Not Compromised
Therapy Function Compromised



■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion


Years	1	2	3	4	5	6	7	at 86 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	99.9%	99.9%	99.8%	99.5%	99.5%	99.5%
Effective Sample Size	58407	45103	33630	23582	14118	6559	810	165

W3DR01

Azure S DR

US Market Release 16Aug2017
CE Approval Date 02Mar2017
Registered USA Implants 66,683
Estimated Active USA Implants 58,407
Normal Battery Depletions 173

Total Malfunctions (USA) 15
Therapy Function Not Compromised 14
 Electrical Component 10
 Possible Early Battery Depletion 2
 Software/Firmware 2
Therapy Function Compromised 1
 Electrical Component 1




■ Including Normal Battery Depletion
 ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 88 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	99.9%	99.8%	99.6%	99.2%	98.7%	98.7%
Effective Sample Size	693189	534521	395621	275289	171478	85267	13708	736

W3SR01

Azure S SR

US Market Release	16Aug2017	Total Malfunctions (USA)	1
CE Approval Date	02Mar2017	Therapy Function Not Compromised	1
Registered USA Implants	13,707	Electrical Component	1
Estimated Active USA Implants	11,243	Therapy Function Compromised	0
Normal Battery Depletions	6		




■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	7	at 86 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	99.9%	99.9%	99.8%	99.5%	99.5%	99.5%
Effective Sample Size	58407	45103	33630	23582	14118	6559	810	165

X2DR01

Astra XT DR MRI SureScan

US Market Release		Total Malfunctions (USA)	
CE Approval Date	02Mar2017	Therapy Function Not Compromised	
Registered USA Implants	1	Therapy Function Compromised	
Estimated Active USA Implants			
Normal Battery Depletions			



Years	1	2	3	4	5	6	7	at 86 mo
Excluding NBD	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%	100.0%
Including NBD	100.0%	100.0%	99.9%	99.9%	99.8%	99.5%	99.5%	99.5%
Effective Sample Size	58407	45103	33630	23582	14118	6559	810	165

X2SR01

Astra XT SR MRI SureScan

US Market Release

Total Malfunctions (USA)

CE Approval Date

02Mar2017


Therapy Function Not Compromised

Registered USA Implants

Estimated Active USA Implants

Therapy Function Compromised

Normal Battery Depletions



Years	Excluding NBD	Including NBD	Effective Sample Size
0			
1			
2			
3			
4			
5			
6			
7			
8			
9			
10			

X3DR01

Astra S DR

US Market Release

Total Malfunctions (USA)

CE Approval Date

02Mar2017


Therapy Function Not Compromised

Registered USA Implants

Estimated Active USA Implants

Therapy Function Compromised

Normal Battery Depletions



Years	Excluding NBD	Including NBD	Effective Sample Size
0			
1			
2			
3			
4			
5			
6			
7			
8			
9			
10			

US Market Release

Total Malfunctions (USA)

CE Approval Date

02Mar2017


Therapy Function Not Compromised

Registered USA Implants

Estimated Active USA Implants

Therapy Function Compromised

Normal Battery Depletions



Years	
Excluding NBD	
Including NBD	
Effective Sample Size	

Method for Estimating Transcatheter Pacing Performance

Micra TPS Performance Analysis

Transcatheter pacing systems (TPS) combine the pacing functions of an IPG with the therapy delivery functions of an implantable lead into a single device implanted inside the heart. Therefore, TPS is subject to complications similar to pacing leads (e.g. cardiac perforation) and malfunctions or battery depletion events similar to an implanted pulse generator (IPG). Although both transvenous systems and Micra IPG experience similar system level major complications, Micra has been shown to reduce the likelihood of major complications at a system level in post-approval registry data.

The performance report information is determined from the analysis of Medtronic Cardiac Rhythm Management (CRM) United States registration, complaint and CareLink™ network data. A TPS model or model family will be included in this report when it has accumulated at least 10,000 implant months and will remain in the report as long as at least 500 devices remain active in the CareLink™ population.

Shortfalls of using returned products to Estimate Micra TPS Performance

Micra TPS devices returned to Medtronic are analyzed to determine whether or not they meet performance limits established by Medtronic. Although returned product analyses are valuable for gaining insight into failure mechanisms, this data cannot be used by itself for determining the survival probability because only a small fraction of Micra devices are explanted and returned to Medtronic for analysis. Some devices are programmed off due to an adverse event, however, they are often not retrieved/ explanted. The devices that are retrieved and returned cannot be assumed to be statistically representative of the performance of the total population for a given model. For this same reason, devices that meet their expected longevity are also not expected to be returned to Medtronic CRM.

The CareLink™ Network

To account for the shortfalls of returned product analysis, a study of de-identified product data on the Medtronic CareLink™ network is used. The number of devices enrolled and transmitting actively enables a population large enough to give a representative volume of normal battery depletions and to provide insight into the complications that may occur after the device was successfully implanted. As the intent of the product performance report is to provide visibility to long-term device performance, the devices reviewed from the CareLink™ Network have been implanted for at least 30 days.

Categorization of Micra TPS Qualifying Complications or Malfunctions for Survival Analysis on CareLink™

For survival estimation, complication and premature battery depletion data from Medtronic's Complaint Handling System is adjudicated and subsequently cross-referenced with an assessment of device performance from the Medtronic CareLink™ network to categorize if the device is 1) functioning normally, 2) has reached normal battery depletion, or 3) has experienced a qualifying malfunction or complication. This categorization is combined with the CareLink™ data for the total number of implants and implant durations to create survival estimates for the likelihood of experiencing a qualifying complication or malfunction, and normal battery depletion. Ultimately, the data is summarized in two survival curves, one with only qualifying complications or malfunctions and the other including normal battery depletion.

Definition of Qualifying Complication or Malfunction

A longevity analysis is completed for all de-identified devices followed on CareLink™ that have reached the Recommended Replacement Time (RRT), to identify devices that experienced possible early battery depletion. These are findings where the actual reported implant time is less than 80% of the expected longevity calculated using the available device diagnostic information.

Methods for Estimating Transcatheter Pacing Performance continued

Additionally, all reported Micra TPS complaints are adjudicated by subject matter experts and medical safety personnel for inclusion as a product performance event given available information. These product performance events are then cross-referenced with the CareLink™ population for inclusion in the survival analysis.

Product Performance events include, but are not limited to, these that occur 30 days after the implant procedure:

- Premature Battery Depletion
- Cardiac Perforation
- Dislodgement
- Failure to Capture
- Elevated Pacing Threshold
- Tine Fracture

Normal Battery Depletion

A longevity analysis is completed for all devices followed on CareLink™ that are at or within 6 months of RRT to identify devices were taken out of service due to normal battery depletion. The population that is within six months of RRT is assessed against the expected longevity of the product. Normal Battery Depletion is defined as the condition when the device has reached its elective replacement indicator(s) with implant time exceeding 80% of the expected longevity calculated using the available device diagnostic information.

Medtronic CRM establishes expected longevity by statistically characterizing the power consumed by the device and the power available from the device battery. This characterization is applied to a number of parameter configurations. The statistical mean value minus three standard deviations is used as the expected longevity for determining if a battery depleted normally. The actual longevity achieved for any device while implanted will depend on the actual programmed parameters and patient factors and may differ significantly from these estimates.

Statistical and Data Analysis Methods

The performance is expressed in terms of device survival estimates, where "survival" refers to the function of the device, not the survival of the patient. These survival estimates are intended to illustrate the probability that a device will survive for a given number of years without a chronic device-related complication.

Active surveillance normally begins at the time of implant and continues until a product performance or censoring event occurs. Of the several different statistical methods available for survival analysis, PPR survival analysis is estimated using the Standard Actuarial Method, with suspensions assumed distributed evenly within the intervals (Cutler-Ederer Method), and incorporated data from these retrospectively enrolled devices. Thus, in some cases sample sizes may fluctuate from one time interval to the next interval.

The survival estimate is the probability that a device is free of a product performance event or normal battery depletion at a given time point. For example, if a survival probability is 95% after 5 years of service, then the device has a 5% chance of experiencing a related complication or battery depletion in the first 5 years following implant.

Since the survival estimate can become very imprecise with small effective sample sizes, Medtronic truncates the survival curve when the effective sample size is less than 100 devices. The survival charts in the Product Performance Report show the effective sample size for each year interval where we have experience. When the effective sample size reaches 100, the next data point is added to the survival curve.

Methods for Estimating Transcatheter Pacing Performance continued

Because the de-identified information pulled from the CareLink™ network allows for assessment of all devices that were taken out of service there are no adjustments done for underreporting of malfunctions or battery depletion.

Definition of Analysis Dataset

To be included in the US survival analysis dataset, the product must have been successfully implanted and on the CareLink™ network for at least 30 days.

US Reports of Acute Observations

In the first weeks following implantation, physiologic responses and performance can vary until long-term stability is attained. Acute performance may be subject to a number of factors, including patient-specific anatomy, clinical conditions and/or varying implant conditions/techniques. After a period of time, the implant and performance stabilizes. It is for this reason that the CareLink™ analysis, which is intended to measure long-term performance, do not include complications that occur within the first 30 days after implant.

Acute performance information, defined as the first month after implant, but not including the day of the implant procedure, is included in our reporting. The source of this information is the Medtronic complaint handling system database that includes events reported to Medtronic. This information is summarized in tables titled "Acute Observations".

Each Event Report received by Medtronic's complaint handling system is assigned one or more Reason for Report codes based on the information received. The Reason for Report codes have been grouped into Acute Observation categories. The categories are:

- Cardiac Perforation
- Dislodgement
- Failure to Capture
- Failure to Sense
- Elevated Pacing Threshold

Although multiple observations are possible for any given Micra, only one observation is reported per device. The observation reported is the observation highest on the list. For example, if an Event Report includes observations for both Cardiac Perforation and Elevated Pacing Thresholds, Cardiac Perforation is reported.

The event reported to Medtronic may or may not have involved clinical action or product returned to Medtronic. The product may have remained implanted and in service.

US Reports of the Day of Implant Observations

Due to the procedural differences with Micra products compared to transvenous leads and IPGs, information about the clinical experience on the day of implant is included in our reporting. The source of this information is the Medtronic complaint handling system database that includes events reported to Medtronic which may be related to either the Micra device or the delivery system. The information is summarized in tables titled "Day of Implant Observations."

Methods for Estimating Transcatheter Pacing Performance **continued**


Each Event Report received by Medtronic's complaint handling system is assigned one or more Reason for Report codes based on the information received. The Reason for Report codes have been grouped into Day of Implant Observation categories. The categories are:

- Cardiac Perforation
- Dislodgement
- Failure to Capture
- Failure to Sense
- Elevated Pacing Threshold

Although multiple observations are possible for any given Micra, only one observation is reported per device. The observation reported is the observation highest on the list. For example, if an Event Report includes observations for both Cardiac Perforation and Elevated Pacing Thresholds, Cardiac Perforation is reported.

The event reported to Medtronic may or may not have involved clinical action or product returned to Medtronic. The product may have remained implanted and in service.

US Market Release	15Jan2020	CareLink Population		CareLink Qualifying Malfunctions/Complications	
CE Approval Date	31Mar2020	Enrolled	34,606	Dislodgements	3
Registered USA Implants	52,012	Active	25,907	Elevated Pacing Threshold	11
		Cumulative Follow-Up Months	819,654	Failure To Capture	7
		Normal Battery Depletion	119	Premature Battery Depletion	8
				Tine Fracture	1



Years	1	2	3	4	at 58 mo
Excluding NBD	99.9%	99.9%	99.9%	99.9%	99.9%
Including NBD	99.9%	99.7%	99.3%	98.7%	97.5%
Effective Sample Size	27377	16406	7729	2271	140

***Acute Observations N = 52,012**

Cardiac Perforation	13
Dislodgement	30
Elevated Pacing Threshold	90
Failure to Capture	46
Failure To Sense	121

***Day of Implant Observations N = 52,012**

Cardiac Perforation	272
Dislodgement	87
Elevated Pacing Threshold	144
Failure to Capture	84
Failure to Sense	39


The rate of perforation for commercially released Micra devices continues to perform acceptably within levels observed within the post-approval clinical study registry. Overall, predicate clinical studies for Micra VR have demonstrated a reduction in the risk of major complications of 63% through 12 months¹ and 57% through 36 months² relative to transvenous pacing systems.

¹El-Chami et al. Heart Rhythm 2018 15(12): 1800-1807.

²Piccini et al. Heart Rhythm 2020 17(5 Supplement) D-PO02-089

*Data in these tables is sourced direct from the MDT complaint handling database summarizing observations reported to Medtronic relative to the registered US implant population.

US Market Release	06Apr2016	CareLink Population		CareLink Qualifying Malfunctions/Complications	
CE Approval Date	14Apr2015	Enrolled	47,833	Cardiac Perforation	8
Registered USA Implants	72,276	Active	29,819	Dislodgements	1
		Cumulative Follow-Up Months	1,595,882	Elevated Pacing Threshold	45
		Normal Battery Depletion	361	Failure To Capture	7
				Premature Battery Depletion	16
				Tine Fracture	1



■ Including Normal Battery Depletion ■ Excluding Normal Battery Depletion

Years	1	2	3	4	5	6	at 83 mo
Excluding NBD	99.9%	99.9%	99.8%	99.8%	99.8%	99.7%	99.5%
Including NBD	99.9%	99.7%	99.3%	98.9%	98.1%	96.5%	92.7%
Effective Sample Size	40738	30023	20358	12504	6813	2430	224

***Acute Observations N = 72,276**

Cardiac Perforation	21
Dislodgement	22
Elevated Pacing Threshold	167
Failure to Capture	83
Failure To Sense	19

***Day of Implant Observations N = 72,276**

Cardiac Perforation	291
Dislodgement	178
Elevated Pacing Threshold	270
Failure to Capture	133
Failure to Sense	72


The rate of perforation for commercially released Micra devices continues to perform acceptably within levels observed within the post-approval clinical study registry. Overall, predicate clinical studies for Micra VR have demonstrated a reduction in the risk of major complications of 63% through 12 months¹ and 57% through 36 months² relative to transvenous pacing systems.

¹EI-Chami et al. Heart Rhythm 2018 15(12): 1800-1807.

²Piccini et al. Heart Rhythm 2020 17(5 Supplement) D-PO02-089

*Data in these tables is sourced direct from the MDT complaint handling database summarizing observations reported to Medtronic relative to the registered US implant population.

US Market Release	20Apr2023	CareLink Population		CareLink Qualifying Malfunctions/Complications	
CE Approval Date	04Jan2024	Enrolled	11,367	Elevated Pacing Threshold	8
Registered USA Implants	29,232	Active	10,684	Failure To Capture	2
		Cumulative Follow-Up Months	77,488		
		Normal Battery Depletion	5		



Years	1	at 18 mo
Excluding NBD	99.9%	99.9%
Including NBD	99.8%	98.8%
Effective Sample Size	2063	131

***Acute Observations N = 29,232**

Cardiac Perforation	4
Dislodgement	8
Elevated Pacing Threshold	52
Failure to Capture	23
Failure To Sense	31

***Day of Implant Observations N = 29,232**

Cardiac Perforation	51
Dislodgement	47
Elevated Pacing Threshold	51
Failure to Capture	37
Failure to Sense	14


The rate of perforation for commercially released Micra devices continues to perform acceptably within levels observed within the post-approval clinical study registry. Overall, predicate clinical studies for Micra VR have demonstrated a reduction in the risk of major complications of 63% through 12 months¹ and 57% through 36 months² relative to transvenous pacing systems.

¹El-Chami et al. Heart Rhythm 2018 15(12): 1800-1807.

²Piccini et al. Heart Rhythm 2020 17(5 Supplement) D-PO02-089

*Data in these tables is sourced direct from the MDT complaint handling database summarizing observations reported to Medtronic relative to the registered US implant population.

US Market Release	20Apr2023	CareLink Population		CareLink Qualifying Malfunctions/Complications	
CE Approval Date	04Jan2024	Enrolled	4,274	Elevated Pacing Threshold	1
Registered USA Implants	10,393	Active	3,986		
		Cumulative Follow-Up Months	29,562		
		Normal Battery Depletion	3		



Years	1	at 17 mo
Excluding NBD	100.0%	100.0%
Including NBD	99.8%	99.8%
Effective Sample Size	758	113

***Acute Observations N = 10,393**

Dislodgement	7
Elevated Pacing Threshold	22
Failure to Capture	13
Failure To Sense	6

***Day of Implant Observations N = 10,393**

Cardiac Perforation	12
Dislodgement	19
Elevated Pacing Threshold	30
Failure to Capture	26
Failure to Sense	4

The rate of perforation for commercially released Micra devices continues to perform acceptably within levels observed within the post-approval clinical study registry. Overall, predicate clinical studies for Micra VR have demonstrated a reduction in the risk of major complications of 63% through 12 months¹ and 57% through 36 months² relative to transvenous pacing systems.

¹El-Chami et al. Heart Rhythm 2018 15(12): 1800-1807.

²Piccini et al. Heart Rhythm 2020 17(5 Supplement) D-PO02-089

*Data in these tables is sourced direct from the MDT complaint handling database summarizing observations reported to Medtronic relative to the registered US implant population.

Method for Estimating Lead Performance

Medtronic Cardiac Rhythm Management (CRM) has tracked lead survival for over 41 years with its multicenter, global chronic lead studies.

Leads Performance Analysis

Implanted leads operate in the challenging biochemical environment of the human body and the body's response to foreign objects. Implanted leads are also subject to mechanical stresses associated with heart motion, body motion, and patient anatomy.

In this environment, pacemaker and defibrillation leads cannot be expected to last forever. While IPGs and ICDs have a battery that will deplete after a predictable length of time, a lead's longevity cannot be predicted easily based on mechanical measurements, nor are there simple indicators that a lead is approaching the end of its service life. Therefore, regular monitoring while implanted, and evaluation of lead integrity upon IPG or ICD replacement, is necessary to determine if a lead may be approaching the end of its service life.

Shortfalls Of Using Returned Product And Complaints To Estimate Lead Performance

Leads and lead segments returned to Medtronic are analyzed to determine whether or not they meet performance limits established by Medtronic. Although returned product analyses are valuable for gaining insight into lead failure mechanisms, this data cannot be used by itself for determining the survival probability of leads because only a small fraction of leads are explanted and returned for analysis. Some leads are modified due to adverse device effect, however may not be explanted. Additionally, those leads that are returned cannot be assumed to be statistically representative of the performance of the total population for a given lead model. Partial or total lead extraction can result in significant damage to a lead, making a definitive analysis of a suspected failure, and its cause, impossible.

To account for the under reporting inherent with lead survival analysis based solely on returned product, some manufacturers add reported complaints where adverse product performance is evident but the product itself has not been returned. The improvement to the accuracy of survival estimates depends on the degree to which all complaints are actually communicated to the manufacturer. Since not all complaints are communicated to the manufacturer, adding complaints to the survival analysis does not completely solve the under reporting problem.

Lead survival probabilities are more appropriately determined through a prospective clinical surveillance study that includes active follow up with the patients. Although Medtronic monitors returned product analysis and complaints, these are not used to determine lead survival estimates.

Medtronic consolidated all cardiac rhythm surveillance registries into the PAN Registry. The PAN Registry is a patient centric surveillance platform which follows patients implanted with Medtronic cardiac rhythm product(s). The Product Performance Report (PPR) tracks PAN Registry enrolled patients to monitor lead performance status in vivo. The PAN Registry is designed to record clinical observations representative of the total clinical experience. Lead survival estimates include both lead hardware failure and lead-related clinical events that are classified as product performance events, and do not differentiate a lead hardware failure from other clinical events such as Failure to capture, perforation, dislodgement, or concurrent pulse generator failure.

PAN Registry

Medtronic has been monitoring the performance of its cardiac therapy products with a multicenter study since 1983 and has evaluated the performance of more than 131,000 leads, with data reported from countries around the world. Throughout this time period, Medtronic has continually worked to adapt systems and processes to more effectively monitor product performance following market release. The following summarizes current registry requirements.

Method for Estimating Lead Performance continued

Medtronic's product surveillance registry is a world-wide study that has a prospective, non-randomized, observational design. A key purpose of the registry is to provide continuing evaluation and periodic reporting of the long-term reliability and performance of Medtronic market-released cardiac rhythm therapy products. Product-related adverse events, indicating the status of the product, are collected to measure product survival probabilities. The data gathered may also be used to support the design and development of new cardiac therapy products. The registry is designed to continue indefinitely, encompassing new products as they become commercially available.

To ensure a sufficiently large and representative source of data, participating clinical sites must meet pre-specified selection criteria. Patients are enrolled upon implantation of a Medtronic Cardiac rhythm product. Every effort is made to ensure participants are representative of the range of clinical environments in which Medtronic cardiac rhythm products are used. Eligible products for enrollment include Medtronic market-released cardiac rhythm therapy products for which additional information to further characterize product performance following market release is desired. Number of enrollments is reviewed regularly to ensure adequate sample size is obtained for each individual product. Enrollment may be capped and follow-up discontinued when sufficient duration and precision is achieved to effectively characterize product survivability.

Enrolled patients are followed in accordance with the standard care practices of their care provider from their implant date until they can no longer be followed (e.g., death, lost to follow-up, etc.). However, to ensure regular patient status assessments are completed, follow-up windows consistent with typical care practices have been established with a minimum annual follow-up requirement. Product-related adverse events, system modifications and changes in patient status (e.g. death and withdrawal from the study) are required to be reported upon occurrence. This active surveillance model ensures a robust dataset for effectively monitoring product performance.

Patients are eligible for enrollment if:

- Patient is intended to be implanted or has been implanted with a Medtronic market-released cardiac lead connected to a market-released CRT, ICD, or IPG device, and the lead is used for a pacing, sensing, or defibrillation application, or
- Patient participated in a qualifying investigational study of a Medtronic cardiac rhythm product that is now market-released; complete implant and follow-up data are available; and the data can be appropriately and legally released

Each site is required to inform Medtronic whenever a lead event has occurred, a lead is modified, or when a patient is no longer participating. Timely, accurate, and complete reporting and analysis of safety information for surveillance is crucial for the protection of patients, clinicians, and the sponsor. Medtronic continually evaluates the quality and integrity of the data through a combination of on-site and centralized monitoring activities.

Lead Complications

Chronic lead performance is characterized by estimating lead related complication free survival probabilities. For analysis purposes, the complication criteria, which align with the AdvaMed 'Industry Guidance for Uniform Reporting of Clinical Performance of Cardiac Rhythm Management Pulse Generators and Leads', are defined below. These criteria do not, however, enable a lead integrity or "hardware" failure to be conclusively differentiated from other clinical events such as an undetected lead dislodgement, perforation, or concurrent pulse generator failure manifested as a sensing or capture problem.

Method for Estimating Lead Performance continued

All reported lead-related adverse events are classified by the reporting investigator and are adjudicated by an independent event adjudication committee¹. A lead-related event with at least one of the following classifications that is adjudicated by the committee as a complication and occurs more than 30 days after implant is considered a product performance event and will contribute to the survival analysis endpoint. Events with an onset date of 30 days or less after the implant are considered procedure related and therefore are not included as product performance events.

Product performance events include, but are not limited to:

- Failure to capture
- Failure to sense/undersensing
- Oversensing
- Elevated pacing thresholds
- Abnormal pacing impedance (based on lead model, but normal range is typically 200 - 2,000 ohms)
- Abnormal defibrillation impedance (based on lead model, but normal range is typically 20 - 200 ohms)
- Lead Insulation breach
- Lead Conductor fracture, confirmed electrically, visually or radiographically
- Extracardiac stimulation
- Cardiac perforation
- Lead dislodgement
- Structural Lead Failure

Data Analysis Methods

The performance of leads is expressed in terms of lead survival estimates, where "survival" refers to the function of the lead, not the survival of the patient. These survival estimates are intended to illustrate the probability that a lead will survive for a given number of years without a chronic lead-related complication.

Active surveillance normally begins at the time of implant and continues until a product performance or censoring event occurs. In some cases, in the PAN Registry, active surveillance of a device starts after the device was implanted. The survival probability of such device is conditional on survival to the time when the device enters the Registry. This phenomenon is called Left-truncation². PPR lead survival analysis is estimated using the Kaplan-Meier method, a statistical method to incorporate data from these retrospectively enrolled devices, left-truncated data, was applied. The statistical technique uses data from existing devices while appropriately adjusting the device survival curves for the time the device was not actively followed in the registry. Thus, in some cases sample sizes may fluctuate from one time interval to the next interval.

On the following pages, each graph includes a survival curve for each lead model. The survival estimates is the probability that a lead is free of a product performance event at a given time point. For example, if a survival probability is 95% after 5 years of service, then the lead has a 5% chance of experiencing a lead-related complication in the first 5 years following implant.

The data in the tables is rounded to the nearest tenth of one percent. Occasionally, a graph may show 100% survival, but have one or more complications. This occurs because even with the complications, the data rounds to 100%.

The survival curves are statistical estimates. As sample size increases and performance experience accumulates, the estimation improves. Confidence intervals are provided as a way to indicate the degree of certainty of the estimates. Greenwood's formula is used to calculate the standard errors, and the log-log method is used to produce the 2-sided 95% confidence bounds.

Method for Estimating Lead Performance continued

Since the survival estimate can become very imprecise with small effective sample sizes, Medtronic truncates the survival curve when the number of leads entering an interval is less than 50 leads. When the number of leads entering an interval reaches 50, the next data point is added to the survival curve. For those lead models that do not have sufficient sample size, a survival curve will not be presented.

Definition of Analysis Dataset

The survival estimates are derived from all device components successfully enrolled as of the data received cut-off date (e.g. date of data entry at a study site). The number of enrollments is listed for each lead model.

This sample is considered to be representative of the worldwide population, and therefore the survival estimates shown should be representative of the performance worldwide of these models.

Criteria for Model Inclusion

Survival probabilities and the associated study information for a model or model family will be published when more than 100 leads have been enrolled and no fewer than 50 leads followed for at least 6 months. Medtronic, at its discretion, may stop providing updated performance information on lead models that received original US market-release approval 20 or more years ago.

Returned Product Analysis Results

Although the returned product analysis data is not used to generate the survival estimates, the data provides valuable insight into the causes of lead malfunction.

For reporting returned product analysis results, Medtronic CRM considers a lead as having malfunctioned whenever the analysis shows that any parameter was outside the performance limits established by Medtronic while implanted and in service. To be considered a malfunction for returned product analysis reporting, the lead must have been returned to Medtronic and analyzed.

The results of the analysis are presented in four categories. The lead reporting categories are:

Conductor Fracture: Conductor malfunction with complete or intermittent loss of continuity that could interrupt current flow (e.g., fractured conductors), including those associated with clavicle flex fatigue or crush damage.

Insulation Breach: A malfunction of the insulation allowing inappropriate entry of body fluids or inappropriate current flow between the conductors, or between the conductor and the body. Examples include cuts, tears, depressions, abrasions, and material degradation.

Crimps/Welds/Bonds: Any malfunction in a conductor or lead body associated with a point of connection.

Other: Malfunctions of specific lead mechanical attributes, such as sensors, connectors, seal rings, or malfunction modes not included in the three categories above.

A lead subject to a safety advisory is not considered to have malfunctioned unless it has been returned to Medtronic CRM and found, through analysis, to actually have performed outside the performance limits established by Medtronic.

Method for Estimating Lead Performance continued

For leads designed for either ventricular or atrial use, the numbers listed in the Returned Product Analysis tables include both.

The numbers of malfunctions listed in the Returned Product Analysis tables are the actual numbers confirmed in the returned product analysis. The numbers of complications listed in the complications tables are the actual numbers observed in the PSR centers around the world.

US Reports of Acute Lead Observations (Occurring within First Month of Service)

In the first weeks following lead implantation, physiologic responses and lead performance can vary until long-term lead stability is attained. Acute (defined as the first month after implant) lead performance may be subject to a number of factors, including patient-specific anatomy, clinical conditions and/or varying implant conditions/techniques. After a period of time, the implant and the lead performance stabilizes. It is for this reason that the Product Surveillance Registry results, which are intended to measure long-term performance, do not include complications that occur within the first 30 days after implant.

Information about the clinical experience in the first month of service is included in our reporting. The source for this information is Medtronic's complaint handling system database. The information is summarized in tables titled "US Reports of Acute Lead Observations."

Each Event Report received by Medtronic's complaint handling system is assigned one or more Reason for Report codes based on the information received. The Reason for Report codes have been grouped into Acute Lead Observation categories. The categories used for this product performance reporting are drawn from the "FDA Guidance for Submission of Research and Marketing Applications for Permanent Pacemaker Leads and for Pacemaker Lead Adapter 510(k) Submissions." The categories are:

- Cardiac Perforation
- Conductor Fracture
- Lead Dislodgement
- Failure to Capture
- Oversensing
- Failure to Sense
- Insulation Breach
- Impedance Abnormal
- Extracardiac Stimulation
- Unspecified

Although multiple observations are possible for any given lead, only one observation is reported per lead. The observation reported is the observation highest on the list. For example, if an Event Report includes observations for both Lead Dislodgement and Failure to Sense, Lead Dislodgement is reported.

The lead event reported to Medtronic may or may not have involved clinical action or product returned to Medtronic. The lead may have remained implanted and in service.

Method for Estimating Lead Performance continued

Estimated Number of Implanted and Active Leads in the United States

In addition to providing the number of leads enrolled in the PSR, we also provide the number of leads registered as implanted and the number remaining active in the United States based on the status recorded in the Medtronic Device Registration Tracking Application (DTrak).

Footnotes:

¹During the evolution of SLS, event adjudication was transitioned from a Medtronic technical review committee to an independent event adjudication committee in 2011. Data analyses include adjudication using both methods.

²Klein, John P., Moeschberger, Melvin L. Survival Analysis Techniques for Censored and Truncated Data, New York: Springer-Verlag New York, Inc., 1997.

US Market Release	03Aug2005
CE Approval	31Jan2003
Registered USA Implants	299,222
Estimated Active USA Implants	262,255
Fixation Type	Fixed Screw
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	43
Insulation Breach	121
Crimp/Weld/Bond	0
Other	25

US Acute Lead Observations

Cardiac Perforation	95
Conductor Fracture	6
Extra Cardiac Stimulation	12
Failure to Capture	741
Failure to Sense	117
Impedance Out of Range	72
Insulation Breach	2
Lead Dislodgement	962
Oversensing	150
Unspecified Clinical Failure	2


Atrial Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,843
Cumulative Months of Follow-Up	96,807
Number of Leads Active in Study	626

Qualifying Complications

Cardiac Perforation	1	Impedance Out of Range	2
Conductor Fracture	3	Insulation (not further defined)	1
Extra Cardiac Stimulation	1	Lead Dislodgement	4
Failure to Capture	4	Oversensing	1
Failure to Sense	3		



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	at 168 mo
%	99.6%	99.3%	99.3%	99.2%	99.1%	98.9%	98.4%	97.9%	97.6%	96.8%	96.8%	96.8%	96.8%	96.8%
#	1,549	1,213	965	750	592	474	387	326	273	223	196	158	92	51


His Bundle Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,505
Cumulative Months of Follow-Up	56,792
Number of Leads Active in Study	816

Qualifying Complications

Extra Cardiac Stimulation	1	Lead Dislodgement	6
Failure to Capture	34	Oversensing	1
Failure to Sense	3	Other	3



Years	1	2	3	4	5	6	at 84 mo
%	98.3%	97.5%	97.1%	96.4%	95.8%	94.6%	94.6%
#	1,269	1,022	768	507	277	124	66


Left Bundle Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	2,189
Cumulative Months of Follow-Up	32,041
Number of Leads Active in Study	1,817

Qualifying Complications

13	
Conductor Fracture	1
Failure to Capture	7
Lead Dislodgement	5



Years	1	2	3	at 42 mo
%	99.5%	99.0%	98.1%	98.1%
#	1,245	454	178	95


Ventricular Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	2,600
Cumulative Months of Follow-Up	101,289
Number of Leads Active in Study	1,460

Qualifying Complications

27	
Failure to Capture	14
Impedance Out of Range	2
Lead Dislodgement	9
Other	2



Years	1	2	3	4	5	6	7	8	9	10	11	12	at 156 mo
%	99.5%	99.1%	98.8%	98.7%	98.5%	97.9%	97.9%	97.5%	97.5%	97.5%	97.5%	97.5%	97.5%
#	2,220	1,627	1,042	681	447	316	249	199	163	125	109	91	55

US Market Release	23Jun2002
CE Approval	01Feb2002
Registered USA Implants	159,708
Estimated Active USA Implants	77,699
Fixation Type	Tines
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	16
Insulation Breach	64
Crimp/Weld/Bond	0
Other	0

US Acute Lead Observations

Cardiac Perforation	38
Conductor Fracture	2
Extra Cardiac Stimulation	4
Failure to Capture	200
Failure to Sense	19
Impedance Out of Range	16
Lead Dislodgement	215
Oversensing	9


Atrial Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	227
Cumulative Months of Follow-Up	30,107
Number of Leads Active in Study	45

Qualifying Complications

Failure to Sense	2
Lead Dislodgement	1



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	at 210 mo
%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%	99.1%
#	214	205	198	183	167	158	148	136	126	117	110	107	99	96	88	77	68	61


Ventricular Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,192
Cumulative Months of Follow-Up	82,469
Number of Leads Active in Study	110

Qualifying Complications

Conductor Fracture	12
Impedance Out of Range	2
Failure to Capture	4
Insulation (not further defined)	2
Lead Dislodgement	2
Other	1



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	at 216 mo
%	99.5%	99.3%	99.2%	99.2%	99.2%	99.2%	99.2%	99.2%	99.2%	98.8%	98.8%	97.6%	97.6%	96.9%	96.9%	96.9%	96.9%	96.9%
#	1,031	880	742	638	499	408	344	300	261	223	193	161	130	123	111	92	73	50

US Market Release	25Feb2004
CE Approval	14Jun2004
Registered USA Implants	866,415
Estimated Active USA Implants	513,159
Fixation Type	Active Screw In
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	134
Insulation Breach	233
Crimp/Weld/Bond	2
Other	23

US Acute Lead Observations

Cardiac Perforation	293
Conductor Fracture	11
Extra Cardiac Stimulation	30
Failure to Capture	460
Failure to Sense	329
Impedance Out of Range	89
Insulation Breach	2
Lead Dislodgement	1,017
Oversensing	185
Unspecified Clinical Failure	10


Atrial Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	5,120
Cumulative Months of Follow-Up	302,539
Number of Leads Active in Study	1,580

Qualifying Complications

Cardiac Perforation	2	Insulation (not further defined)	3
Conductor Fracture	3	Lead Dislodgement	14
Failure to Capture	9	Oversensing	2
Failure to Sense	3	Other	2



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	at 210 mo
%	99.7%	99.7%	99.6%	99.5%	99.4%	99.0%	98.9%	98.9%	98.9%	98.9%	98.6%	98.6%	98.4%	97.6%	97.6%	97.6%	97.6%	97.6%
#	3,842	3,202	2,749	2,398	2,029	1,701	1,494	1,300	1,122	940	729	521	399	295	191	123	76	52


Ventricular Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,778
Cumulative Months of Follow-Up	123,023
Number of Leads Active in Study	243

Qualifying Complications

Conductor Fracture	1	Impedance Out of Range	2
Extra Cardiac Stimulation	1	Lead Dislodgement	1
Failure to Capture	6	Other	2
Failure to Sense	1		



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	at 198 mo
%	99.7%	99.7%	99.7%	99.6%	99.4%	99.4%	99.0%	99.0%	98.8%	98.6%	98.6%	98.6%	98.1%	98.1%	98.1%	96.9%	96.9%
#	1,459	1,312	1,162	1,001	808	695	571	491	431	371	301	238	201	161	118	75	64

US Market Release	17Sep1998
CE Approval	15Apr1998
Registered USA Implants	186,244
Estimated Active USA Implants	35,894
Fixation Type	Tines
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	21
Insulation Breach	101
Crimp/Weld/Bond	0
Other	0

US Acute Lead Observations

Cardiac Perforation	4
Conductor Fracture	4
Extra Cardiac Stimulation	1
Failure to Capture	35
Impedance Out of Range	2
Insulation Breach	1
Lead Dislodgement	35
Oversensing	1
Unspecified Clinical Failure	1

Product Surveillance Registry Results


Number of Leads Enrolled in Study	1,202
Cumulative Months of Follow-Up	70,346
Number of Leads Active in Study	7

Qualifying Complications

Conductor Fracture	3
Extra Cardiac Stimulation	1
Failure to Capture	12

21

Impedance Out of Range	1
Lead Dislodgement	4



Years	1	2	3	4	5	6	7	8	9	10	11	at 144 mo
%	98.8%	98.7%	98.5%	98.1%	97.8%	97.4%	97.4%	97.4%	97.4%	97.4%	97.4%	97.4%
#	921	822	734	629	515	402	333	279	239	158	94	57

US Market Release	23Jun2002
CE Approval	01Feb2002
Registered USA Implants	127,590
Estimated Active USA Implants	74,399
Fixation Type	J-shape, tines
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	14
Insulation Breach	26
Crimp/Weld/Bond	0
Other	0

US Acute Lead Observations


Cardiac Perforation	6
Conductor Fracture	1
Extra Cardiac Stimulation	1
Failure to Capture	203
Failure to Sense	100
Impedance Out of Range	12
Lead Dislodgement	298
Oversensing	19
Unspecified Clinical Failure	4

Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,782
Cumulative Months of Follow-Up	88,920
Number of Leads Active in Study	561

Qualifying Complications

Conductor Fracture	2	Lead Dislodgement	7
Failure to Capture	8		



Years	1	2	3	4	5	6	7	8	9	10	11	12	at 150 mo
%	99.3%	99.3%	99.2%	98.9%	98.6%	98.4%	98.4%	98.4%	98.4%	98.4%	98.4%	98.4%	98.4%
#	1,408	1,099	872	687	552	466	392	331	265	188	132	80	63

US Market Release	05Oct1998
CE Approval	15Apr1998
Registered USA Implants	89,801
Estimated Active USA Implants	19,586
Fixation Type	J-shape, tines
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	17
Insulation Breach	34
Crimp/Weld/Bond	0
Other	0

US Acute Lead Observations


Failure to Capture	10
Failure to Sense	2
Insulation Breach	1
Lead Dislodgement	37
Oversensing	2
Unspecified Clinical Failure	2

Product Surveillance Registry Results

Number of Leads Enrolled in Study	369
Cumulative Months of Follow-Up	23,164
Number of Leads Active in Study	26

Qualifying Complications

9	
Failure to Capture	4
Failure to Sense	1
Lead Dislodgement	3
Other	1



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	at 162 mo
%	97.7%	97.7%	97.7%	97.7%	97.0%	97.0%	97.0%	96.1%	96.1%	96.1%	96.1%	96.1%	96.1%	96.1%
#	204	182	167	158	134	126	110	105	99	88	82	78	60	54

US Market Release	03Jun1998
CE Approval	05Jun1997
Registered USA Implants	100,059
Estimated Active USA Implants	18,195
Fixation Type	Tines
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	16
Insulation Breach	47
Crimp/Weld/Bond	1
Other	0

US Acute Lead Observations

Cardiac Perforation	2
Conductor Fracture	2
Failure to Capture	23
Impedance Out of Range	4
Insulation Breach	1
Lead Dislodgement	30
Unspecified Clinical Failure	9


Atrial Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	425
Cumulative Months of Follow-Up	42,516
Number of Leads Active in Study	1

Qualifying Complications

3	
Failure to Capture	2
Lead Dislodgement	1



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	at 198 mo
%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%	99.5%
#	411	391	358	322	289	252	219	185	152	128	107	92	74	64	58	51	50

Ventricular Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	992
Cumulative Months of Follow-Up	35,924
Number of Leads Active in Study	6

Qualifying Complications

13	
Failure to Capture	7
Impedance Out of Range	1
Failure to Sense	2
Lead Dislodgement	1
Other	2



Years	1	2	3	4	5	6	7	8	9	10	11	at 144 mo
%	99.3%	99.1%	98.8%	98.4%	98.4%	97.0%	97.0%	96.4%	96.4%	96.4%	96.4%	96.4%
#	474	391	304	264	230	192	167	144	114	95	73	54

US Market Release	31Aug2000
CE Approval	12Aug1999
Registered USA Implants	3,513,770
Estimated Active USA Implants	2,017,444
Fixation Type	Active Screw In
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	1,615
Insulation Breach	1,710
Crimp/Weld/Bond	4
Other	208

US Acute Lead Observations

Cardiac Perforation	1,808
Conductor Fracture	38
Extra Cardiac Stimulation	117
Failure to Capture	2,881
Failure to Sense	1,845
Impedance Out of Range	514
Insulation Breach	17
Lead Dislodgement	5,787
Oversensing	1,035
Unspecified Clinical Failure	26

Atrial Placement


Product Surveillance Registry Results

Number of Leads Enrolled in Study	14,201
Cumulative Months of Follow-Up	704,948
Number of Leads Active in Study	4,837

Qualifying Complications

128

Cardiac Perforation	2	Impedance Out of Range	13
Conductor Fracture	15	Insulation (not further defined)	3
Extra Cardiac Stimulation	3	Lead Dislodgement	49
Failure to Capture	19	Oversensing	3
Failure to Sense	12	Other	9



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	at 240 mo
%	99.6%	99.4%	99.3%	99.1%	98.8%	98.6%	98.5%	98.3%	98.3%	98.2%	98.2%	98.0%	97.8%	97.3%	97.2%	97.2%	96.7%	96.7%	96.7%	96.7%
#	10,115	7,972	6,542	5,460	4,591	3,810	3,227	2,626	2,054	1,643	1,311	1,002	773	603	480	399	313	226	144	77

Ventricular Placement


Product Surveillance Registry Results

Number of Leads Enrolled in Study	3,433
Cumulative Months of Follow-Up	174,388
Number of Leads Active in Study	512

Qualifying Complications

39

Cardiac Perforation	1	Impedance Out of Range	4
Conductor Fracture	10	Insulation (not further defined)	1
Failure to Capture	13	Lead Dislodgement	5
Failure to Sense	1	Oversensing	2
		Other	2



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	at 222 mo
%	99.5%	99.3%	99.2%	99.0%	98.8%	98.4%	98.2%	97.9%	97.6%	97.6%	97.3%	97.3%	96.5%	95.5%	95.5%	95.5%	95.5%	95.5%	95.5%
#	2,268	1,950	1,658	1,380	1,155	917	741	616	533	462	376	284	215	176	142	122	93	72	62

US Market Release	08Feb2011
CE Approval	21Jan2009
Registered USA Implants	207,827
Estimated Active USA Implants	125,134
Fixation Type	Active Screw In
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	120
Insulation Breach	224
Crimp/Weld/Bond	0
Other	12

US Acute Lead Observations

Cardiac Perforation	213
Conductor Fracture	4
Extra Cardiac Stimulation	18
Failure to Capture	144
Failure to Sense	29
Impedance Out of Range	9
Insulation Breach	2
Lead Dislodgement	312
Oversensing	32


Atrial Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	3,142
Cumulative Months of Follow-Up	148,543
Number of Leads Active in Study	1,255

Qualifying Complications

Conductor Fracture	3	Lead Dislodgement	12
Failure to Capture	3	Oversensing	2
		Other	1



Years	1	2	3	4	5	6	7	8	9	10	11	12	at 150 mo
%	99.8%	99.6%	99.6%	99.4%	99.3%	98.9%	98.4%	98.4%	98.4%	98.1%	98.1%	98.1%	98.1%
#	2,528	2,202	1,878	1,460	766	453	397	354	317	294	211	111	74


Ventricular Placement

Product Surveillance Registry Results

Number of Leads Enrolled in Study	3,073
Cumulative Months of Follow-Up	145,681
Number of Leads Active in Study	1,237

Qualifying Complications

Conductor Fracture	3	Impedance Out of Range	2
Failure to Capture	8	Lead Dislodgement	3
Failure to Sense	1	Oversensing	2
		Other	1



Years	1	2	3	4	5	6	7	8	9	10	11	12	at 150 mo
%	99.7%	99.7%	99.6%	99.6%	99.6%	99.3%	98.3%	98.0%	98.0%	98.0%	98.0%	97.5%	97.5%
#	2,526	2,183	1,851	1,426	735	423	375	332	300	277	197	104	72

US Market Release	03Jun1998
CE Approval	25Sep1997
Registered USA Implants	141,709
Estimated Active USA Implants	28,955
Fixation Type	Tines
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	28
Insulation Breach	73
Crimp/Weld/Bond	0
Other	1

US Acute Lead Observations


Cardiac Perforation	7
Conductor Fracture	3
Extra Cardiac Stimulation	3
Failure to Capture	49
Failure to Sense	7
Impedance Out of Range	1
Insulation Breach	3
Lead Dislodgement	72
Oversensing	1
Unspecified Clinical Failure	8

Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,218
Cumulative Months of Follow-Up	54,819
Number of Leads Active in Study	11

Qualifying Complications

Extra Cardiac Stimulation	1	Impedance Out of Range	1
Failure to Capture	3	Lead Dislodgement	5



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	at 162 mo
%	99.5%	99.3%	99.2%	98.9%	98.9%	98.6%	98.6%	98.6%	98.6%	98.6%	97.8%	97.8%	97.8%	97.8%
#	814	652	517	421	335	265	219	174	149	133	110	82	57	53

US Market Release	03Jun1998
CE Approval	05Jun1997
Registered USA Implants	64,872
Estimated Active USA Implants	14,210
Fixation Type	Tines
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	24
Insulation Breach	45
Crimp/Weld/Bond	0
Other	0

US Acute Lead Observations


Conductor Fracture	1
Failure to Capture	31
Failure to Sense	2
Impedance Out of Range	1
Lead Dislodgement	39
Unspecified Clinical Failure	3

Product Surveillance Registry Results

Number of Leads Enrolled in Study	370
Cumulative Months of Follow-Up	9,549
Number of Leads Active in Study	8

Qualifying Complications

Failure to Capture	5
Impedance Out of Range	1
Lead Dislodgement	1
Oversensing	1



Years	1	2	3	4	5	6	at 84 mo
%	100.0%	98.2%	97.2%	96.0%	96.0%	96.0%	96.0%
#	141	107	84	77	63	55	55

US Market Release	03Jun1998
CE Approval	25Sep1997
Registered USA Implants	37,334
Estimated Active USA Implants	9,760
Fixation Type	Tines
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	7
Insulation Breach	7
Crimp/Weld/Bond	0
Other	0

US Acute Lead Observations


Cardiac Perforation	1
Failure to Capture	4
Failure to Sense	3
Lead Dislodgement	43
Oversensing	1
Unspecified Clinical Failure	1

Product Surveillance Registry Results

Number of Leads Enrolled in Study	722
Cumulative Months of Follow-Up	39,887
Number of Leads Active in Study	24

Qualifying Complications

5	
Failure to Capture	3
Lead Dislodgement	2



Years	1	2	3	4	5	6	7	8	9	10	11	12	at 156 mo
%	99.6%	99.3%	99.3%	98.9%	98.9%	98.9%	98.9%	98.9%	98.9%	98.9%	98.9%	98.9%	98.9%
#	523	432	351	299	249	197	169	154	132	118	102	71	51

US Market Release	25Jun2001
CE Approval	23Mar2001
Registered USA Implants	17,612
Estimated Active USA Implants	5,448
Fixation Type	Tines
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	16
Insulation Breach	18
Crimp/Weld/Bond	0
Other	0

US Acute Lead Observations


Failure to Capture	4
Lead Dislodgement	14
Unspecified Clinical Failure	2

Product Surveillance Registry Results

Number of Leads Enrolled in Study	43
Cumulative Months of Follow-Up	4,761
Number of Leads Active in Study	10

Qualifying Complications

3	
Conductor Fracture	1
Insulation (not further defined)	1
Oversensing	1



Years	at 0 mo
%	100.0%
#	

6721 Epicardial Patch

US Market Release	31Mar1994
CE Approval	01Jan1993
Registered USA Implants	3,434
Estimated Active USA Implants	858
Fixation Type	Suture
Pace Sense Polarity	n/a
Steroid Indicator	None

US Returned Product Analysis

Conductor Fracture	15
Insulation Breach	1
Crimp/Weld/Bond	0
Other	0

US Acute Lead Observations


Cardiac Perforation	1
Conductor Fracture	2
Failure to Capture	4
Failure to Sense	2
Impedance Out of Range	24
Oversensing	1

Product Surveillance Registry Results

Number of Leads Enrolled in Study	417
Cumulative Months of Follow-Up	24,242
Number of Leads Active in Study	5

Qualifying Complications

Conductor Fracture	21	Impedance Out of Range	4
Failure to Capture	8	Insulation (not further defined)	2
Other	16		



Years	1	2	3	4	5	6	7	8	at 108 mo
%	96.6%	94.6%	92.1%	91.0%	89.3%	84.7%	83.3%	83.3%	83.3%
#	347	319	274	221	190	136	102	66	57

US Market Release	02Sep2004
CE Approval	
Registered USA Implants	354
Estimated Active USA Implants	63
Fixation Type	Tines
Pace Sense Polarity	True Bipolar/One Coil
Steroid Indicator	Yes

US Returned Product Analysis


Conductor Fracture	5
Insulation Breach	0
Crimp/Weld/Bond	0
Other	0

US Acute Lead Observations

Unspecified Clinical Failure	1
------------------------------	---

Product Surveillance Registry Results

Number of Leads Enrolled in Study	4
Cumulative Months of Follow-Up	332
Number of Leads Active in Study	



Years	at 0 mo
%	100.0%
#	

US Market Release	02Sep2004
CE Approval	
Registered USA Implants	8,081
Estimated Active USA Implants	1,148
Fixation Type	Active Screw In
Pace Sense Polarity	True Bipolar/One Coil
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	670
Insulation Breach	1
Crimp/Weld/Bond	0
Other	5

US Acute Lead Observations

Cardiac Perforation	1
Conductor Fracture	2
Failure to Capture	1
Failure to Sense	1
Lead Dislodgement	1
Oversensing	3
Unspecified Clinical Failure	1

Product Surveillance Registry Results


Number of Leads Enrolled in Study	311
Cumulative Months of Follow-Up	18,179
Number of Leads Active in Study	6

Qualifying Complications

Conductor Fracture	36
Failure to Capture	3
Failure to Sense	1

59

Impedance Out of Range	10
Lead Dislodgement	2
Oversensing	7



Years	1	2	3	4	5	6	7	at 90 mo
%	98.2%	96.2%	93.1%	88.3%	82.2%	74.3%	72.3%	72.3%
#	261	232	204	166	137	104	70	56

US Market Release	01Nov2008
CE Approval	31Mar2008
Registered USA Implants	69,799
Estimated Active USA Implants	40,897
Fixation Type	Active Screw In
Pace Sense Polarity	True Bipolar/One Coil
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	505
Insulation Breach	14
Crimp/Weld/Bond	0
Other	44

US Acute Lead Observations


Cardiac Perforation	30
Conductor Fracture	3
Extra Cardiac Stimulation	2
Failure to Capture	43
Failure to Sense	16
Impedance Out of Range	38
Insulation Breach	1
Lead Dislodgement	69
Oversensing	71
Unspecified Clinical Failure	5

Product Surveillance Registry Results

Number of Leads Enrolled in Study	3,026
Cumulative Months of Follow-Up	176,512
Number of Leads Active in Study	542

Qualifying Complications

Cardiac Perforation	1	Impedance Out of Range	8
Conductor Fracture	26	Lead Dislodgement	8
Extra Cardiac Stimulation	1	Oversensing	9
Failure to Capture	8	Other	6
Failure to Sense	1	Unspecified Clinical Failure	1



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	at 168 mo
%	99.5%	99.3%	99.0%	98.6%	98.5%	98.1%	97.6%	96.8%	96.0%	94.9%	94.7%	94.5%	93.5%	93.5%
#	2,497	2,053	1,673	1,367	1,162	1,008	844	708	615	510	407	278	140	73

US Market Release	02Aug2012
CE Approval	12Jul2012
Registered USA Implants	435,164
Estimated Active USA Implants	359,135
Fixation Type	Active Screw In
Pace Sense Polarity	True Bipolar/One Coil
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	890
Insulation Breach	37
Crimp/Weld/Bond	2
Other	105

US Acute Lead Observations

Cardiac Perforation	212
Conductor Fracture	24
Extra Cardiac Stimulation	35
Failure to Capture	529
Failure to Sense	185
Impedance Out of Range	156
Insulation Breach	3
Lead Dislodgement	719
Oversensing	417

Product Surveillance Registry Results


Number of Leads Enrolled in Study	10,106
Cumulative Months of Follow-Up	449,463
Number of Leads Active in Study	3,588

Qualifying Complications

Cardiac Perforation	2
Conductor Fracture	57
Extra Cardiac Stimulation	1
Failure to Capture	15
Failure to Sense	1

125

Impedance Out of Range	10
Insulation (not further defined)	3
Lead Dislodgement	24
Oversensing	6
Other	5
Unspecified Clinical Failure	1



Years	1	2	3	4	5	6	7	8	9	10	at 132 mo
%	99.6%	99.5%	99.2%	99.0%	98.7%	98.3%	97.9%	97.5%	96.9%	96.3%	95.9%
#	8,054	5,983	4,720	3,853	3,232	2,583	2,047	1,376	742	356	104

6937A Transvene SVC-CS

US Market Release	06Apr2001
CE Approval	
Registered USA Implants	3,178
Estimated Active USA Implants	1,654
Fixation Type	Passive
Pace Sense Polarity	One Coil
Steroid Indicator	None

US Returned Product Analysis

Conductor Fracture	6
Insulation Breach	0
Crimp/Weld/Bond	0
Other	0

US Acute Lead Observations


Cardiac Perforation	1
Conductor Fracture	3
Impedance Out of Range	2
Lead Dislodgement	1
Oversensing	2
Unspecified Clinical Failure	2

Product Surveillance Registry Results

Number of Leads Enrolled in Study	127
Cumulative Months of Follow-Up	14,613
Number of Leads Active in Study	7

Qualifying Complications

Conductor Fracture	6	Impedance Out of Range	2
		Insulation (not further defined)	2
		Lead Dislodgement	1
		Other	1
		Unspecified Clinical Failure	4



Years	1	2	3	4	5	6	7	8	9	at 114 mo
%	99.2%	99.2%	99.2%	98.3%	95.6%	93.6%	92.5%	91.3%	88.8%	88.8%
#	119	117	114	109	98	87	80	73	61	55

US Market Release	13Dec2000
CE Approval	05Nov1999
Registered USA Implants	44,865
Estimated Active USA Implants	11,716
Fixation Type	Tines
Pace Sense Polarity	True Bipolar/Two Coils
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	242
Insulation Breach	4
Crimp/Weld/Bond	1
Other	4

US Acute Lead Observations


Conductor Fracture	2
Failure to Capture	17
Failure to Sense	3
Impedance Out of Range	10
Lead Dislodgement	24
Oversensing	18
Unspecified Clinical Failure	6

Product Surveillance Registry Results

Number of Leads Enrolled in Study	640
Cumulative Months of Follow-Up	39,169
Number of Leads Active in Study	52

Qualifying Complications

Conductor Fracture	18	Impedance Out of Range	6
Failure to Capture	4	Oversensing	3
Failure to Sense	1	Other	1
		Unspecified Clinical Failure	1



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	at 162 mo
%	100.0%	99.8%	99.2%	97.3%	94.8%	91.7%	91.1%	90.6%	89.9%	89.9%	89.1%	88.3%	86.2%	86.2%
#	502	418	352	290	228	192	165	148	133	118	109	94	66	57

US Market Release	05Jan2016
CE Approval	12Sep2013
Registered USA Implants	4,812
Estimated Active USA Implants	4,183
Fixation Type	Tines
Pace Sense Polarity	True Bipolar/Two Coils
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	2
Insulation Breach	0
Crimp/Weld/Bond	0
Other	0

US Acute Lead Observations


Cardiac Perforation	1
Failure to Capture	7
Failure to Sense	2
Impedance Out of Range	1
Lead Dislodgement	9
Oversensing	6

Product Surveillance Registry Results

Number of Leads Enrolled in Study	67
Cumulative Months of Follow-Up	2,491
Number of Leads Active in Study	38

Qualifying Complications

Conductor Fracture	1
--------------------	---



Years	at 12 mo
%	100.0%
#	54

US Market Release	12Nov2001
CE Approval	04Oct2001
Registered USA Implants	375,772
Estimated Active USA Implants	124,006
Fixation Type	Active Screw In
Pace Sense Polarity	True Bipolar/Two Coils
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	1,441
Insulation Breach	104
Crimp/Weld/Bond	4
Other	198

US Acute Lead Observations

Cardiac Perforation	29
Conductor Fracture	26
Extra Cardiac Stimulation	2
Failure to Capture	83
Failure to Sense	36
Impedance Out of Range	61
Insulation Breach	4
Lead Dislodgement	124
Oversensing	141
Unspecified Clinical Failure	20

Product Surveillance Registry Results


Number of Leads Enrolled in Study	4,600
Cumulative Months of Follow-Up	305,785
Number of Leads Active in Study	419

Qualifying Complications

Cardiac Perforation	1
Conductor Fracture	42
Failure to Capture	8
Failure to Sense	2

105

Impedance Out of Range	13
Insulation (not further defined)	6
Lead Dislodgement	5
Oversensing	21
Other	4
Unspecified Clinical Failure	3



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	at 240 mo
%	99.5%	99.3%	99.0%	98.7%	98.2%	97.9%	97.5%	97.1%	96.7%	96.2%	95.9%	95.5%	95.1%	94.7%	94.4%	92.9%	92.9%	91.4%	89.5%	86.9%
#	3,308	2,909	2,552	2,262	2,032	1,788	1,547	1,387	1,244	1,091	926	742	595	437	287	187	127	111	85	59

US Market Release	13Feb2012
CE Approval	12Mar2010
Registered USA Implants	141,144
Estimated Active USA Implants	96,879
Fixation Type	Active Screw In
Pace Sense Polarity	True Bipolar/Two Coils
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	268
Insulation Breach	15
Crimp/Weld/Bond	1
Other	37

US Acute Lead Observations


Cardiac Perforation	40
Conductor Fracture	16
Extra Cardiac Stimulation	12
Failure to Capture	128
Failure to Sense	49
Impedance Out of Range	39
Insulation Breach	1
Lead Dislodgement	243
Oversensing	91

Product Surveillance Registry Results

Number of Leads Enrolled in Study	2,445
Cumulative Months of Follow-Up	140,838
Number of Leads Active in Study	510

Qualifying Complications

Conductor Fracture	15	Impedance Out of Range	1
Failure to Capture	4	Lead Dislodgement	1
Failure to Sense	4	Oversensing	2
		Other	1



Years	1	2	3	4	5	6	7	8	9	10	11	at 144 mo
%	99.7%	99.5%	99.4%	99.4%	99.0%	98.9%	98.2%	97.9%	97.4%	97.2%	97.2%	97.2%
#	1,911	1,598	1,384	1,174	1,023	858	736	627	526	412	287	95

US Market Release	02Sep2004
CE Approval	
Registered USA Implants	10,381
Estimated Active USA Implants	1,622
Fixation Type	Tines
Pace Sense Polarity	True Bipolar/Two Coils
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	219
Insulation Breach	3
Crimp/Weld/Bond	0
Other	6

US Acute Lead Observations


Conductor Fracture	2
Failure to Capture	7
Lead Dislodgement	7
Oversensing	1
Unspecified Clinical Failure	3

Product Surveillance Registry Results

Number of Leads Enrolled in Study	40
Cumulative Months of Follow-Up	2,329
Number of Leads Active in Study	2

Qualifying Complications

Conductor Fracture	5
Impedance Out of Range	1



Years	at 0 mo
%	100.0%
#	

US Market Release	02Sep2004
CE Approval	
Registered USA Implants	186,211
Estimated Active USA Implants	23,768
Fixation Type	Active Screw In
Pace Sense Polarity	True Bipolar/Two Coils
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	8,183
Insulation Breach	37
Crimp/Weld/Bond	3
Other	119

US Acute Lead Observations

Cardiac Perforation	10
Conductor Fracture	52
Failure to Capture	31
Failure to Sense	19
Impedance Out of Range	20
Insulation Breach	5
Lead Dislodgement	22
Oversensing	37
Unspecified Clinical Failure	24

Product Surveillance Registry Results


Number of Leads Enrolled in Study	986
Cumulative Months of Follow-Up	58,228
Number of Leads Active in Study	27

Qualifying Complications

Conductor Fracture	78
Failure to Capture	5
Failure to Sense	6

136

Impedance Out of Range	19
Insulation (not further defined)	2
Lead Dislodgement	1
Oversensing	21
Other	3
Unspecified Clinical Failure	1



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	at 180 mo
%	98.6%	96.5%	93.4%	91.0%	88.2%	84.5%	81.6%	79.0%	78.0%	76.6%	71.0%	68.5%	66.2%	63.5%	63.5%
#	719	626	532	458	392	343	281	236	187	153	125	96	79	65	55

6996 Sub-Q Lead


		US Returned Product Analysis		US Acute Lead Observations	
US Market Release	11Jun2001	Conductor Fracture	39	Cardiac Perforation	1
CE Approval	19Dec1997	Insulation Breach	0	Failure to Capture	1
Registered USA Implants	5,857	Crimp/Weld/Bond	0	Impedance Out of Range	19
Estimated Active USA Implants	2,653	Other	0	Insulation Breach	1
Fixation Type	Suture on Anchor Sleeve			Lead Dislodgement	3
Pace Sense Polarity	One Coil			Oversensing	1
Steroid Indicator	None				

Product Surveillance Registry Results

Number of Leads Enrolled in Study	56
Cumulative Months of Follow-Up	2,620
Number of Leads Active in Study	4

Qualifying Complications

Conductor Fracture	1	Impedance Out of Range	3
--------------------	---	------------------------	---



Years	at 0 mo
%	100.0%
#	

US Market Release	20Oct2023
CE Approval	17Feb2023
Registered USA Implants	1,792
Estimated Active USA Implants	1,731
Fixation Type	Shaped passive fixation
Pace Sense Polarity	True Bipolar/Two Coils
Steroid Indicator	None

US Returned Product Analysis

Conductor Fracture
Insulation Breach
Crimp/Weld/Bond
Other

US Acute Lead Observations


Cardiac Perforation	0
Failure to Capture	2
Failure to Sense	4
Impedance Out of Range	21
Lead Dislodgement	12
Oversensing	39

Product Surveillance Registry Results

Number of Leads Enrolled in Study	556
Cumulative Months of Follow-Up	1,151
Number of Leads Active in Study	542

Qualifying Complications

Lead Dislodgement	1
-------------------	---



Years	at 6 mo
%	99.5%
#	81

2187 Attain LV

US Market Release	28Aug2001
CE Approval	
Registered USA Implants	11,921
Estimated Active USA Implants	975
Fixation Type	Distal Continuous Curve
Pace Sense Polarity	Unipolar
Steroid Indicator	None

US Returned Product Analysis

Conductor Fracture	1
Insulation Breach	3
Crimp/Weld/Bond	0
Other	3

US Acute Lead Observations


Extra Cardiac Stimulation	1
Failure to Capture	3
Failure to Sense	1
Lead Dislodgement	9

Product Surveillance Registry Results

Number of Leads Enrolled in Study	140
Cumulative Months of Follow-Up	7,282
Number of Leads Active in Study	4

Qualifying Complications

Failure to Capture	3
--------------------	---



Years	1	2	3	at 48 mo
%	99.1%	98.0%	98.0%	98.0%
#	101	85	65	52

4193 Attain OTW

US Market Release	03May2002
CE Approval	22Dec2000
Registered USA Implants	100,665
Estimated Active USA Implants	12,055
Fixation Type	Double Curve
Pace Sense Polarity	Unipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	92
Insulation Breach	31
Crimp/Weld/Bond	0
Other	15

US Acute Lead Observations


Extra Cardiac Stimulation	18
Failure to Capture	11
Lead Dislodgement	45
Oversensing	1
Unspecified Clinical Failure	2

Product Surveillance Registry Results

Number of Leads Enrolled in Study	805
Cumulative Months of Follow-Up	42,759
Number of Leads Active in Study	17

Qualifying Complications

Conductor Fracture	1	Impedance Out of Range	2
Extra Cardiac Stimulation	10	Lead Dislodgement	16
Failure to Capture	20	Unspecified Clinical Failure	3



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	at 168 mo
%	96.0%	95.1%	94.4%	94.1%	93.4%	92.6%	92.2%	91.7%	91.2%	90.5%	90.5%	88.4%	87.2%	87.2%
#	569	444	376	304	252	228	193	171	139	118	97	78	63	51

US Market Release	24Aug2004
CE Approval	14Jul2003
Registered USA Implants	114,259
Estimated Active USA Implants	28,267
Fixation Type	Double Curve
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	48
Insulation Breach	167
Crimp/Weld/Bond	0
Other	2

US Acute Lead Observations

Cardiac Perforation	2
Conductor Fracture	3
Extra Cardiac Stimulation	49
Failure to Capture	43
Impedance Out of Range	9
Lead Dislodgement	153
Oversensing	2
Unspecified Clinical Failure	4

Product Surveillance Registry Results


Number of Leads Enrolled in Study	1,654
Cumulative Months of Follow-Up	101,664
Number of Leads Active in Study	119

Qualifying Complications

Conductor Fracture	2
Extra Cardiac Stimulation	11
Failure to Capture	22

68

Insulation (ESC)	1
Insulation (not further defined)	2
Lead Dislodgement	30



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	at 186 mo
%	98.6%	97.4%	96.7%	96.1%	95.6%	94.3%	94.1%	93.5%	93.5%	93.3%	93.0%	93.0%	93.0%	91.6%	91.6%	91.6%
#	1,238	1,046	897	769	697	616	506	430	369	322	264	192	154	100	69	63

US Market Release	15Aug2008
CE Approval	13May2005
Registered USA Implants	17,447
Estimated Active USA Implants	6,147
Fixation Type	Deployable Lobe Fixation
Pace Sense Polarity	Unipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	10
Insulation Breach	3
Crimp/Weld/Bond	0
Other	2

US Acute Lead Observations


Extra Cardiac Stimulation	30
Failure to Capture	21
Impedance Out of Range	4
Lead Dislodgement	30
Unspecified Clinical Failure	1

Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,486
Cumulative Months of Follow-Up	89,261
Number of Leads Active in Study	106

Qualifying Complications

Conductor Fracture	4	Impedance Out of Range	2
Extra Cardiac Stimulation	18	Insulation (not further defined)	6
Failure to Capture	9	Lead Dislodgement	5
		Other	1



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	at 162 mo
%	99.1%	98.5%	98.1%	97.6%	97.0%	96.7%	96.3%	95.6%	95.3%	95.3%	93.8%	93.8%	93.8%	93.8%
#	1,243	1,072	924	747	620	510	416	323	263	209	163	118	78	59

US Market Release	15May2009
CE Approval	24Jul2007
Registered USA Implants	69,226
Estimated Active USA Implants	27,613
Fixation Type	Double Curve
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	28
Insulation Breach	2
Crimp/Weld/Bond	0
Other	9

US Acute Lead Observations

Cardiac Perforation	3
Conductor Fracture	2
Extra Cardiac Stimulation	99
Failure to Capture	68
Failure to Sense	1
Impedance Out of Range	12
Insulation Breach	1
Lead Dislodgement	228
Oversensing	1
Unspecified Clinical Failure	2


Product Surveillance Registry Results

Number of Leads Enrolled in Study	2,325
Cumulative Months of Follow-Up	122,201
Number of Leads Active in Study	162

Qualifying Complications

Conductor Fracture	3
Extra Cardiac Stimulation	17
Failure to Capture	37
Other	87

Impedance Out of Range	2
Insulation (not further defined)	1
Lead Dislodgement	23
Other	4



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	at 168 mo
%	98.0%	97.3%	96.6%	95.9%	95.7%	95.2%	94.9%	94.2%	93.9%	93.7%	93.7%	93.2%	93.2%	93.2%
#	1,892	1,503	1,196	968	798	640	507	416	351	286	223	160	106	51

US Market Release	01Apr2011
CE Approval	18Dec2009
Registered USA Implants	35,324
Estimated Active USA Implants	17,117
Fixation Type	Double Curve
Pace Sense Polarity	Dual Electrodes
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	4
Insulation Breach	0
Crimp/Weld/Bond	2
Other	4

US Acute Lead Observations

Cardiac Perforation	2
Conductor Fracture	1
Extra Cardiac Stimulation	65
Failure to Capture	39
Impedance Out of Range	11
Insulation Breach	4
Lead Dislodgement	120


Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,473
Cumulative Months of Follow-Up	80,295
Number of Leads Active in Study	168

Qualifying Complications

Conductor Fracture	1	38
Extra Cardiac Stimulation	12	
Failure to Capture	9	

Lead Dislodgement	14
Other	2



Years	1	2	3	4	5	6	7	8	9	10	11	at 144 mo
%	98.7%	97.9%	97.7%	97.4%	96.8%	96.4%	96.4%	96.1%	96.1%	95.8%	95.8%	95.8%
#	1,163	941	776	660	556	474	409	330	284	225	138	52

US Market Release	01Aug2014
CE Approval	01Jan2013
Registered USA Implants	126,764
Estimated Active USA Implants	97,709
Fixation Type	Double Curve
Pace Sense Polarity	Quadripolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	7
Insulation Breach	0
Crimp/Weld/Bond	0
Other	27

US Acute Lead Observations


Cardiac Perforation	7
Conductor Fracture	1
Extra Cardiac Stimulation	247
Failure to Capture	186
Failure to Sense	1
Impedance Out of Range	51
Lead Dislodgement	275

Product Surveillance Registry Results

Number of Leads Enrolled in Study	2,277
Cumulative Months of Follow-Up	116,840
Number of Leads Active in Study	531

Qualifying Complications

Extra Cardiac Stimulation	5	Lead Dislodgement	16
Failure to Capture	5	Other	3



Years	1	2	3	4	5	6	7	8	9	at 114 mo
%	99.1%	98.7%	98.6%	98.6%	98.4%	98.3%	98.3%	98.3%	98.3%	98.3%
#	1,904	1,619	1,380	1,163	948	688	507	319	147	78

4396 Attain Ability Straight

US Market Release	31Mar2011
CE Approval	18Dec2009
Registered USA Implants	8,492
Estimated Active USA Implants	4,379
Fixation Type	Tines
Pace Sense Polarity	Dual Electrodes
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	5
Insulation Breach	1
Crimp/Weld/Bond	0
Other	0

US Acute Lead Observations


Cardiac Perforation	1
Conductor Fracture	2
Extra Cardiac Stimulation	21
Failure to Capture	14
Lead Dislodgement	35

Product Surveillance Registry Results

Number of Leads Enrolled in Study	485
Cumulative Months of Follow-Up	26,686
Number of Leads Active in Study	69

Qualifying Complications

Extra Cardiac Stimulation	1	Insulation (not further defined)	1
Failure to Capture	4	Lead Dislodgement	4



Years	Months After Implant									
	1	2	3	4	5	6	7	8	9	at 120 mo
%	99.8%	99.2%	98.2%	98.2%	97.7%	97.7%	97.7%	97.7%	97.7%	97.7%
#	386	310	271	236	198	157	128	108	94	60

4398 Attain Performa Straight

US Market Release	10Dec2014
CE Approval	01Jan2013
Registered USA Implants	44,022
Estimated Active USA Implants	34,958
Fixation Type	Tines
Pace Sense Polarity	Quadripolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	4
Insulation Breach	0
Crimp/Weld/Bond	0
Other	8

US Acute Lead Observations


Cardiac Perforation	8
Conductor Fracture	1
Extra Cardiac Stimulation	116
Failure to Capture	84
Impedance Out of Range	15
Lead Dislodgement	46

Product Surveillance Registry Results

Number of Leads Enrolled in Study	2,138
Cumulative Months of Follow-Up	85,359
Number of Leads Active in Study	1,018

Qualifying Complications

Extra Cardiac Stimulation	1	Impedance Out of Range	1
Failure to Capture	7	Lead Dislodgement	8



Months After Implant

Years	1	2	3	4	5	6	7	at 96 mo
%	99.7%	99.6%	99.3%	99.1%	98.9%	98.9%	98.4%	98.4%
#	1,745	1,396	1,076	795	553	319	151	61

US Market Release	10Dec2014
CE Approval	01Jan2013
Registered USA Implants	82,691
Estimated Active USA Implants	66,788
Fixation Type	S-shape
Pace Sense Polarity	Quadripolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	7
Insulation Breach	0
Crimp/Weld/Bond	0
Other	15

US Acute Lead Observations

Cardiac Perforation	12
Conductor Fracture	2
Extra Cardiac Stimulation	148
Failure to Capture	111
Impedance Out of Range	40
Lead Dislodgement	96
Oversensing	1


Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,388
Cumulative Months of Follow-Up	67,043
Number of Leads Active in Study	365

Qualifying Complications

Extra Cardiac Stimulation	3
Failure to Capture	1
Failure to Sense	1

17 Lead Dislodgement 12



Years	1	2	3	4	5	6	7	8	at 102 mo
%	99.2%	99.0%	98.9%	98.7%	98.5%	98.5%	98.5%	98.0%	98.0%
#	1,168	980	811	678	550	369	222	116	71

4798 Attain Stability Quad

US Market Release	03Jun2019
CE Approval	24Apr2017
Registered USA Implants	63,583
Estimated Active USA Implants	59,000
Fixation Type	Non-electrically Active Side Fixation
Pace Sense Polarity	Quadripolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	1
Insulation Breach	0
Crimp/Weld/Bond	0
Other	15

US Acute Lead Observations

Cardiac Perforation	10
Conductor Fracture	2
Extra Cardiac Stimulation	122
Failure to Capture	143
Impedance Out of Range	47
Lead Dislodgement	124
Oversensing	1


Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,505
Cumulative Months of Follow-Up	34,824
Number of Leads Active in Study	908

Qualifying Complications

Conductor Fracture	1
Extra Cardiac Stimulation	2
Failure to Capture	4

Lead Dislodgement 2



Years	1	2	3	4	5	at 66 mo
%	99.7%	99.5%	99.5%	99.1%	99.1%	99.1%
#	1,055	582	333	168	89	61

US Market Release	06Sep1996
CE Approval	01Jan1993
Registered USA Implants	24,378
Estimated Active USA Implants	6,762
Fixation Type	Suture
Pace Sense Polarity	Unipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	302
Insulation Breach	65
Crimp/Weld/Bond	1
Other	0

US Acute Lead Observations


Cardiac Perforation	1
Conductor Fracture	1
Failure to Capture	11
Failure to Sense	8
Impedance Out of Range	21
Oversensing	2
Unspecified Clinical Failure	3

Product Surveillance Registry Results

Number of Leads Enrolled in Study	234
Cumulative Months of Follow-Up	7,531
Number of Leads Active in Study	3

Qualifying Complications

18	
Conductor Fracture	10
Failure to Capture	4
Failure to Sense	1
Insulation (not further defined)	1
Oversensing	2



Years	1	2	3	at 48 mo
%	97.9%	95.1%	94.0%	85.6%
#	118	100	82	60

US Market Release	16Sep1999
CE Approval	21Apr1998
Registered USA Implants	66,017
Estimated Active USA Implants	36,191
Fixation Type	Suture
Pace Sense Polarity	Bipolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	157
Insulation Breach	98
Crimp/Weld/Bond	0
Other	1

US Acute Lead Observations

Cardiac Perforation	1
Conductor Fracture	4
Extra Cardiac Stimulation	7
Failure to Capture	107
Failure to Sense	15
Impedance Out of Range	22
Insulation Breach	1
Lead Dislodgement	8
Oversensing	33


Product Surveillance Registry Results

Number of Leads Enrolled in Study	1,055
Cumulative Months of Follow-Up	74,317
Number of Leads Active in Study	181

Qualifying Complications

Conductor Fracture	34
Extra Cardiac Stimulation	2
Failure to Capture	30
Failure to Sense	3
Other	111

Impedance Out of Range	5
Insulation (not further defined)	6
Lead Dislodgement	1
Oversensing	27
Other	3



Years	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	at 186 mo
%	99.4%	97.6%	96.0%	94.4%	93.2%	91.2%	89.5%	89.2%	85.5%	84.5%	80.2%	77.8%	75.0%	72.0%	71.3%	71.3%
#	827	748	669	582	513	447	398	343	266	218	176	151	119	91	68	56

US Market Release	03Dec1992
CE Approval	01Jan1993
Registered USA Implants	58,041
Estimated Active USA Implants	12,595
Fixation Type	Fixed Screw
Pace Sense Polarity	Unipolar
Steroid Indicator	None

US Returned Product Analysis

Conductor Fracture	35
Insulation Breach	2
Crimp/Weld/Bond	0
Other	1

US Acute Lead Observations

Cardiac Perforation	3
Extra Cardiac Stimulation	6
Failure to Capture	126
Failure to Sense	4
Impedance Out of Range	15
Lead Dislodgement	4
Oversensing	2
Unspecified Clinical Failure	1

Product Surveillance Registry Results


Number of Leads Enrolled in Study	475
Cumulative Months of Follow-Up	18,155
Number of Leads Active in Study	44

Qualifying Complications

Conductor Fracture	5
Extra Cardiac Stimulation	1
Failure to Capture	21
Failure to Sense	2

38

Impedance Out of Range	3
Lead Dislodgement	3
Oversensing	2
Other	1



	Months After Implant							
Years	1	2	3	4	5	6	7	at 90 mo
%	95.0%	91.9%	90.3%	90.3%	89.0%	87.4%	86.4%	85.1%
#	237	192	162	145	123	98	79	64

US Market Release	10Sep1998
CE Approval	15Apr1997
Registered USA Implants	9,662
Estimated Active USA Implants	2,200
Fixation Type	Tines
Pace Sense Polarity	Quadripolar
Steroid Indicator	Yes

US Returned Product Analysis

Conductor Fracture	8
Insulation Breach	3
Crimp/Weld/Bond	0
Other	0

US Acute Lead Observations


Extra Cardiac Stimulation	1
Failure to Capture	3
Failure to Sense	2
Lead Dislodgement	7
Oversensing	2

Product Surveillance Registry Results

Number of Leads Enrolled in Study	570
Cumulative Months of Follow-Up	15,952
Number of Leads Active in Study	2

Qualifying Complications

8	
Conductor Fracture	3
Failure to Capture	2
Failure to Sense	3



Years	1	2	3	4	5	6	at 84 mo
%	99.7%	99.3%	99.3%	97.9%	97.1%	97.1%	94.1%
#	288	218	160	132	105	78	56

Method for Estimating Insertable Cardiac Monitor Performance

Insertable Cardiac Monitor (ICM) Performance Analysis

The Reveal LINQ™ and LINQ II™ ICMs are small, leadless devices that are inserted under the skin, in the chest that records subcutaneous ECG. These ICMs can be subject to malfunctions, similar to other implanted devices.

The performance report information is determined from the analysis of available complaint and available CareLink™ network data. An ICM model will be included in this report when it has accumulated at least 10,000 implant months and will remain in the report as long as at least 500 devices remain active in the CareLink™ population.

Using returned product data and CareLink™ to Estimate Insertable Cardiac Monitor Performance

ICMs returned to Medtronic are analyzed to determine whether or not they meet performance limits established by Medtronic. Although returned product analyses are valuable for gaining insight into failure mechanisms, this data can be limited in determining the survival probability as not all ICMs are returned to Medtronic for analysis. As ICMs are diagnostic devices, it is possible for a device not to be returned after meeting the device designated longevity or the patient receiving a diagnosis of their condition.

For certain malfunctions relating to oversensing for LINQ II™, CareLink™ Network data is leveraged. This data is related to FA1368: LINQ II™ Insertable Cardiac Monitoring Systems (LNQ22) with Potential for Amplified Noise.

Qualifying Complication or Malfunctions

All reported ICM complaints are adjudicated by subject matter experts for inclusion as a product performance event given available information. These product performance events are then cross-referenced with the CareLink™ population for inclusion in the survival analysis.

Product Performance events include, but are not limited to:

- Amplified Noise due to moisture (FA1368) – this only affects LINQ II™
- Premature Battery Depletion
- Electrical Component
- Software/Firmware
- Other

The CareLink™ Network

As noted previously, the CareLink™ Network is leveraged for data related to FA1368, in addition to determining the inclusion of product performance events in the survival analysis.

Statistical and Data Analysis Methods

The performance is expressed in terms of device survival estimates, where "survival" refers to the function of the device, not the survival of the patient. These survival estimates are intended to illustrate the probability that a device will survive for a given number of years without a chronic device-related complication.

Method for Estimating Insertable Cardiac Monitor Performance continued

Active surveillance normally begins at the time of implant and continues until a product performance or censoring event occurs. Of the several different statistical methods available for survival analysis, PPR survival analysis is estimated using the Standard Actuarial Method, with suspensions assumed distributed evenly within the intervals (Cutler-Ederer Method), and incorporated data from these retrospectively enrolled devices. Thus, in some cases sample sizes may fluctuate from one time interval to the next interval.

The survival estimate is the probability that a device is free of a product performance event at a given time point. For example, if a survival probability is 95% after 5 years of service, then the device has a 5% chance of experiencing a related complication in the first 5 years following implant.


Since the survival estimate can become very imprecise with small effective sample sizes, Medtronic truncates the survival curve when the effective sample size is less than 100 devices. The survival charts in the Product Performance Report show the effective sample size for each year interval where we have experience. When the effective sample size reaches 100, the next data point is added to the survival curve.

Because the information pulled from the CareLink™ network allows for assessment of all devices that were taken out of service there are no adjustments done for underreporting of malfunctions.

Definition of Analysis Dataset

To be included in the survival analysis dataset, the product must have been successfully implanted and on the CareLink™ network for at least 30 days.

US Market Release	03Jul2020	CareLink Population		Qualifying Malfunctions/Complications	
CE Approval Date	05Nov2019	Enrolled	364,817	Amplified Noise due to moisture (FA1368)	1,204
Serial Number Prefix	RLB	Active	287,480	Electrical Component	8
Mass	3.4 g	Cumulative Follow-Up Months	6,319,375	Other	27
Volume	1.4 cc			Premature Battery Depletion	245
Estimated Longevity	4.5 years			Software/Firmware	95



Years	1	2	3	4	at 56 mo
Freedom From Malfunction	99.6%	99.5%	99.4%	99.4%	99.3%
Effective Sample Size	218212	108811	39949	7902	124

ICD and CRT-D Charge Time Performance

Medtronic continues its commitment to providing updated information on charge time performance.

Introduction

Information on charge time performance of Medtronic products is presented in this section of the CRM Product Performance Report. Medtronic implemented the collection of charge time data on July 1, 1999. The data are collected via our ongoing active clinical study of long-term system performance called the Product Surveillance Registry. The study protocol requests device data be routinely taken and sent to Medtronic at no more than 6-month intervals.

In our analysis performed for this report, only charge times resulting from full energy charges are considered. To ensure consistent reporting across devices, the charge time reported at implant represents the last charge time available from date of implant. When more than one charge time is available in a 6-month interval, a conservative approach has been adopted whereby only the maximum charge time in each 6-month interval is reported. As charge time is directly proportional to the time elapsed since the last capacitor reformation, charges occurring within 15 days of a previous charge are excluded. This precludes the reporting of overly optimistic results.


Data from over 20,000 devices contribute to the charge time data in this report. By tracking and reporting this charge time data, Medtronic is able to ascertain the actual performance of its charging circuitry. The insight gained through this information is applied to Medtronic's ongoing efforts to provide charge times that are short and consistent over the life of the product.

Charge time data for ICD and CRT-D models are presented using boxplots at 6-month intervals. The shaded box on the plots represents the middle half of the data – the Interquartile Range (IQR). The white line in the middle of each box is the median charge time. The top of the box representing the IQR is the third quartile or the 75th percentile (i.e., 75% of all charge times fall below this line), whereas the bottom of the box represents the first quartile or the 25th percentile. Vertical lines are drawn from the quartiles to the farthest value not more than 1.5 times the interquartile range. Any values more extreme than the vertical lines are considered outliers.

ICD and CRT-D Charge Time Performance


D204VRM, D214VRM, D224VRC, D234VRC

Model Number	Brand
D214VRM	Secura VR
D234VRC	Secura VR




D264DRG, D284DRG, D384DRx, D394DRx

Model Number	Brand
D264DRM	Maximo II DR
D284DRG	Maximo II DR
D384DRG	Cardia DR
D394DRG	Egida DR




D264TRM, D284TRK, D384TRx, D394TRx

Model Number	Brand
D284TRK	Maximo II CRT-D
D394TRG	Egida CRT-D




D264VRM, D284VRC, D384VRx, D394VRx

Model Number	Brand
D264VRM	Maximo II VR
D284VRC	Maximo II VR
D384VRG	Cardia VR
D394VRG	Egida VR




D274VRC, D294VRC

Model Number	Brand
D294VRC	Virtuoso II VR



D314DRx


Model Number	Brand
D314DRG	Protecta XT DR
D314DRM	Protecta XT DR



ICD and CRT-D Charge Time Performance


D314TRx

Model Number	Brand
D314TRG	Protecta XT CRT-D




D334DRx, D364DRx

Model Number	Brand
D364DRG	Protecta DR
D364DRM	Protecta DR




D334TRx, D364TRx

Model Number	Brand
D364TRG	Protecta CRT-D
D364TRM	Protecta CRT-D




D334VRx, D364VRx

Model Number	Brand
D364VRG	Protecta VR
D364VRM	Protecta VR




D354DRx

Model Number	Brand
D354DRG	Protecta XT DR
D354DRM	Protecta XT DR



D354TRx


Model Number	Brand
D354TRG	Protecta XT CRT-D
D354TRM	Protecta XT CRT-D



ICD and CRT-D Charge Time Performance


D354VRx

Model Number	Brand
D354VRG	Protecta XT VR
D354VRM	Protecta XT VR



DDxxxx, DR


Model Number	Brand
DDBB1D1	Evera XT
DDBB1D4	Evera XT
DDBB2D1	Evera XT
DDBB2D4	Evera XT
DDBC3D1	Evera S
DDBC3D4	Evera S
DDMB1D1	Evera MRI XT
DDMB1D4	Evera MRI XT
DDMB2D1	Evera MRI XT
DDMB2D4	Evera MRI XT
DDMC3D1	Evera MRI S
DDMC3D4	Evera MRI
DDMD3D1	Primo
DDMD3D4	Primo
DDME3D1	Mirro
DDME3D4	Mirro



ICD and CRT-D Charge Time Performance


DTxxxxx, CRT-D

Model Number	Brand
DTBA1D1	Viva XT
DTBA1D4	Viva XT
DTBA1Q1	Viva Quad XT
DTBA1QQ	Viva Quad XT
DTBA2D1	Viva XT
DTBA2D4	Viva XT
DTBA2Q1	Viva Quad XT
DTBA2QQ	Viva Quad XT
DTBB1D1	Viva S
DTBB1D4	Viva S
DTBB1Q1	Viva Quad S
DTBB1QQ	Viva Quad S
DTBB2D1	Viva S
DTBB2D4	Viva S
DTBB2QQ	Viva Quad S
DTBC2D1	Brava
DTBC2D4	Brava
DTBC2Q1	Brava Quad
DTBC2QQ	Brava Quad
DTBX1QQ	Viva Quad C
DTBX2QQ	Viva Quad C
DTMA1D1	Claria MRI
DTMA1D4	Claria MRI
DTMA1Q1	Claria MRI
DTMA1QQ	Claria MRI
DTMA2D1	Claria MRI
DTMA2D4	Claria MRI
DTMA2Q1	Claria MRI
DTMA2QQ	Claria MRI
DTMB1D1	Amplia MRI
DTMB1D4	Amplia MRI
DTMB1Q1	Amplia MRI
DTMB1QQ	Amplia MRI
DTMB2D1	Amplia MRI
DTMB2D4	Amplia MRI
DTMB2Q1	Amplia MRI
DTMB2QQ	Amplia MRI
DTMC1D1	Compia MRI
DTMC1QQ	Compia MRI
DTMC2D1	Compia MRI
DTMC2D4	Compia MRI
DTMC2QQ	Compia MRI



DVEA3E4, DVEX2E4, DVEX3E4


Model Number	Brand
DVEA3E4	Aurora EV-ICD



ICD and CRT-D Charge Time Performance

DVxxxxx, VR

Model Number	Brand
DVAB1D1	Visia AF
DVAB1D4	Visia AF
DVAB2D1	Visia AF XT
DVAC3D1	Visia AF S
DVBB1D1	Evera XT
DVBB1D4	Evera XT
DVBB2D1	Evera XT
DVBB2D4	Evera XT
DVBC3D1	Evera S
DVBC3D4	Evera S
DVFB1D1	Visia MRI AF
DVFB1D4	Visia MRI AF
DVFB2D1	Visia MRI AF XT
DVFB2D4	Visia MRI AF XT
DVFC3D1	Visia MRI AF S
DVFC3D4	Visia MRI AF S
DVMB1D4	Evera MRI XT
DVMB2D1	Evera MRI XT
DVMB2D4	Evera MRI XT
DVMC3D1	Evera MRI S
DVMC3D4	Evera MRI S
DVMD3D1	Primo
DVMD3D4	Primo
DVME3D1	Mirro
DVME3D4	Mirro



Potential for Autonomous Cursor Motion

CareLink™ 2090 Programmer

Original Date of Communication: July 2024

STATUS UPDATE – APRIL 2025

As of 18 April 2025, Medtronic has 540 reports of autonomous cursor behavior including reports identified during the software update. There are no additional reports of unintended therapy delivered. There have been no reports of permanent harm or death associated with this behavior.

ORIGINAL COMMUNICATION - JULY 2024

Medtronic CareLink™ 2090 programmers with serial number prefixes PKK0 and PKK1 have the potential for autonomous cursor motion when Finger Touchscreen capability is enabled by software version 3.2 or higher. Through 11 June 2024, Medtronic has received 23 reports of autonomous cursor behavior, with 2 reports of unintended therapy delivered when the programmer was not under the control of trained personnel during a patient session. There have been no reports of permanent harm or death associated with this behavior.

If a programmer is unattended while in an active patient device session, a risk to patients may exist if an autonomous cursor motion engages in unintended programming. Medtronic estimates that 1.0% of Model 2090 programmers with serial number prefixes PKK0 or PKK1 could display this behavior when updated to software version 3.2 or higher.

INSTRUMENT MANAGEMENT RECOMMENDATIONS:

Software updates are necessary to maintain proper programmer function. Medtronic representatives will assist in performing the software update on all Medtronic CareLink™ 2090 programmers and assess proper function. Medtronic representatives will assist with returning programmers needing repair or replacement.

LINQ II ICM Potential for Amplified Noise June 2024

LINQ II™ Insertable Cardiac Monitoring Systems

Original Date of Communication: June 2024

STATUS UPDATE – APRIL 2025

As of 30 March 2025, Medtronic has identified 1,204 (1.86%) devices that have exhibited these characteristics, with zero (0) reports of serious harm. Medtronic estimates the projected rate of this issue to be 3.4% at 2 years or 8.7% at 4.5 years for the identified subset, and patient management recommendations are unchanged.

ORIGINAL COMMUNICATION – JUNE 2024

In November 2023, Medtronic communicated that a specific subset of LINQ II insertable cardiac monitors (ICMs) underwent a manufacturing process that may allow for moisture to impact electrode performance and create the potential for amplified noise and/or overall signal reduction of the ICM. This noise pattern is different from occasional noise due to device position/migration, patient activity, or external electromagnetic interference.

During continued investigation, Medtronic identified additional devices that have the potential for amplified noise. The identified subset now includes 64,700 total devices. Based on CareLink analysis and reported complaints as of 01 May 2024, 553 (0.85%) devices have exhibited these characteristics, with zero (0) reports of serious harm. Medtronic estimates the projected rate of this issue to be 2.9% at 2 years or 6.2% at 4.5 years for the identified subset. If an amplified noise pattern occurs, potential harms include missed/delayed diagnosis, delayed medical intervention, and early device replacement. **Medtronic recently implemented manufacturing changes to address this issue.** Overall LINQ II freedom from malfunction, including this issue, is projected to be 98.51% at 4.5 years.

PATIENT MANAGEMENT RECOMMENDATIONS:

In consultation with our Independent Physician Quality Panel (IPQP), Medtronic emphasizes continued care of patients implanted with an ICM per the existing device labeling. These recommendations are reflective of the November 2023 communication.

- Please encourage enrollment in and regular transmissions to CareLink.
 - Medtronic will continue to apply recurring algorithmic searches on CareLink for the specific amplified noise pattern and notify the clinician if present. No further action is required for patients regularly transmitting to CareLink.
- For patients not followed in CareLink:
 - Consider whether enrolling in CareLink is an option, per HRS/EHRA/APHRS/LAHR guidance.¹ CareLink monitoring will reduce the potential for episodes caused by amplified noise to overwrite true episodes before they are reviewed.

Customer Communications

- If noise interferes with the ability to assess the patient's rhythm or reason for monitoring, contact Medtronic Technical Services for assistance (dxhelp@medtronic.com or 1-800-929-4043).
- If the ICM is no longer in use, no further action is necessary.

¹Ferrick A, et. al. (2023). 2023 HRS/EHRA/APHRS/LAHR expert consensus statement on practical management of the remote device clinic. *Heart Rhythm*, 20(9), e92-e144.

SOFTWARE UPDATE AVAILABLE - Increased Potential for Reduced Energy or No Energy Delivered During HV Therapy When Programmed AX>B (2023)

Claria, Amplia, Compia, Viva, Brava, Visia AF, Evera, Primo, Mirro ICDs/CRT-Ds

Original Date of Communication: May 2023


Devices managed with an updated programming system are no longer in scope of the May 2023 communication.

STATUS UPDATE – APRIL 2025

As of 14 April 2025, Medtronic has identified 40 devices (representing 0.0033% of devices distributed worldwide) that experienced the issue described in the communication. No permanent harms or deaths have been attributed to this issue. When programmed B>AX, the occurrence rate remains in line with the historic non-advisory population as described in the original communication. To support the recommendations in the May 2023 communication, Medtronic reviewed more than 680,000 shock events that occurred in over 228,000 unique devices.

Medtronic has released a software update that aligns programming systems with the programming recommendations in most geographies. Devices managed with an updated programming system are no longer in scope of the May 2023 communication.

The overall reliability of these devices remains high. The graph below shows device survival probability inclusive of all confirmed malfunctions.



ORIGINAL COMMUNICATION - MAY 2023

This communication advises physicians of a rare potential for reduced- or no-energy output during high voltage (HV) therapy (typically 0-12J) in implantable cardioverter defibrillators (ICDs) and cardiac resynchronization therapy defibrillators (CRT-Ds) manufactured with a specific (glassed) feedthrough, including currently available ICDs and CRT-Ds. Through 10 April 2023, Medtronic has identified 27 devices from approximately 816,000 devices (representing 0.003%) distributed worldwide that have experienced a reduced- or no-energy HV therapy due to the issue described in this letter. There have been no deaths due to this issue in the glassed feedthrough device population.

With current field programming, devices with a glassed feedthrough may experience increased potential (projected to be 0.02% at 5 years) for reduced- or no-energy output during HV therapy. In some cases, a persistent 50% drop in all pacing lead impedances may be displayed; lead function, however, is not affected. See Issue Details below for further information on root cause, projected rates, risks, and potential harms.

A broader analysis was conducted to determine the incidence of device related reduced- or no-energy HV therapy events outside of the above population, with implants dating back to 2012. Using these historical observed events, projected rates for this population of devices is 0.002% at five (5) years and 0.006% at nine (9) years. This analysis identified two deaths from the historical population in which there was evidence suggesting a device-related reduced- or no-energy HV therapy occurred.

Should this issue occur, pacing, sensing, episode detection, anti-tachycardia pacing (ATP) therapies, battery longevity, and telemetry remain functional. **When devices in the glassed feedthrough population are programmed**

exclusively in the B>AX configuration, the potential for a reduced- or no-energy HV therapy is 0.002% at five (5) years and 0.005% at nine (9) years, comparable to historical device performance.

PATIENT MANAGEMENT RECOMMENDATIONS:

Medtronic recognizes that each patient requires unique clinical considerations. Based on internal investigation and external consultation with our Independent Physician Quality Panel (IPQP), Medtronic provides the following guidance:

- **Prophylactic device replacement is NOT recommended.**
 - The risk of mortality for patients after reprogramming is 0.001% at 9 years and is less than the risk of patient mortality due to complications associated with device replacement (0.032% - 0.043%^{1,2,3}).
- **Program all HV therapy pathways B>AX in all therapy zones to minimize the risk for this issue.**
 - Note: Using "Get Medtronic Nominals" **will require manual reprogramming of Rx5 and Rx6 to B>AX** for all ventricular therapies.
 - For patients remotely followed via CareLink, your Medtronic representative will provide a report to assist with identifying patients who may have one or more HV therapy pathways programmed AX>B. You may contact your local representative to obtain an updated copy of the report at any time.
- **Prioritize reprogramming patients who have both a history of HV therapy and Rx1 programmed AX>B.**
 - Rx1 provides the greatest statistical likelihood to resolve an arrhythmia, and therefore it is important to minimize risk of a reduced- or no-energy HV therapy in the first sequence.
- **For remaining patients with AX>B programming in any HV therapy sequence, schedule the next follow-up for in-clinic reprogramming, with the appropriate discretion,** to minimize potential for reduced- or no-energy HV therapies to occur.
- Per standard practice, check tachyarrhythmia episodes to determine effectiveness of therapies that have been delivered.
 - Instruct patients to contact the clinic if they receive HV therapy or hear an audible tone coming from their device.
 - Verify delivered energy is consistent with programmed energy in the Episode Summary. Note: The image below highlights an episode experiencing intermittent reduced HV therapies (red boxes).

Customer Communications

Episode Summary				
Initial Type	VF (spontaneous)			
Duration	27 sec			
A/V Max Rate	Unknown/231 bpm			
V. Median	231 bpm (260 ms)			
Activity at onset	Active, Sensor = 118 bpm			
Last Therapy	VF Rx3: Defib, Successful			
Therapies	Delivered	Charge	Obms	Energy
VF Rx 1 Burst	During Charging			
VF Rx 1 Defib	0.7 J	9.97 sec	<20 ohms	0.0 35 J
VF Rx 2 Defib	0.3 J	0.77 sec	<20 ohms	33 35 J
VF Rx 3 Defib	35.7 J	0.49 sec	93 ohms	33 35 J
Termination				

Example screen illustrating reduced-energy shocks for an Evera XT DR device.

- After an SCP event, pacing, sensing, episode detection, and anti-tachycardia pacing (ATP) therapies are not impacted; additionally, HV charging, battery longevity and Bluetooth telemetry are not impacted.
- **Contact Medtronic Technical Services (1-800-929-4043) or your local representative if one of the following is observed as these may be an indication of either a device or lead-related issue:**
 - Reduced- or no-energy HV therapy is displayed in Episode Text (regardless of programmed pathway)
 - A persistent drop of approximately 50% in RA, RV and LV pacing lead impedance measurements as this may be an indication of increased potential for a future reduced- or no-energy therapy.

ISSUE DETAILS:

There is an increased potential for a reduced- or no-energy HV therapy in the AX>B configuration when all the following conditions are met:

- The device has a glassed feedthrough (manufactured after July 2017).
- There is significant separation of the layers of insulation materials in the feedthrough components of the device header.
- An unintended current pathway forms within the void created by the insulation separation, capable of conducting high levels of current during HV therapy.

When an unintended current pathway is detected during HV therapy, the Short Circuit Protection (SCP) feature may trigger (see Appendix A). This behavior can be intermittent; both full-energy and reduced-energy HV therapies within the same episode have been observed. SCP events may also be lead related; for both lead-related and device-related unintended current pathways, the defibrillation waveform is truncated early in the energy delivery sequence, resulting in reduced or no energy being delivered (~0-12J).

RATES OF OCCURRENCE FOR DEVICE-RELATED REDUCED- or NO-ENERGY HV THERAPY:

Through 10 April 2023, 27 devices with glassed feedthroughs have been identified as experiencing a reduced or no-energy HV therapy (0.003% out of 816,000 distributed devices). Of these, 26 were in devices with an AX>B delivered pathway. Based on an analysis of patients with a glassed feedthrough device and with a history of HV therapy, the observed rate for this issue is 0.03%. See Table 1 below for projected rates of occurrence and risk for harm.

Potential harms related to reduced- or no-energy HV therapy include failure to terminate an arrhythmia, which could lead to death, as well as complications associated with device replacement and/or unnecessary lead replacement if the reduced- or no-energy HV therapy is erroneously attributed to a lead failure.

TABLE 1: Comparison of projected event rates and risk of catastrophic harm, with and without reprogramming

Population		Projected event rate (one or more reduced- or no-energy HV therapies in an episode)	Mortality risk of this issue, considering the probability a sequence of six HV therapies fails to terminate an arrhythmia
Glassed feedthrough devices with current field programming (~816,000 devices)	Overall Population of Devices	0.02% @ 5 years*	0.004% @ 5 years*
	For Patients with History of HV Therapy	0.48% @ 5 years*	0.08% @ 5 years*
Glassed feedthrough devices after all HV pathways reprogrammed B>AX	Overall Population of Devices	0.005% @ 9 years**	0.001% @ 9 years**
	For Patients with History of HV Therapy	0.04% @ 9 years**	0.01% @ 9 years**
Historical devices dating back to 2012 (~651,000 devices)	Overall Population of Devices	0.006% @ 9 years**	0.001% @ 9 years**
	For Patients with History of HV Therapy	0.05% @ 9 years**	0.01% @ 9 years**

* A timeframe of 5 years is used for this projection given the average implant duration of the devices in scope of this communication is approximately 4 years and allowing for adequate time for reprogramming.

** A time frame of 9 years is used based on the weighted average for device longevity for ICDs and CRTDs.

Medtronic is updating Instructions for Use, HV therapy nominals and the programmer interfaces for these device models to be consistent with information in this letter. Medtronic will release additional information once the necessary regulatory approvals have been received, as applicable in the local region.

APPENDIX A – ADDITIONAL DETAILS ON SHORT CIRCUIT PROTECTION (SCP)

Short Circuit Protection (SCP) is a safety feature that can only occur during high voltage (HV) therapy. SCP is designed to truncate energy delivery to protect the device when an unintended current pathway is detected during a shock. An SCP event can occur when an unintended current pathway develops either in the lead or in the device.

Contact Medtronic for additional guidance if you believe an SCP event occurred.

Programming all high voltage delivery pathways to B>AX will minimize the effect of an unintended current pathway in the device header (e.g., feedthrough), thereby minimizing the potential for an SCP event. Comparing the programmed energy to the delivered energy in the Episode Text for a treated episode can be used to identify SCP events. The lead impedance in the Episode Text will also display <20 ohms. Refer to the sections below on how to identify if an SCP event has occurred.

For Evera/Visia AF/Primo/Mirro ICDs and Viva/Brava/Claria/Amplia/Compia CRT-D devices:

Identifying SCP events: For episodes in which a high voltage therapy was delivered, if there is a significantly reduced-energy or no-energy shock delivered, the device may have experienced an SCP event. In each stored episode, delivered energy is displayed in the Episode Text and on the Interval Plot. If an SCP event occurred, this text may indicate that a lower energy was delivered and <20 ohm impedance value. Note: there is currently no available SCP alert in this family of devices.

Clinicians can consider enabling the device audible alert and/or CareAlert for "Number of Shocks Delivered in an Episode," with "N = 1." Turning this alert "ON" may help ensure the patient and/or clinic is made aware when a HV therapy is delivered by the device.

If an SCP event occurs during a manual shock delivery, there will be no episode data. For manual shock deliveries, review the shock impedance and the delivered energy from the Last High Voltage Therapy section on the Battery and Lead Measurement screen of the programmer.

¹Tarakji KG, et al. Antibacterial Envelope to Prevent Cardiac Implantable Device Infection. The New England Journal of Medicine. 2019; 380(20):1895-1905.

²Medtronic Data on File. MDT2260884-CRHF CIED Infection Report; Agile: MDT2260884, Version 2.0, 11/02/2015.

³Birnie D, et al. Complications associated with defibrillation threshold testing: The Canadian experience. Heart Rhythm. 2008; 5(3):387-90.

SOFTWARE UPDATE AVAILABLE - Increased Potential for Reduced Energy or No Energy Delivered During HV Therapy When Programmed AX>B (2023)

Cobalt™ XT, Cobalt™ and Crome™ ICDs and CRT-Ds

Original Date of Communication: May 2023


Devices managed with an updated programming system are no longer in scope of the May 2023 communication.

STATUS UPDATE - APRIL 2025

As of 14 April 2025, Medtronic has identified 40 devices (representing 0.0033% of devices distributed worldwide) that experienced the issue described in the communication. No permanent harms or deaths have been attributed to this issue. When programmed B>AX, the occurrence rate remains in line with the historic non-advisory population as described in the original communication. To support the patient management recommendations in the May 2023 communication, Medtronic reviewed more than 680,000 shock events that occurred in over 228,000 unique devices.

Medtronic has released a software update that aligns programming systems with the programming recommendations in most geographies. Devices managed with an updated programming system are no longer in scope of the May 2023 communication.

The overall reliability of these devices remains high. The graph below shows device survival probability inclusive of all confirmed malfunctions.



ORIGINAL COMMUNICATION - MAY 2023

This communication advises physicians of a rare potential for reduced- or no-energy output during high voltage (HV) therapy (typically 0-12J) in implantable cardioverter defibrillators (ICDs) and cardiac resynchronization therapy defibrillators (CRT-Ds) manufactured with a specific (glassed) feedthrough, including currently available ICDs and CRT-Ds. Through 10 April 2023, Medtronic has identified 27 devices from approximately 816,000 devices (representing 0.003%) distributed worldwide that have experienced a reduced- or no-energy HV therapy due to the issue described in this letter. There have been no deaths due to this issue in the glassed feedthrough device population.

With current field programming, devices with a glassed feedthrough may experience increased potential (projected to be 0.02% at 5 years) for reduced- or no-energy output during HV therapy. In some cases, a persistent 50% drop in all pacing lead impedances may be displayed; lead function, however, is not affected. See Issue Details below for further information on root cause, projected rates, risks, and potential harms.

A broader analysis was conducted to determine the incidence of device related reduced- or no-energy HV therapy events outside of the above population, with implants dating back to 2012. Using these historical observed events, projected rates for this population of devices is 0.002% at five (5) years and 0.006% at nine (9) years. This analysis identified two deaths from the historical population in which there was evidence suggesting a device-related reduced- or no-energy HV therapy occurred.

Should this issue occur, pacing, sensing, episode detection, anti-tachycardia pacing (ATP) therapies, battery longevity, and telemetry remain functional. **When devices in the glassed feedthrough population are programmed**

exclusively in the B>AX configuration, the potential for a reduced- or no-energy HV therapy is 0.002% at five (5) years and 0.005% at nine (9) years, comparable to historical device performance.

PATIENT MANAGEMENT RECOMMENDATIONS:

Medtronic recognizes that each patient requires unique clinical considerations. Based on internal investigation and external consultation with our Independent Physician Quality Panel (IPQP), Medtronic provides the following guidance:

- **Prophylactic device replacement is NOT recommended.**
 - The risk of mortality for patients after reprogramming is 0.001% at 9 years and is less than the risk of patient mortality due to complications associated with device replacement (0.032% - 0.043%^{1,2,3}).
- **Program all HV therapy pathways B>AX in all therapy zones to minimize the risk for this issue.**
 - Note: Using "Get Medtronic Nominals" **will require manual reprogramming of Rx5 and Rx6 to B>AX** for all ventricular therapies.
 - For patients remotely followed via CareLink, your Medtronic representative will provide a report to assist with identifying patients who may have one or more HV therapy pathways programmed AX>B. You may contact your local representative to obtain an updated copy of the report at any time.
- **Prioritize reprogramming patients who have both a history of HV therapy and Rx1 programmed AX>B.**
 - Rx1 provides the greatest statistical likelihood to resolve an arrhythmia, and therefore it is important to minimize risk of a reduced- or no-energy HV therapy in the first sequence.
- **For remaining patients with AX>B programming in any HV therapy sequence, schedule the next follow-up for in-clinic reprogramming, with the appropriate discretion,** to minimize potential for reduced- or no-energy HV therapies to occur.
- Per standard practice, check tachyarrhythmia episodes to determine effectiveness of therapies that have been delivered.
 - Instruct patients to contact the clinic if they receive HV therapy or hear an audible tone coming from their device.
 - Verify delivered energy is consistent with programmed energy in the Episode Summary. Note: The image below highlights an episode experiencing intermittent reduced HV therapies (red boxes).

Customer Communications

Episode Summary				
Initial Type	VF (spontaneous)			
Duration	27 sec			
A/V Max Rate	Unknown/231 bpm			
V. Median	231 bpm (260 ms)			
Activity at onset	Active, Sensor = 118 bpm			
Last Therapy	VF Rx3: Defib, Successful			
Therapies	Delivered	Charge	Obms	Energy
VF Rx 1 Burst	During Charging			
VF Rx 1 Defib	0.7 J	9.97 sec	<20 ohms	0.0 35 J
VF Rx 2 Defib	0.3 J	0.77 sec	<20 ohms	33 35 J
VF Rx 3 Defib	35.7 J	0.49 sec	93 ohms	33 35 J
Termination				

Example screen illustrating reduced-energy shocks for an Evera XT DR device.

- After an SCP event, pacing, sensing, episode detection, and anti-tachycardia pacing (ATP) therapies are not impacted; additionally, HV charging, battery longevity and Bluetooth telemetry are not impacted.
- **Contact Medtronic Technical Services (1-800-929-4043) or your local representative if one of the following is observed as these may be an indication of either a device or lead-related issue:**
 - Reduced- or no-energy HV therapy is displayed in Episode Text (regardless of programmed pathway)
 - A persistent drop of approximately 50% in RA, RV and LV pacing lead impedance measurements as this may be an indication of increased potential for a future reduced- or no-energy therapy.

ISSUE DETAILS:

There is an increased potential for a reduced- or no-energy HV therapy in the AX>B configuration when all the following conditions are met:

- The device has a glassed feedthrough (manufactured after July 2017).
- There is significant separation of the layers of insulation materials in the feedthrough components of the device header.
- An unintended current pathway forms within the void created by the insulation separation, capable of conducting high levels of current during HV therapy.

When an unintended current pathway is detected during HV therapy, the Short Circuit Protection (SCP) feature may trigger (see Appendix A). This behavior can be intermittent; both full-energy and reduced-energy HV therapies within the same episode have been observed. SCP events may also be lead related; for both lead-related and device-related unintended current pathways, the defibrillation waveform is truncated early in the energy delivery sequence, resulting in reduced or no energy being delivered (~0-12J).

RATES OF OCCURRENCE FOR DEVICE-RELATED REDUCED- or NO-ENERGY HV THERAPY:

Through 10 April 2023, 27 devices with glassed feedthroughs have been identified as experiencing a reduced or no-energy HV therapy (0.003% out of 816,000 distributed devices). Of these, 26 were in devices with an AX>B delivered pathway. Based on an analysis of patients with a glassed feedthrough device and with a history of HV therapy, the observed rate for this issue is 0.03%. See Table 1 below for projected rates of occurrence and risk for harm.

Potential harms related to reduced- or no-energy HV therapy include failure to terminate an arrhythmia, which could lead to death, as well as complications associated with device replacement and/or unnecessary lead replacement if the reduced- or no-energy HV therapy is erroneously attributed to a lead failure.

TABLE 1: Comparison of projected event rates and risk of catastrophic harm, with and without reprogramming

Population		Projected event rate (one or more reduced- or no-energy HV therapies in an episode)	Mortality risk of this issue, considering the probability a sequence of six HV therapies fails to terminate an arrhythmia
Glassed feedthrough devices with current field programming (~816,000 devices)	Overall Population of Devices	0.02% @ 5 years*	0.004% @ 5 years*
	For Patients with History of HV Therapy	0.48% @ 5 years*	0.08% @ 5 years*
Glassed feedthrough devices after all HV pathways reprogrammed B>AX	Overall Population of Devices	0.005% @ 9 years**	0.001% @ 9 years**
	For Patients with History of HV Therapy	0.04% @ 9 years**	0.01% @ 9 years**
Historical devices dating back to 2012 (~651,000 devices)	Overall Population of Devices	0.006% @ 9 years**	0.001% @ 9 years**
	For Patients with History of HV Therapy	0.05% @ 9 years**	0.01% @ 9 years**

* A timeframe of 5 years is used for this projection given the average implant duration of the devices in scope of this communication is approximately 4 years and allowing for adequate time for reprogramming.

** A time frame of 9 years is used based on the weighted average for device longevity for ICDs and CRTDs.

Medtronic is updating Instructions for Use, HV therapy nominals and the programmer interfaces for these device models to be consistent with information in this letter. Medtronic will release additional information once the necessary regulatory approvals have been received, as applicable in the local region.

APPENDIX A – ADDITIONAL DETAILS ON SHORT CIRCUIT PROTECTION (SCP)

Short Circuit Protection (SCP) is a safety feature that can only occur during high voltage (HV) therapy. SCP is designed to truncate energy delivery to protect the device when an unintended current pathway is detected during a shock. An SCP event can occur when an unintended current pathway develops either in the lead or in the device.

Contact Medtronic for additional guidance if you believe an SCP event occurred.

Programming all high voltage delivery pathways to B>AX will minimize the effect of an unintended current pathway in the device header (e.g., feedthrough), thereby minimizing the potential for an SCP event. Comparing the programmed energy to the delivered energy in the Episode Text for a treated episode can be used to identify SCP events. The lead impedance in the Episode Text will also display <20 ohms. Refer to the sections below on how to identify if an SCP event has occurred.

For Cobalt/Crome ICD and CRT-D devices:

Identifying SCP events during high voltage (HV) therapy delivery: When an SCP event has occurred, the device will issue an RV Defibrillation Lead Impedance CareAlert. Audible and wireless alerts are issued, if enabled. These devices are shipped with alerts enabled and are nominally active once implant detection completes.

A CareAlert will occur simultaneous with HV therapy delivery and will specifically report "RV Defib Lead Impedance 0 Ω " with the same time stamp as the therapy delivery. The alert condition is displayed in the CareAlert Events log (Data >> CareAlert Events). The Quick Look Observations will also display "Alert: RV defib lead impedance warning on Mmm/DD/YYYY." SCP events are not reflected in the long-term lead impedance trends.

Following an SCP event in devices with both the SVC Coil and Active Can enabled, the device will turn off the SVC coil. If the SCP event is caused by a lead integrity issue involving the SVC coil, turning off the SVC coil can allow any subsequent shocks to be delivered using the RV Coil -to- Active Can vector. The programmed therapy parameters will indicate the SVC Coil has been automatically disabled. The SVC coil will remain off until it is turned back on by the clinician.

If an SCP event occurs during a manual shock delivery, there will be no episode data. For manual shock deliveries, review the shock impedance and the delivered energy from the Last High Voltage Therapy section on the Battery and Lead Measurement screen of the programmer.

¹Tarakji KG, et al. Antibacterial Envelope to Prevent Cardiac Implantable Device Infection. The New England Journal of Medicine. 2019; 380(20):1895-1905.

²Medtronic Data on File. MDT2260884-CRHF CIED Infection Report; Agile: MDT2260884, Version 2.0, 11/02/2015.

³Birnie D, et al. Complications associated with defibrillation threshold testing: The Canadian experience. Heart Rhythm. 2008; 5(3):387-90.

SOFTWARE UPDATE AVAILABLE - Potential for Intermittent-Reduced-Energy Shock Due To Short Circuit Protection Event (2022)

Cobalt™ XT, Cobalt™ and Crome™ ICDs and CRT-Ds

Original Date of Communication: June 2022

STATUS UPDATE - APRIL 2025

Manufacturing updates may increase device programming options. Contact Medtronic Technical Services for details.

As of 10 August 2022, a software release is now available for CareLink™ SmartSync™ Device Managers (SmartSync). Once a SmartSync tablet has been updated with software application D00U005 version 7.1.1 (or higher), the programmer will deploy a device update to Cobalt and Crome implantable cardioverter defibrillators (ICDs) and cardiac resynchronization therapy defibrillators (CRT-Ds) to prevent the potential for an intermittent, second-phase Short Circuit Protection (SCP) event during high-voltage (HV) therapy delivery. This software update was previously announced as part of an advisory communication Medtronic issued in June 2022 (see original communication posted below).

Through 14 April 2025, Medtronic has confirmed 160 devices (representing 0.08% of devices distributed worldwide) have experienced a second-phase SCP event. In all events, ~79% of the programmed shock energy was delivered. No events have occurred in devices programmed B>AX that have the August 2022 software update. No permanent harms or deaths have been directly attributed to this issue.

Medtronic representatives are available to work with clinicians to ensure all SmartSync tablets in their facility(s) are updated with application software D00U005 version 7.1.1 (or higher). The software can be installed by connecting each SmartSync tablet to the internet, opening the SmartSync App and accepting the on-screen prompts.

As disclosed in the June 2022 patient management recommendations, patients will require an in-clinic visit for the update to be installed into their device via interrogation with an updated SmartSync tablet. Once installed, the update will allow devices to deliver the full programmed shock energy. Programming B>AX pathway and Active Can enabled is still required. On-screen messaging will reinforce these programming recommendations.

Clinicians can identify if a patient's device has successfully received the update by viewing the displayed Configuration ID and confirming the first number in the sequence is as indicated below:

- 11-1-0 or higher for Cobalt/Crome VR devices
- 10-1-0 or higher for Cobalt/Crome DR and CRT-D devices

The Device Configuration ID can be found under the “Device Information” section of the SmartSync Parameters Report, or for CareLink patients, under the Transmission Details page by selecting <More Reports > ‘Parameters.’

ORIGINAL COMMUNICATION - JUNE 2022

This communication provides notice of the potential for reduced shock energy (~79% of programmed energy) during high-voltage (HV) therapy for Cobalt and Crome implantable cardioverter defibrillators (ICDs) and cardiac resynchronization therapy defibrillators (CRT-Ds). Through 03 June 2022, Medtronic has identified 27 devices (0.03% of devices distributed worldwide) that have experienced a reduced-energy shock, which is accompanied by a Short Circuit Protection (SCP) alert. Medtronic has not received any reports of permanent harm or death due to this issue. Medtronic has submitted a device software update to address this issue and anticipates it will be available for download into implanted devices <beginning third/fourth quarter of calendar year 2022>, pending regulatory approvals.

ISSUE SUMMARY:

Short Circuit Protection (SCP) alerts trigger during HV therapy during the first- or second-phase of the HV biphasic waveform delivery. This communication focuses on second-phase SCP events that are the result of a secondary, low-level current pathway detected in the HV circuitry.

- A second-phase SCP event **will deliver approximately 79%** of programmed energy as a monophasic waveform.
- **Defibrillation efficacy is reduced by ~1%** for this type of SCP event when HV therapy is programmed to 40J, considering cumulative success across the full series of shocks (Rx1 through Rx6).

Based on analysis of peer-reviewed literature as well as CareLink data on shock efficacy from more than 279,000 episodes*, termination success rates for 32J (~79% of 40J), monophasic shocks versus 40J biphasic shocks are estimated in Table 1. Termination success may vary depending on individual patient risk factors and medication use.

TABLE 1

	Normal Operation (40J, Biphasic delivery)	Second-phase SCP (32J, Monophasic delivery)
Estimated First Shock Success* (in VF Zone)	89%	85%
Estimated Cumulative Success Shocks 1-6*	99%	98%

*Medtronic data on file; May 2022.

- While 0.03% has been observed to date, Medtronic projects 0.18%** of the ~80,000 distributed devices may experience a second-phase SCP event within 24 months of service life, when considering the probability for these SCP events increases over time, and the likelihood a patient will need HV therapy during that time.

- For the population of patients who received HV therapy, the observed rate was 0.77%. When projecting for this population, the chance of encountering a second-phase SCP event is ~5.0%** at 24 months.

**The above projections are based on calculations without the planned device software update. Once installed, this update, in addition to the programming recommendations, will resolve occurrences of second-phase SCP events.

Potential harms related to a second-phase SCP event include failure to terminate the arrhythmia due to reduced-delivered-energy, a theoretical risk of proarrhythmia, and complications associated with device replacement, including unnecessary lead replacement due to misinterpretation of the SCP alert.

- **While not observed clinically**, Medtronic estimates the risk for proarrhythmia is 0.002% in the AX>B configuration, and improbable in the B>AX configuration (less than 0.00004%), with Active Can pathway enabled. These risks may be higher when Active Can is disabled.
- The overall risk for **patient mortality due to this issue is estimated to be 0.002%** at 24 months when combining the likelihood a patient will need therapy with the probability an arrhythmia fails to terminate after six sequences of 32J monophasic shocks.
 - Comparatively, the risk of **patient mortality due to complications associated with device replacement is 0.032% - 0.043%**^{1,2,3}

PATIENT MANAGEMENT RECOMMENDATIONS AND CONSIDERATIONS:

SCP events are evident to the patient and clinician. Devices will issue an audible tone and, for patients enrolled in CareLink, a wireless CareAlert will report *RV Defib lead impedance 0 ohms*.

Medtronic recognizes that each patient requires unique clinical considerations. Based on internal investigation and external consultation with our Independent Physician Quality Panel (IPQP), Medtronic recommends:

- **Prophylactic device replacement is NOT recommended.**
- Remote monitoring with normal frequency of follow-up per clinic protocol, with patients' next follow-up scheduled in-clinic to allow for device reprogramming (if necessary):
 - **Programming all HV therapies to 40J with a B>AX pathway and Active Can/SVC Coil set with Active Can enabled across all therapy zones.**
- Contact Medtronic Technical Services (1-800-723-4636) or your local representative if an *RV Defib Lead Impedance Alert* reporting zero (0) ohms is observed – as this is an indicator that an SCP event was detected during HV therapy.
 - Importantly, if the delivered energy during the episode is ~79% of the programmed energy AND the SCP alert indicates an RV Defib Lead impedance alert reporting exactly zero (0) ohms, this is an indication of a second-phase SCP event (as described in this letter) and not a lead issue.

- Consider device replacement only after observing and confirming the cause of an SCP event with a Medtronic representative, with the understanding a device has an ~81% probability of delivering subsequent reduced-energy shocks, and with the understanding an update for implanted devices is anticipated to be available beginning <third quarter/fourth quarter> of calendar year 2022.

Note: The software update will require an additional in-clinic follow-up in order for it to be installed into a patient's device. The update will ensure the full shock energy is delivered in the presence of a secondary, low-level current pathway in the HV circuitry.

- After an SCP event, pacing, sensing, episode detection, and anti-tachycardia pacing (ATP) therapies are not impacted; additionally, HV charging, battery longevity and Bluetooth telemetry are not impacted.

¹ Tarakji KG, et al. Antibacterial Envelope to Prevent Cardiac Implantable Device Infection. The New England Journal of Medicine. 2019; 380(20):1895-1905.

² Medtronic Data on File. MDT2260884-CRHF CIED Infection Report; Agile: MDT2260884, Version 2.0, 11/02/2015.

³ Birnie D, et al. Complications associated with defibrillation threshold testing: The Canadian experience. Heart Rhythm. 2008; 5(3):387-90.

Software Update Available to Correct Potential for SmartSync Telemetry Error

CareLink SmartSync™ Device Manager supporting Cobalt™ and Crome™ ICDs and CRT-Ds

Original Date of Communication: April 2022

STATUS UPDATE - APRIL 2025

As of 22 April 2025, Medtronic has received 219 reports of the SmartSync Telemetry Error in Cobalt and Crome devices. No serious adverse events or permanent harms have been reported.

ORIGINAL COMMUNICATION – APRIL 2022

Medtronic is notifying health care professionals of a **software update for CareLink SmartSync™ Device Managers** (SmartSync) that will address a telemetry error that may occur with Medtronic Cobalt™ and Crome™ implantable cardioverter defibrillators (ICDs) and cardiac resynchronization therapy defibrillators (CRT-Ds). Specifically, **software application D00U005 version 6.0.3 will deploy an update** to implanted devices that will correct the potential for temporary suspension of some device features (details below) due to a telemetry error involving inductive (non-Bluetooth) telemetry. As of 22 March 2022, 0.3% of devices have experienced this issue. No serious adverse events or permanent harms have been reported due to this error.

Medtronic representatives will work with you to ensure all SmartSync tablets in your facility are updated with application software D00U005 version 6.0.3 or higher. Once the software has been installed on a tablet, a patient's device will automatically receive an update (to prevent the telemetry error) during their next SmartSync session.

Details:

Some Cobalt and Crome devices may encounter a persistent "session-active" flag following the use of inductive telemetry. The persistent session-active flag is the result of a telemetry connection error that can occur when intermittent or disrupted signals manifest while communicating with the device at the end of the telemetry session. Inductive telemetry with a Cobalt/Crome device typically occurs during device interrogation with a CareLink Express™ Mobile reader head. A persistent session-active flag will result in temporary suspension of the following features (if available in the device) until the flag is cleared:

- Battery voltage measurements
- Capture Management™

Customer Communications

- Atrial Lead Position Check™
- AdaptivCRT™, EffectivCRT™ diagnostic, and EffectivCRT™ During AF
- Wavelet™ template management
- Battery conditioning charges

Potential risks include loss of pacing or inadequate CRT support, and/or loss of Recommended Replacement Time (RRT) indicator.

When battery measurements are suspended for more than seven days, the longevity estimator cannot calculate a value and the estimator will display a grey bar with "???" Longevity estimates will be unavailable for approximately 82 weeks. A device that experiences a persistent session-active flag can be manually cleared via a specific sequence of steps, using a non-Bluetooth SmartSync telemetry session. Contact Medtronic Technical Services at 800-723-4636 for further instruction. After the persistent flag is manually cleared, the above features will automatically be restored. Remaining longevity estimates will resume approximately 82 weeks after the date the flag is cleared. The issue is unlikely to result in clinical impact to the patient given the features listed above can be restored with an in-clinic SmartSync programmer session.

Devices manufactured after July 2021 have already received the software update and are not susceptible to the described behavior. Refer to Appendix A (below) for details on how to identify which Cobalt/Crome devices have already received the update.

Patient Management Recommendations:

We realize that each patient requires unique clinical considerations. In consultation with our Independent Physician Quality Panel (IPQP), Medtronic recommends continuing normal follow-up frequency per local clinic protocol.

Once the software is installed on a SmartSync tablet, please follow these recommendations:

- **Patients routinely seen in the clinic** will automatically receive the update during their next interrogation using an updated SmartSync tablet (D00U005 version 6.0.3 or higher). No additional programming of the device is required.
- **Patients followed remotely who do not have regularly scheduled in-clinic sessions** should have their next follow-up session conducted in clinic using an updated SmartSync tablet (D00U005 version 6.0.3 or higher). No additional programming of the device is required.

Note: If a patient's device displays a grey longevity estimator bar with "???", the device may have a persistent session-active flag. Contact Medtronic Technical Services at 800-723-4636 for assistance.

APPENDIX A

How to Confirm a Patient's Device Has Received the Update?

Each device will display a Device Configuration ID after interrogation by an updated SmartSync tablet, or after transmitting to CareLink. The Device Configuration ID can be found via the Parameters Report as noted below:

For SmartSync - the following is available from the Parameters Report PDF file.

The image shows a PDF report titled 'Parameters' from Medtronic. It contains the following sections:

- Header:** Medtronic logo, 'Parameters' title, Device: Cobalt™ XT DR DCPA2D4, Serial Number, Date of Interrogation: 13-Dec-2021 14:51:37, Patient ID, Physician.
- Additional Features:**

Rate Drop Response	Off
Sleep	Off
Non-Comp Atrial Pacing	On
NCAP Interval	300 ms
MRI SureScan	Off
PMF Intervention	On
PVC Response	On
V. Safety Pacing	On
- Device Information:**

Device	Medtronic	Cobalt XT DR DCPA2D4	RSM	Implanted: 27-Sep-2021
Atrial	Medtronic	5076 Coaptant Evolution MR1	PUN1	Implanted: 27-Sep-2021
RV/SVC	Medtronic	8513 Microprint Quattro MR0	TDK	Implanted: 27-Sep-2021
- Device Configuration ID:** 2-1-0
- Notes:**

Image: Sample SmartSync-generated Parameters Report showing updated Device Configuration ID.

For CareLink - the following is available from the Transmission Details page by selecting 'More Reports' > 'Parameters.'

The image shows a screenshot of the CareLink web interface. The 'Parameters' section is highlighted, showing the following details:

- Mode:** ODD
- Pacing Details:**

	Atrial	RV
Sensitivity	0.30 mV	0.30 mV
Sense Polarity	Bipolar	Bipolar
- Refractory/Blanking:**

PVAB Interval	150 ms
PVAB Method	Partial
A. Blank Post AS	100 ms
V. Blank Post VS	120 ms
- Additional Features:**

Rate Drop Response	Off
MRI SureScan	Off
- Device Information:**

Device	Medtronic	Cobalt DR DCPB1D1	RSN60004B	Implanted: 09-Jun-2021
--------	-----------	-------------------	-----------	------------------------
- Device Configuration ID:** 2-1-0


Image: Sample CareLink Parameters Report showing updated Device Configuration ID.

How do I update my SmartSync™ application software for the issue described in the April/May 2022 communication?

On any tablet, you can update to the most recent version for all applications resident on that tablet by simply connecting to the internet and either automatically discover if new software is available by launching the SmartSync App (see images below), OR manually discover if new software is available by navigating to the Software Information

Customer Communications

screen and perform "Check for Updates." Contact your local Medtronic representative or Medtronic Technical Services at 800-638-1991 if you need assistance.



How do I confirm if a SmartSync tablet has already been installed with the updated software?

On any tablet, you can confirm the application software version for any device family by:

1. Selecting the MENU in the upper right corner of the SmartSync App [1]
2. Selecting PROFILE [2]
3. Selecting the SOFTWARE tab and scrolling through the SOFTWARE INFO list [3]

If the software update for this issue has already been installed, you will see the following versions listed:

- The Common/Platform application version is 3.6.4 (or higher)
- is 6.0.3 (or higher)

Customer Communications



LINQ II - Brady, Pause and PVC Detections Disabled Following Electrical Reset

LINQ II™ Insertable Cardiac Monitoring Systems

Original Date of Communication: June 2021

STATUS UPDATE - APRIL 2025

Medtronic released a software update in August 2023 to address this issue in LINQ II ICMs manufactured from July 2020 to June 2021. The update is available through SmartSync App 3.12.4 or higher.

Medtronic implemented a manufacturing update to all newly manufactured LINQ II ICMs released into distribution to prevent a partial electrical reset from disabling Brady, Pause and PVC event detection.

Updated LINQ II ICMs can be identified, before being implanted, by the GTIN that is printed under the barcode on the box. The new U.S. LINQ II GTIN ends with "002."

As a reminder, unused LINQ II devices manufactured prior to June 2021 were requested to be returned to Medtronic per the Original advisory (dated June 2021) – these devices cannot be updated in the field and will continue to be susceptible to the issue.

ORIGINAL COMMUNICATION - JUNE 2021

This notice is to inform you that LINQ II insertable cardiac monitors (ICMs) that undergo a partial electrical reset appear to be programmed "ON," but are no longer able to detect and report Brady, Pause and PVC events to clinicians. Medtronic estimates that 0.21% of LINQ II ICMs have experienced a partial electrical reset resulting in the inability to detect Brady, Pause and PVC events. While there is a potential for underreporting due to lack of awareness that an electrical reset has occurred, there have been zero (0) serious or permanent harms or deaths reported as a result of this issue. After a partial electrical reset, these Brady, Pause and PVC episode types will not be reported to the clinician.

- A correction for currently implanted LINQ II ICMs is not available.
- We are requesting that hospitals quarantine all LINQ II ICMs on hospital shelves. Physicians should cease implanting any remaining LINQ II ICMs that may remain in shelf stock and return any unused product to Medtronic.
- There will be an update for future manufactured LINQ II ICMs, which is anticipated to be available in the U.S. July 2021.

This letter contains a description of the information known to date and patient management recommendations.

ISSUE DESCRIPTION

Medtronic has identified that LINQ II ICMs that undergo a partial electrical reset appear to be programmed "ON," but are no longer able to detect and report Brady, Pause, and PVC events. A partial electrical reset is normal behavior that can occur when the device detects a possible issue with the device software. However, an error in the partial electrical reset implementation is causing this unintended behavior.

All LINQ II ICM devices currently in distribution are susceptible to this issue. Through 10 May 2021, Medtronic has received 37 complaints related to an electrical reset. The projected rate of a LINQ II ICM experiencing a partial electrical reset that results in the inability to detect Brady, Pause, and PVC events is 0.73% at 36 months. Complaint data suggests the majority of electrical resets were associated with Electromagnetic Interference (EMI) due to cardioversion or electrocautery. Potential harms include those associated with the risk of a delayed medical intervention or missed diagnosis for Brady, Pause, and/or PVC events, and an explant procedure.

If a partial electrical reset occurs, CareLink™ and Reveal LINQ™ Mobile Manager (LMM) will continue to indicate that detection parameters are "ON;" however, Brady, Pause, and PVC events will not be automatically collected. The Patient Assistant (Patient Activator) will continue to function to manually trigger ECG collection, store the tracing and mark symptoms.

Tachy and AT/AF detections are not affected by a partial electrical reset.

HOSPITAL RISK MANAGER ACTIONS (U.S. CUSTOMERS ONLY)

Medtronic is requesting customers with affected product on hand to take the following actions:

1. Identify and quarantine all unused affected Medtronic LINQ II ICMs.
2. Return all unused affected product in your inventory to Medtronic. Contact Medtronic Customer Service at 1-800-848-9300 to initiate a product return. Your local Medtronic Representative can assist you as necessary in initiating the return of this product.
3. Please share this notification with the Cardiology and cardiac monitoring departments, Pacemaker/Device Clinic leadership, and physicians who implant or manage patients with LINQ II insertable cardiac monitors (ICMs).
4. Complete the enclosed Confirmation Form and email to RS.CFQFCA@medtronic.com

PATIENT MANAGEMENT RECOMMENDATIONS

If an electrical reset has never occurred, all detection criteria are being monitored and recorded as programmed. Continue with normal follow-up per local clinic protocols for these patients.

Identifying if an electrical reset has occurred:

For patients who are actively followed on CareLink in the U.S: During our investigation of this issue, we identified patients whose device showed evidence of a partial electrical reset as of 10 May 2021. For those clinicians with identified patients, a supplemental letter was provided. If you have not received a supplemental letter, then none of your patients who are actively transmitting on CareLink were identified as having a recorded electrical reset event during our investigation.

All patients, including those on CareLink, should be carefully monitored for reports of an electrical reset condition. Follow instructions below.

- **During in person or remote follow-up:** If a device experiences an electrical reset, clinicians will be informed via programmer pop-up or CareLink display message. Actively monitor for these notifications at each patient follow-up, and contact Medtronic Technical Services should you receive an alert. **Note:** Once cleared, electrical reset notifications are no longer accessible.
- **Retroactively:** Review the Brady lifetime episode counter from the most recent session report (CareLink or in-office). If a report is not available, consider scheduling a follow-up for each patient being monitored for Brady, Pause or PVC events. Review the Brady lifetime episode counter:
 - If the lifetime count for Brady is non-zero, a partial electrical reset has **not** occurred.
 - If the lifetime count for Brady is zero, and the Brady detection parameter indicates it is "ON," a partial electrical reset **may** have occurred. Contact Medtronic Technical Services for assistance by emailing RS.LINQElectricalResetFCA@medtronic.com (U.S.) OR calling 1-800-929-4043 (U.S.).

Patients with a confirmed partial electrical reset:

- Medtronic medical staff, in consultation with our Independent Physician Quality Panel, recommends against device replacement for patients being monitoring for Tachy or AT/AF; continue normal patient follow-up.
- When monitoring for Brady, Pause, or PVC events, device replacement may be appropriate. Consider the following before device replacement:
 - It is important to note that the Patient Assistant (Patient Activator) will continue to manually mark symptoms even after a partial electrical reset. Patient-activated recordings are not impacted by this issue.
 - If replacement is desirable, consider Reveal LINQ with TruRhythm™ or alternative ICM. While Reveal LINQ devices are also susceptible to this issue (see correction notice, *Reveal LINQ™ with TruRhythm™ Insertable Cardiac Monitoring Systems Brady & Pause Detections Disabled Following Partial Electrical Reset*), the observed rate is 0.049% for Reveal LINQ with TruRhythm ICMs compared to 0.21% for LINQ II ICMs.

Note: Implanted Reveal LINQ with TruRhythm ICMs have the ability to receive a future software update to correct this issue, which will be implemented via the Model 2090 and Encore™ programmers, and is anticipated to be available in the U.S. late calendar year 2021.

- Future manufactured LINQ II devices will have a correction for this issue implemented during manufacturing pending regulatory approval of the corrective fix, but initial supply may be limited.
- As a reminder, per the LINQ II ICM's Instructions for Use, contact Medtronic anytime an electrical reset occurs.

Reveal LINQ with TruRhythm - Brady & Pause Detections Disabled Following Electrical Reset

Reveal LINQ™ with TruRhythm™ Insertable Cardiac Monitoring Systems

Original Date of Communication: June 2021

STATUS UPDATE - APRIL 2025

This advisory is being addressed via a software update. Medtronic CareLink (2090) and Encore (29901) Programmer software, SW026 version 8.3, is available to correct a low rate of occurrence issue with Reveal LINQ ICMs where Brady and Pause detections are disabled following a partial electrical reset.

Reveal LINQ ICMs that are interrogated in-office with an updated 2090 or Encore programmer are no longer susceptible to this issue. This corrective fix to the device cannot be delivered with the Reveal LINQ™ Mobile Manager (LMM). Until the update is installed, future partial electrical resets may disable Brady and/or Pause detections as described in the June 2021 communication.

Note: The immediate availability of the software release is specific to countries that follow FDA approval, or that do not require software to be regulated. Release timing may differ for other geographies including those that require CE Mark approval. Check with your local Medtronic representative to determine if the software update is available in your region.

Please work with your local Medtronic Representative to update all 2090 and Encore device programmers. In addition, Medtronic requests you follow the below patient management recommendations:

Patient Management Recommendations:

- Reveal LINQ ICMs with a confirmed partial electrical reset will receive the corrective fix for this issue immediately by the device clinician completing the following steps:
 1. Interrogate the ICM with an updated 2090 or Encore programmer (software application SW026 version 8.3). The corrective fix is automatically installed during initial interrogation. To confirm an ICM has successfully received the update, refer to Appendix A.
 2. Per the Instructions for Use (IFU), following any electrical reset, verify ICM parameters are set appropriately for the patient and reprogram if necessary.
- For Reveal LINQ ICMs that have not experienced a partial electrical reset, an update will occur during the next in-clinic visit in which an updated Model 2090 or Encore programmer installed with software application SW026 version 8.3 (or higher) is used to interrogate the ICM. Partial electrical resets will disable Brady and/or Pause detections as described in the June 2021 communication until the update is installed on to the patient's ICM.

- For patients who are actively followed on CareLink, continue routine monitoring for CareAlerts and verify notification settings for electrical resets.
- Per the IFU, notify your Medtronic representative if an electrical reset occurs. If a partial electrical reset is confirmed, the patient's ICM will require reprogramming.
- During the programmer session, the corrective fix will be installed automatically.

ORIGINAL COMMUNICATION - JUNE 2021

This notice is to inform you that Reveal LINQ with TruRhythm ICMs that undergo a partial electrical reset appear to be programmed "ON," but are no longer able to detect and report Brady and Pause events to clinicians. Medtronic estimates that 0.049% of Reveal LINQ with TruRhythm ICMs have experienced a partial electrical reset resulting in the inability to detect Brady and Pause events. While there is a potential for underreporting due to lack of awareness that an electrical reset has occurred, there have been zero (0) serious or permanent harms or deaths reported as a result of this issue. After a partial electrical reset, these Brady and Pause episode types will not be reported to the clinician.

- Currently implanted/distributed Reveal LINQ with TruRhythm ICMs will receive a future software update to correct this issue delivered via the Model 2090 and Encore™ programmers. The corrective fix is anticipated to be available late calendar year 2021 (U.S.). Availability of the software will be communicated once Medtronic has obtained the necessary regulatory approvals.
- There will be an update for future manufactured Reveal LINQ with TruRhythm ICMs, which is anticipated to be available in the U.S. October 2021. Medtronic will inform physicians once this manufacturing update is implemented into newly manufactured Reveal LINQ with TruRhythm ICMs.

ISSUE DESCRIPTION

Medtronic has identified that Reveal LINQ with TruRhythm ICMs that undergo a partial electrical reset appear to be programmed "ON," but are no longer able to detect and report Brady and Pause events. A partial electrical reset is normal behavior that can occur when the device detects a possible issue with the device software. However, an error in the partial electrical reset implementation is causing this unintended behavior.

All Reveal LINQ with TruRhythm ICMs currently in distribution are susceptible to this issue. Through 10 May 2021, Medtronic has received 87 complaints related to an electrical reset. The projected rate of a Reveal LINQ with TruRhythm ICM experiencing a partial electrical reset that results in the inability to detect Brady and Pause events is 0.056% at 36 months. Complaint data suggests the majority of electrical resets were associated with Electromagnetic Interference (EMI) due to cardioversion or electrocautery. Potential harms include those associated with the risk of a delayed medical intervention or missed diagnosis for Brady and Pause events, and an explant procedure.

If a partial electrical reset occurs, CareLink™, Model 2090 and Encore programmer software and Reveal LINQ™ Mobile Manager (LMM) will continue to indicate that detection parameters are "ON;" however, Brady and Pause

events will not be automatically collected. The Patient Assistant (Patient Activator) will continue to function to manually trigger ECG collection, store the tracing, and mark symptoms.

Tachy and AT/AF detections are not affected by a partial electrical reset.

HOSPITAL RISK MANAGER ACTIONS (U.S. CUSTOMERS ONLY)

1. Please share this notification with the Cardiology and cardiac monitoring departments, Pacemaker/Device Clinic leadership, and physicians who implant or manage patients with LINQ II insertable cardiac monitors (ICMs).
2. Complete the enclosed Confirmation Form and email to RS.CFQFCA@medtronic.com

PATIENT MANAGEMENT RECOMMENDATIONS

If an electrical reset has never occurred, all detection criteria are being monitored and recorded as programmed. Continue with normal follow-up per local clinic protocols for these patients.

Identifying if an electrical reset has occurred:

For patients who are actively followed on CareLink in the U.S: During our investigation of this issue, we identified patients whose device showed evidence of a partial electrical reset as of 10 May 2021. For those clinicians with identified patients, a supplemental letter was provided. If you have not received a supplemental letter, then none of your patients who are actively transmitting on CareLink were identified as having a recorded electrical reset event during our investigation.

All patients, including those on CareLink, should be carefully monitored for reports of an electrical reset condition. Follow instructions below.

- **During in person or remote follow-up:** If a device experiences an electrical reset, clinicians will be informed via programmer pop-up or CareLink display message. Actively monitor for these notifications at each patient follow-up, and contact Medtronic Technical Services should you receive an alert. Note: Once cleared, electrical reset notifications are no longer accessible.
- **Retroactively:** Review the Brady lifetime episode counter from the most recent session report (CareLink or in-office). If a report is not available, consider scheduling a follow-up for each patient being monitored for Brady or Pause events. Review the Brady lifetime episode counter:
 - If the lifetime count for Brady is non-zero, a partial electrical reset has **not** occurred.
 - If the lifetime count for Brady is zero, and the Brady detection parameter indicates it is "ON," a partial electrical reset **may** have occurred. Contact Medtronic Technical Services for assistance by emailing RS.LINQElectricalResetFCA@medtronic.com (U.S.) OR calling 1-800-929-4043 (U.S.).

Patients with a confirmed partial electrical reset:

- Medtronic medical staff, in consultation with our Independent Physician Quality Panel, recommends against device replacement for patients being monitoring for Tachy or AT/AF; continue normal patient follow-up.
- When monitoring for Brady or Pause events, it is important to note that the Patient Assistant (Patient Activator) will continue to manually mark symptoms even after a partial electrical reset. Patient-activated recordings are not impacted by this issue. If patients require monitoring for Brady and/or Pause events, and it is not acceptable to wait for the software update to become available (see details below), consider device replacement. Recognize that exposure to EMI could introduce this issue for new device implants that occur before the manufacturing update is implemented anticipated in the U.S. in October 2021.
- As a reminder, per the Reveal LINQ with TruRhythm ICM's Instructions for Use, contact Medtronic anytime an electrical reset occurs.

FUTURE SOFTWARE UPDATE AVAILABILITY

Medtronic is developing a programmer-delivered software update to correct this issue for Reveal LINQ with TruRhythm ICMs currently implanted or in distribution. Anticipated availability in the U.S. is late calendar year 2021; Medtronic representatives will inform you of the availability and work with you to install the software onto clinic and hospital 2090 and Encore programmers. LMM application software will be unable to deliver the software update for this issue. In order for patients with Reveal LINQ with TruRhythm ICMs to receive the update, the device will need to be interrogated with an updated 2090 or Encore programmer.

APPENDIX A

Reveal LINQ™ with TruRhythm™ Insertable Cardiac Monitoring Systems

Brady & Pause Detections Disabled Following Partial Electrical Reset

Software Update Available


How do I update my Model 2090 and Encore programmers with the Reveal LINQ with TruRhythm software described in the November 2021 communication)?

Software update SW026 version 8.3 can be installed onto all Model 2090 or Encore programmers through either the Medtronic Software Distribution Network (SDN) or via a USB. Medtronic representatives will work with you to install the software.

How do I confirm if a Model 2090 or Encore programmer has already been installed with the updated software (SW026 v8.3)?

From the *Find Patient* screen on the programmer, Tap the Programmer Icon, select "Software," and scroll through the list of installed applications to find Reveal LINQ Software Version 8.3.

Customer Communications



How do I confirm if a patient's Reveal LINQ ICM has received the software update?

Clinicians can confirm if a patient's ICM has received the software update by verifying the Device Configuration ID via a 2090 or Encore programmer (LMM will not display the Configuration ID). To locate the Device Configuration ID, enter a follow-up session and Print the Parameters Report. All Reveal LINQ ICMs that have received the software update will have their Configuration ID ending in a "1" (e.g. X-X-X-1).

NOTE: The Reveal LINQ Device Configuration ID can be viewed on CareLink beginning January 2022 after a manual transmission is completed by the patient.

Parameters			
Symptom Four 7.5 min Episodes			
	Detection	Interval (Rate)	Duration
Tachy	Off	340 ms (176 bpm)	16 beats
Brady	Off	2000 ms (30 bpm)	4 beats
Pause	Off		3 sec
AT/AF Detection			
AT/AF Detection Off			
Sensing			
Sensitivity		0.035 mV (35 µV)	
Blank after Sense		300 ms	
Sensing Threshold Decay Delay		200 ms	
Device Data Collection			
Reason for Monitoring	Suspected AF		
Device Date/Time	26-Aug-2021 06:44		
Wireless Transmission Time	00:00		
Wireless Data Priority	Pause, Tachy, Brady		
Device Data Collection	On		
Device Information			
Device	Medtronic	REVEAL LINQ Linq11	RLA511585S
Device Configuration ID:	0-0-0-1		
Implanted:	23-Mar-2021		
History			

Potential for Shortened RRT-to-EOS in Subset of ICDs and CRT-Ds

Subset of ICDs and CRT-Ds

Original Date of Communication: February 2021

STATUS UPDATE - APRIL 2025

As of 16 April 2025, approximately 122,113 devices susceptible to this issue are estimated to still be active worldwide. Observed rate of occurrence is 0.17% and projected rate for the affected population of devices remains 0.22%. Devices with higher pacing outputs and high pacing percentages (e.g., CRT-D devices) have the lowest probability of occurrence (refer to Appendix A of the original communication for further details – see below). No permanent patient harms have been reported due to this issue.

ORIGINAL COMMUNICATION – FEBRUARY 2021

In February 2021, Medtronic informed physicians of a potential issue for a subset of Implantable Cardioverter Defibrillators (ICDs) and Cardiac Resynchronization Therapy Defibrillators (CRT-Ds). Medtronic has identified that a small percentage of implanted cardiac devices, from a well-defined subset, may experience a shortened Recommended Replacement Time (RRT) to End of Service (EOS) interval following an earlier-than-expected RRT observation. The subset of ICDs and CRT-Ds affected by this issue were last implanted in February 2019 and manufactured with a specific battery design that is no longer being distributed.

We have received no reports of permanent harm to patients as a result of this issue.

Approximately 339,900 devices susceptible to this issue are estimated to still be active worldwide. Through 4 January 2021, confirmed events (observed rate 0.07%) have involved a rapid drop in battery voltage ranging from days to months, with unexpected RRT as one of the primary reported observations. For those devices in which RRT triggered earlier than expected, the median time from RRT to the EOS observation was 14 days. In a small number of the cases, no output/no telemetry was reported prior to device replacement. Medtronic projects approximately 0.22% of the affected device population may experience this issue during their service life.

The rapid depletion is caused by a latent shorting mechanism involving lithium plating, resulting from a thermal gradient between the anode and cathode components of the battery. **Devices with higher pacing outputs and high pacing percentages (e.g. CRT-D devices) have the lowest probability of occurrence (refer to Appendix A – see below).** Conversely, devices with low current drain (evidenced by a longer overall service time from implant to RRT) have a higher probability of experiencing this issue. Importantly, the probability of this issue developing is constant after approximately three years of service time.

Patient Management Guidance

We realize that each patient requires unique clinical considerations. In consultation with our Independent Physician Quality Panel (IPQP), Medtronic **recommends** the following:

- **Continue normal follow-up per local clinical protocol.**
 - Recognize that patients who require significant pacing support and high voltage therapy have the lowest risk for this issue - See Appendix A for additional details.
 - Where possible, take advantage of the CareLink™ home monitoring system and the wireless low battery voltage CareAlert.
 - The low battery voltage audible alert is shipped On with high-urgency tones; Remind patients to contact their clinic if they hear an audible alert, particularly since patients may be opting to delay clinic visits due to COVID-19 guidance.
 - Inform a Medtronic Representative of any unexpected device behaviors.
 - Be aware that the inability to interrogate the device, or to transmit data, may be an indicator that the device has experienced this issue.
- **If unexpected RRT is observed, prompt replacement of the device should occur commensurate with the underlying clinical situation of the patient:**
 - For non-pacing dependent patients or for primary prevention ICD patients, replacement within 1 week of an unexpected RRT notification is recommended.
 - For pacing dependent patients, immediate replacement is recommended following an unexpected RRT notification.

Note: For all patients, this issue can also manifest as an unexpected change in the remaining longevity estimate that cannot be attributed to programming changes, or changes in use conditions.

Medtronic medical staff in consultation with the IPQP **recommends against prophylactic replacement** due to the low rate of occurrence and the low potential for permanent harm when prompt replacement occurs in response to an unexpected RRT.

Patients and clinicians may determine if a specific device is affected by looking up the serial number on Medtronic's Product Performance website: productperformance.medtronic.com


APPENDIX A

The table below provides a comparison of sample use conditions and their associated, projected service time (Implant to Recommended Replacement Time), along with their cumulative and per-year risk of encountering rapid depletion due to a latent shorting mechanism in the battery. Devices with higher pacing outputs and high pacing percentages have the lowest probability of occurrence. There have been no reports of permanent harm to patients as a result of this issue.

Probability (Risk per Year) of Rapid Depletion due to this Issue as a Function of Service Time

Projected Service Time * (based on sample programmed settings and use conditions)	Projected Risk per Year & Total Cumulative Risk at end of service time++	Notes/Example
12-year service time	0.11% per year, 0.98% cumulative	VR ICD patient with 0% pacing and no shocks delivered
10.25-year service time	0.070% per year, 0.50% cumulative	VR ICD patient with 50% pacing history and two (2) or fewer shocks per year
9-year service time	0.045% per year, 0.27% cumulative	DR ICD patient with little-to-no pacing history (e.g. 10%AP, 25%VP, and two (2) or fewer shocks per year)
8.25-year service time	0.033% per year, 0.18% cumulative	DR ICD patient with complete heart block (10% AP and 100% VP, and two (2) or fewer shocks per year)
7-year service time	0.017% per year, 0.075% cumulative	CRT-D patient with 15% AP, 90% RVP, 100% LVP, and two (2) or fewer shocks per year
* Assumes current drain remains stable throughout life of device (i.e. No change in remaining longevity due to reprogramming or changes in use conditions)	++ Per annum risk of issue becomes constant after approximately 3 years of service time. Cumulative risk = early risk plus annual risk over the projected service time.	A output = 1.5V, 0.4ms, 500 ohms RV output = 2.0V, 0.4ms, 500 ohms LV output = 2.5V, 0.4ms, 500 ohms Average pacing rate = 75 bpm

Customer Communications



Cumulative Probability is the expected risk for a device to experience this issue between implant and end of service. When risk is evaluated for a device that has reached a service life beyond 3 years, the remaining risk can be estimated based on the yearly risk value shown.

The Population Average (0.22%) is the cumulative probability for the full subset of devices susceptible to this issue. This value takes into account expected longevity and patient mortality. Not all devices with projected service time of 12 years will be in service all 12 years.

Key Points:
Slope of the curve reflects the risk per year based on sample service times of 7, 9, and 12 years.

Slope (risk per year) is constant after approximately 3 years of service time.

Performance Note: Potential for no output/no telemetry condition in subset of IPG and CRT-P products due to ceramic capacitor leakage pathway

Azure™ and Astra™ pacemakers, and Percepta™, Serena™ and Solara™ CRT-P

Original Date of Communication: May 2019

STATUS UPDATE - APRIL 2025

As of 16 April 2025, there have been a total of 33 confirmed events worldwide associated with this failure mode. No additional deaths, beyond the two deaths previously disclosed*, have been reported since the October 2020 update. Confirmed events included reports of no output, premature depletion and electrical reset that reverted to VVI 65 operation.

Medtronic's ongoing monitoring of these failures have allowed us to improve long-term modeling projections for future device failures. The overall projected lifetime rate of occurrence for the remaining active devices manufactured with the original low voltage capacitor is projected to be 0.025%.

All products in distribution are unaffected. The specific low voltage capacitor susceptible to this issue was last used in manufacturing 01 June 2019. Patient management recommendations remain unchanged from the original posting (refer to the May 2019 text below).

*Assessment for cause of death determined loss of pacing therapy could not be ruled out as a contributing factor.

ORIGINAL COMMUNICATION - MAY 2019

Medtronic has identified a rare but potentially serious failure mode in a population of Azure™ and Astra™ pacemakers, and Percepta™, Serena™ and Solara™ cardiac resynchronization therapy pacemakers (CRT-P), manufactured with a specific multilayer ceramic capacitor. These devices continue to perform within reliability projections.

While inherently very reliable, a known failure mode of these capacitors is the potential for internal cracking that can be caused by thermal-mechanical stress during manufacturing. Under rare conditions, internal cracking within a capacitor may result in the development of a leakage pathway, causing high current drain and leading to rapid battery depletion. While the issue presents as rapid battery depletion, this is not a battery performance issue.

As of April 26, 2019, three complaints out of ~266,700 devices distributed worldwide since February 2017, have been received that included a no output /no telemetry scenario resulting from rapid battery depletion. Battery depletion due to this issue can range from several days to several weeks. One of these reported events contributed to a patient

Customer Communications

death. The three confirmed failures occurred within 9 months post implant. The projected rate for this issue is 0.0028%, with the most susceptible period for a leakage pathway to develop in the capacitor being the first 12 months post implant.

Based on the low predicted rate of failure and the recent implementation of process and component enhancements, Medtronic expects few, if any, additional events to occur. Medtronic, in consultation with our Independent Physician Quality Panel, does not recommend device replacement. Physicians should continue normal patient follow-up in accordance with standard practice, and where possible, continue to utilize the low battery voltage wireless CareAlert™ (shipped ON), together with remote monitoring via CareLink™ home monitor or the MyCareLink Heart™ mobile app. Per the instructions for use, at each follow-up, verify the status of the implanted system as well as the clinical effectiveness of the device. Pay attention to any unexpected changes in remaining longevity estimates or the inability to interrogate the device and/or transmit data.

Contact Medtronic Technical Services if you have concerns on a specific patient.

Brady Technical Services | rs.techservices@medtronic.com | 800-505-4636

Potential Conductor Wire Fracture

6930, 6931, 6948, 6949 Sprint Fidelis Defibrillation Leads

Original Date of Communication: October 2007

STATUS UPDATE - APRIL 2025

As of April 21, 2025, of the initial implant population of 205,600 in the United States, approximately 26,600 remain implanted. According to Product Surveillance Registry results, lead survival is estimated to be 63.5% (+6.6/-6.0%) at 180 months. As the implanted population ages and the sample size increases for each time interval, the accuracy of the estimated survival probability will increase as shown by tighter confidence intervals.

Initial Affected Population	Number of Confirmed Advisory Related Events	Estimated Remaining Active Population
279,500 Worldwide (205,600 United States)	7,338 Worldwide (5,270 United States)	36,300 Worldwide (26,600 United States)

ORIGINAL COMMUNICATION - OCTOBER 2007

PRODUCT

All Model 6930, 6931, 6948, and 6949 implantable defibrillation leads

ADVISORY

There are two primary locations where chronic conductor fractures have occurred on Sprint Fidelis leads: 1) the distal portion of the lead, affecting the anode (ring electrode) and 2) near the anchoring sleeve tie-down, predominantly affecting the cathode (helix tip electrode), and occasionally the high voltage conductor. These two locations account for approximately 90% of the chronic fractures identified in Returned Product Analysis (RPA). The remaining 10% of chronic fractures occurred in the DF-1 connector leg and the proximal portion of the RV coil. High voltage conductor fractures could result in the inability to deliver defibrillation therapy. Anode or cathode conductor fractures (at either location) may present clinically as increased impedance, oversensing, increased interval counts, multiple inappropriate shocks, and/or loss of pacing output.

PATIENT MANAGEMENT RECOMMENDATIONS (UPDATED APRIL 2011)

The Lead Integrity Alert (LIA) provides three days advance notice prior to inappropriate therapy to 76% of patients with lead fractures¹. As a result, we strongly recommend that all Sprint Fidelis patients who have the ability to upgrade to Lead Integrity Alert do so promptly. Also ensure that high voltage lead impedance alerts (maximum of 100 ohms) are programmed. When a lead fracture is suspected or confirmed, immediate patient attention is strongly recommended. Physicians should inform their patients to seek medical attention without delay if they experience unexpected shocks.

- **If a Fidelis lead fracture of any type has occurred, we recommend implanting a new high voltage lead with or without extraction of the Fidelis lead.**
- In patients with normal device function and no manifestation of lead fracture, no action is recommended. The risk of prophylactic intervention appears to be greater than serious injury resulting from lead fracture even for pacemaker dependent patients, except in select individual patient circumstances as determined by the physician.
- In the event of a device change-out or upgrade procedure, with no manifestation of lead fracture, consider the patient age and lead model data above, as well as patient life expectancy, co-morbidities, ease of extraction related to implant time, patient preference, etc., for the following options:
 - Leave a properly performing lead intact.
 - Implant a new ICD lead without extraction of the existing lead.
 - Carefully consider all factors before prophylactic placement of a pace-sense lead. Data shows an increased risk of high voltage conductor fracture if a pace-sense conductor fracture has previously occurred.
 - Individual patient circumstances may warrant extracting and implanting a new ICD lead. If warranted, Medtronic's Independent Physician Quality Panel recommends the lead extraction procedure be performed by a physician with extensive lead extraction experience.²

Footnotes:

1: Swerdlow C, Gunderson, B, et al. "Downloadable Algorithm to Reduce Inappropriate Shocks Caused by Fractures of Implantable Cardioverter-Defibrillator Leads", *Circulation*, November 2008, 118: 2122-2129.

2: "Transvenous Lead Extraction: Heart Rhythm Society Expert Consensus on Facilities, Training, Indications, and Patient Management", *Heart Rhythm*, Vol 6, No 7, July 2009.

Product Education Brief: Daily Restarts of SmartSync iPad Tablets

CareLink SmartSync™ Device Manager Application and Pacing System Analyzer (PSA)

Original Date of Communication: March 2025

Overview

This Product Education Brief emphasizes the recommendation to **restart tablets daily as part of SmartSync management**. This may be completed by powering the tablet off between uses. Daily tablet restarts facilitate optimal SmartSync App performance. Tablets should also be restarted upon receipt of a “low tablet resources” pop-up, as indicated in the pop-up text.

Impacts of Restarting the SmartSync Tablet

Restarting the tablet clears the RAM (random-access memory) and closes any running apps, which improves overall tablet performance. When RAM is not cleared, or when other apps run concurrently with the SmartSync app, the following behavior may be observed:

- Unexpected app closures*
- App unresponsiveness*
- System error messages as a result of limited available system memory


Should any of the above behaviors occur, the user should force close all open apps on the tablet, restart the tablet, and then restart the SmartSync app.

* Note that if the app disconnects during an analyzer session, pacing will continue at the programmed parameters for 60 minutes if the surgical/patient cables are connected to implantable device leads or 5 minutes without connection to implantable device leads.

How to Restart Apple iPad

How to restart an iPad tablet without a Home button:

Education Briefs



1. Press and hold either volume button and the top button until the power off slider appears
2. Drag the slider, then wait 30 seconds for your device to turn off.
 - If your device is frozen or unresponsive, force restart your device: Press and quickly release the volume up button, press and quickly release the volume down button, then press and hold the top button. When the Apple logo appears, release the button.
3. To turn the tablet back on, press and hold the top button until you see the Apple logo.

How to restart an iPad tablet with a Home button:



1. Press and hold the top button until the power off slider appears
2. Drag the slider, then wait for 30 seconds for your device to turn off.
 - If your tablet is frozen or unresponsive, force restart your tablet: Press and hold the top button and the Home button at the same time. When the Apple logo appears, release both buttons.
3. To turn your tablet back on, press and hold the top button until you see the Apple logo.

Product Education Brief: Micra TPS Monitoring and End of Service Behavior

Micra™ TPS devices

Original Date of Communication: January 2025

Overview

This Product Education Brief provides a description of specific monitoring and follow-up recommendations, including a description of end of service behavior.

This Product Education Brief is consistent with existing labeling and reiterates precautions listed in the product's Instructions for Use (IFU).

Micra Instructions for Use

The Micra IFUs are available on the Medtronic electronic manuals website (manuals.medtronic.com).

Monitoring, Follow-Up, and End of Service

Remote monitoring of patients with a Micra device is recommended in the 2023 HRS/EHRA/APHRS/LAHR expert consensus statement.

The 2023 HRS/EHRA/APHRS/LAHR expert consensus statement on practical management of the remote device clinic states that in patients with CIEDs, remote monitoring (RM) is recommended as part of the standard of care. The consensus statement advises that patients with CIEDs on RM, in the absence of continuous connectivity, such as with Micra, remote transmissions are recommended at least every 3-12 months. It is also noted that there are some circumstances (e.g., if a patient is pacemaker dependent) where the transmission frequency may match or exceed every 3-6 months. As the device approaches elective replacement, the frequency of transmissions should be increased to every 1-3 months. The frequency recommendations apply whether the patient transmits data remotely or has an in-office visit.

Following recommended replacement time / elective replacement indicator (RRT/ERI) and as described in the IFU, the Micra device sets battery end of service (EOS) after measuring voltage $\leq 2.5V$ on 3 consecutive daily automatic measurements. When the Micra device reaches EOS, it permanently deactivates pacing and switches to the Device Off mode. This EOS mode differs from Medtronic transvenous pacemakers. The IFU indicates that the estimated time from RRT to EOS is 6 months, and the approximate time from ERI to EOS is 3 months. See the device manuals for detailed information regarding the implant procedure, indications, contraindications, warnings, precautions, MRI conditions for use, and potential complications/adverse events.

Product Performance

Medtronic monitors and evaluates product performance, including battery performance, for all Micra devices. The device performance data is published on our product performance website productperformance.medtronic.com. Reports of Micra premature battery depletion and normal battery depletion are included on the product performance website.

In addition, Medtronic continues to collaborate with physicians and regulatory agencies to improve patient outcomes and clinical experience as part of our dedication to patient safety and product effectiveness.

References

- medtronicacademy.com
- Micra MC1VR01 Clinician Manual - M042502C001
- Ferrick, A. M., et al. (2023). 2023 HRS/EHRA/APHRS/LAHR expert consensus statement on practical management of the remote device clinic. *Europace*, 25(5), eead123.

Product Education Brief: Alert Threshold for Lead Impedances

Azure™ and Astra™ pacemakers, and Percepta™, Serena™ and Solara™ CRT-P

Original Date of Communication: April 2023

Overview

This Product Education Brief describes a lead impedance alert behavior in Medtronic Azure™, Astra™, Percepta™, Serena™, and Solara™ devices that may be observed when the pacing lead impedance displayed is slightly above the programmed lower alert threshold.

Details regarding CareAlerts and impedance measurements

In Azure, Astra, Percepta, Serena, and Solara devices, a CareAlert may be triggered when the lead impedance displayed is slightly above the programmed lower alert threshold.

See Figure 1 and 2 for an example case where the lower threshold for a lead impedance alert is programmed at 200 ohms. In the Lead Impedance Trend (Figure 1) the precise measured impedance value is 209 ohms; and yet in the CareAlert display (Figure 2) an alert was triggered indicating the impedance crossed the 200 ohm impedance threshold, and suggesting the impedance is 190 ohms. This scenario can happen due to the range of impedance values used for alert monitoring. In this scenario, the devices are functioning within design specifications and tolerances.




Figure 1– Lead Impedance Trend showing precise impedance values over time

Date/Time	Event	Threshold
01-Jul-2020 03:00:00	*A. Unipolar lead impedance 190 ohms.	200 ohms

Figure 2– CareAlert triggered showing 190 ohms impedance value with a 200ohm alert threshold

While most leads are programmed to pace bipolar, every night the device will assess all lead impedance vectors. Unipolar measurements tend to be lower than bipolar measurements in the same system. Therefore, the nightly unipolar lead impedance measurement will likely be closer to the programmed lower alert threshold than the bipolar measurement.

The impedance value range for alerts is never higher than the measured impedance, so all impedance values that are below the programmed threshold will generate an alert as programmed.

Patient Management

Clinicians should continue to follow patients per their standard clinical practice when a CareAlert triggers for lead impedances. Impedance alerts are designed to prompt the clinician to review impedances in the Lead Impedance Trend Log. The Lead Impedance Trend records precise lead impedance measurements over time. These historic impedance measurements provide evidence as to whether the impedance is stable and functioning as designed (albeit near the programmed threshold); or unstable and necessitating further assessment and/or close monitoring. Note: If impedance values reported in the Lead Impedance Trend indicate impedances are below the programmed lower threshold, then the lead should be assessed for possible insulation breach per standard clinical practice.

Procedure Education Brief: Micra TPS Implant

Micra™ TPS devices

Original Date of Communication: November 2021

Overview

This Medtronic Procedure Education Brief provides a reminder of specific implant procedure safety recommendations included in the current labeling for Micra™ VR and Micra™ AV Transcatheter Pacing System (TPS) specifically from the Micra Instructions for Use (IFU) and the Micra implanter training program. Following instructions provided in the IFU and implanter training can reduce the risk of cardiac perforation, especially considerations for delivery system steering, repositioning the device, and patient selection.

Micra IFU and Implant Procedure Training

The Micra IFU is available on the Medtronic electronic manuals website (<https://manuals.medtronic.com/manuals/main/region>). Implanter training material for implanters who have attended training, and been certified to implant Micra TPS, can be found on the Medtronic Academy website. These instructional materials provide recommendations that limit implant complications such as:

- patient selection considerations to minimize perforation risk
- steering the delivery system with the use of fluoroscopy
- identifying implant location at the right ventricular septum with the use of contrast-enhanced fluoroscopy
- confirming position on the septum with contrast-enhanced fluoroscopy prior to deployment
- considerations for repositioning the device
- ensuring attending staff are prepared to manage pericardial effusion and tamponade, including immediate access to echocardiography equipment and availability of a pericardiocentesis kit
- recognizing clinical signs and symptoms of pericardial effusion and tamponade in order to minimize clinical response time
- preparedness for cardiac surgical intervention

Micra Safety and Effectiveness Data

On 17 November 2021, the US FDA posted a Letter to Healthcare Providers (*Leadless Pacing Systems: Risk of Major Complications Related to Cardiac Perforation During Implantation - Letter to Health Care Providers*) reminding physicians about the rare but possible risk of cardiac perforations associated with leadless pacemaker implantation. They reiterated the specific recommendations from Medtronic Micra implant training and IFU (reviewed above). This communication can be found here: <https://www.fda.gov/medical-devices/medical-device-safety/letters-health-care-providers>.

While regulatory agencies and Micra implanters are aware that cardiac perforation is a known risk, the FDA Letter to Healthcare Providers included data for which implanters may be seeking additional context. The letter indicated that

risk of cardiac perforation between transvenous pacing implants and Micra implants are similar, and that Micra implant complication rates are within expectations. The letter also indicated that in some scenarios, when a perforation occurs with a Micra implant procedure, the severity of the perforation complications can be higher than when a perforation occurs with a transvenous implant procedure.

Data from our Global Complaint Handling database suggests that the Micra rate of perforation is 0.6% and the rate of perforation related death is 0.13% out of over 100,000 implants worldwide. The Micra real-world perforation rate is in-line with, or lower than, the perforation rate observed in pre-market or post-market clinical studies¹.

Since Micra received pre-market approval in 2016, Medtronic has continuously monitored its safety and effectiveness. Multiple studies have shown that Micra has a high rate of implant success (exceeding 99%)^{2,3}. Additionally, in the global Micra Investigational Device Exemption (IDE) Trial, Micra has been shown to reduce the risk for major complications compared to transvenous implants (through 12-months) by 48%², and in the global Micra Post Approval Registry by 63%³.

Medtronic is further assessing the outcomes of Micra in the Micra Coverage with Evidence Development (CED) Study, which is a continuously enrolling, observational, cohort study evaluating complications, utilization, and outcomes of Micra.

Recent publications from the Micra CED Study in July 2021⁴ and November 2021⁵ based on 5,746 Micra patients and 9,622 with contemporaneously implanted transvenous single-chamber pacing patients show that at time of implant, Micra patients tend to be sicker than the transvenous single-chamber pacemaker population. Micra patients have a higher comorbidity burden as measured by the Charlson comorbidity index (5.1 vs 4.6, $P < 0.001$) and a higher rate of end stage renal disease (12.0% vs 2.3%, $P < 0.001$)⁴. Acute and longer-term outcomes reported in these publications are shown in the table below.

Measure	Unadjusted Results	Results Adjusted for Patient Medical History
	(Micra vs Transvenous-VVI)	(Micra vs Transvenous-VVI)
Acute (30-day) device-related complications including dislodgement, infection, pocket complications ⁴	1.4% vs 2.6% ($P < 0.001$)	1.4% vs 2.5% ($P < 0.001$)
Total acute (30-day) complications ⁴	8.4% vs 7.3% ($P = 0.02$)	7.7% vs 7.4% ($P = 0.49$)
Cardiac perforation/effusion ⁴	0.8% vs 0.4% ($P < 0.001$)	0.8% vs 0.4% ($P < 0.001$)
30-day all-cause mortality ⁵	4.4% vs 3.8% ($P = 0.10$)	4.0% vs 4.4% ($P = 0.60$)

2-year reintervention rate ⁵	3.0% vs 4.8% (P=0.006)	3.1% vs 4.9% (P=0.003)
2-year chronic complications ⁵	4.9% vs 6.5% (P<0.001)	4.6% vs 6.5% (P<0.001)
2-year all-cause mortality ⁵	34.0% vs 31.6% (P=0.002)	31.4% vs 32.5% (P=0.37)

Medtronic monitors and evaluates product performance and publishes device performance data on our product performance website <http://productperformance.medtronic.com>. In addition, Medtronic continues to collaborate with physicians and regulatory agencies to improve patient outcomes and clinical experience as part of our dedication to patient safety and product effectiveness.

¹ Micra IDE: 1.8% (13/726), Micra post-approval registry 0.8% (15/1811), Micra Coverage with Evidence Development 0.8% (47/5746)

² Reynolds et al. *NEJM* 2016; 374(6): 533-541.

³ El-Chami et al. *Heart Rhythm* 2018; 15(12): 1800-1807.

⁴ Piccini et al. *JAMA Cardiology* 2021; 6(10): 1187-1195.

⁵ El-Chami et al. *EHJ* 2021; ePub ahead of print

Legacy Models

Medtronic, at its discretion, may stop providing updated performance information on models in alignment with the inclusion criteria defined in the methods for estimating. Listed below are the final product performance reports for legacy models.

GENERATORS

Cardiac Resynchronization Therapy (CRT) Defibrillators

Product Name	Model	Final Issue
Cardia CRT-D	D384TRG	2023 2nd Edition (Issue 89)
Concerto CRT-D	C154DWK	2016 1st Edition (Issue 74)
Concerto CRT-D	C164AWK	2016 1st Edition (Issue 74)
Concerto CRT-D	C174AWK	2016 1st Edition (Issue 74)
Concerto II CRT-D	D274TRK	2023 2nd Edition (Issue 89)
Concerto II CRT-D	D294TRK	2023 2nd Edition (Issue 89)
Consulta CRT-D	D204TRM	2023 2nd Edition (Issue 89)
Consulta CRT-D	D214TRM	2023 2nd Edition (Issue 89)
Consulta CRT-D	D224TRK	2024 2nd Edition (Issue 91)
Consulta CRT-D	D234TRK	2023 2nd Edition (Issue 89)
InSync II Marquis	7289	2012 1st Edition (Issue 66)
InSync Maximo	7303	2012 1st Edition (Issue 66)
InSync Maximo	7304	2016 1st Edition (Issue 74)
InSync Sentry	7297	2012 1st Edition (Issue 66)
InSync Sentry	7299	2016 1st Edition (Issue 74)
Maximo II CRT-D	D264TRM	2023 2nd Edition (Issue 89)
Protecta CRT-D	D334TRG	2023 2nd Edition (Issue 89)
Protecta CRT-D	D334TRM	2023 2nd Edition (Issue 89)
Protecta XT CRT-D	D314TRM	2023 2nd Edition (Issue 89)

Cardiac Resynchronization Therapy (CRT) Pacemakers

Product Name	Model	Final Issue
InSync	8040	2016 1st Edition (Issue 74)
InSync III	8042	2023 2nd Edition (Issue 89)

Implantable Cardioverter Defibrillators (ICDs)

Product Name	Model	Final Issue
Entrust AT	D153ATG	2019 2nd Edition (Issue 81)
Entrust AT	D154ATG	2019 2nd Edition (Issue 81)
Entrust DR	D153DRG	2019 2nd Edition (Issue 81)
Entrust DR	D154DRG	2019 2nd Edition (Issue 81)
Entrust Escudo	D144DRG	2019 2nd Edition (Issue 81)

Implantable Cardioverter Defibrillators (ICDs) continued

Product Name	Model	Final Issue
Entrust Escudo	D144VRC	2019 2nd Edition (Issue 81)
Entrust VR	D153VRC	2019 2nd Edition (Issue 81)
Entrust VR	D154VRC	2019 2nd Edition (Issue 81)
GEM	7227B	2011 1st Edition (Issue 64)
GEM	7227Cx	2011 1st Edition (Issue 64)
GEM	7227D	2011 1st Edition (Issue 64)
GEM	7227E	2011 1st Edition (Issue 64)
GEM DR	7271	2011 1st Edition (Issue 64)
GEM III DR	7275	2012 1st Edition (Issue 66)
GEM III VR	7231Cx	2016 1st Edition (Issue 74)
Intrinsic	7288	2016 1st Edition (Issue 74)
Marquis DR	7274	2016 1st Edition (Issue 74)
Marquis VR	7230B	2019 2nd Edition (Issue 81)
Marquis VR	7230Cx	2019 2nd Edition (Issue 81)
Marquis VR	7230E	2019 2nd Edition (Issue 81)
Maximo DR	7278	2017 1st Edition (Issue 76)
Maximo VR	7232B	2019 2nd Edition (Issue 81)
Maximo VR	7232Cx	2023 2nd Edition (Issue 89)
Maximo VR	7232E	2019 2nd Edition (Issue 81)
Onyx	7290Cx	2013 1st Edition (Issue 68)
Protecta DR	D334DRG	2023 2nd Edition (Issue 89)
Protecta DR	D334DRM	2023 2nd Edition (Issue 89)
Protecta VR	D334VRG	2023 2nd Edition (Issue 89)
Protecta VR	D334VRM	2023 2nd Edition (Issue 89)
Protecta XT VR	D314VRG	2024 2nd Edition (Issue 91)
Protecta XT VR	D314VRM	2024 2nd Edition (Issue 91)
Secura DR	D204DRM	2023 2nd Edition (Issue 89)
Secura DR	D214DRM	2023 2nd Edition (Issue 89)
Secura DR	D224DRG	2024 2nd Edition (Issue 91)
Secura DR	D234DRG	2023 2nd Edition (Issue 89)
Secura VR	D204VRM	2023 2nd Edition (Issue 89)
Secura VR	D224VRC	2023 2nd Edition (Issue 89)
Virtuoso DR	D154AWG	2019 2nd Edition (Issue 81)
Virtuoso DR	D164AWG	2023 2nd Edition (Issue 89)
Virtuoso II DR	D274DRG	2023 2nd Edition (Issue 89)
Virtuoso II DR	D294DRG	2023 2nd Edition (Issue 89)
Virtuoso II VR	D274VRC	2023 2nd Edition (Issue 89)
Virtuoso VR	D154VWC	2019 2nd Edition (Issue 81)
Virtuoso VR	D164VWC	2023 2nd Edition (Issue 89)

Implantable Pulse Generators (IPGs)

Product Name	Model	Final Issue
Advisa DR	A4DR01	2019 1st Edition (Issue 80)
AT500	AT501	2013 1st Edition (Issue 68)
EnPulse	E2D01	2017 2nd Edition (Issue 77)
EnPulse	E2D03	2017 2nd Edition (Issue 77)
EnPulse DR	E1DR01	2017 2nd Edition (Issue 77)
EnPulse DR	E1DR03	2017 2nd Edition (Issue 77)
EnPulse DR	E1DR06	2017 2nd Edition (Issue 77)
EnPulse DR	E1DR21	2017 2nd Edition (Issue 77)
EnPulse DR	E2DR01	2017 2nd Edition (Issue 77)
EnPulse DR	E2DR03	2017 2nd Edition (Issue 77)
EnPulse DR	E2DR06	2017 2nd Edition (Issue 77)
EnPulse DR	E2DR21	2017 2nd Edition (Issue 77)
EnPulse DR	E2DR31	2017 2nd Edition (Issue 77)
EnPulse DR	E2DR33	2017 2nd Edition (Issue 77)
EnPulse SR	E2SR01	2017 2nd Edition (Issue 77)
EnPulse SR	E2SR03	2017 2nd Edition (Issue 77)
EnPulse SR	E2SR06	2017 2nd Edition (Issue 77)
EnPulse VDD	E2VDD01	2017 2nd Edition (Issue 77)
EnRhythm DR	P1501DR	2023 2nd Edition (Issue 89)
EnRhythm MRI	EMDR01	2017 1st Edition (Issue 76)
Kappa 400 DR	KDR401	2017 1st Edition (Issue 76)
Kappa 400 DR	KDR403	2017 1st Edition (Issue 76)
Kappa 400 SR	KSR401	2017 1st Edition (Issue 76)
Kappa 400 SR	KSR403	2017 1st Edition (Issue 76)
Kappa 600 DR	KDR601	2012 1st Edition (Issue 66)
Kappa 600 DR	KDR603	2012 1st Edition (Issue 66)
Kappa 600 DR	KDR606	2012 1st Edition (Issue 66)
Kappa 600 DR	KDR651	2012 1st Edition (Issue 66)
Kappa 600 DR	KDR653	2012 1st Edition (Issue 66)
Kappa 700 DR	KD700	2017 1st Edition (Issue 76)
Kappa 700 DR	KD701	2017 1st Edition (Issue 76)
Kappa 700 DR	KD703	2017 1st Edition (Issue 76)
Kappa 700 DR	KD706	2017 1st Edition (Issue 76)
Kappa 700 DR	KDR700	2017 1st Edition (Issue 76)
Kappa 700 DR	KDR701	2017 1st Edition (Issue 76)
Kappa 700 DR	KDR703	2017 1st Edition (Issue 76)
Kappa 700 DR	KDR706	2017 1st Edition (Issue 76)
Kappa 700 DR	KDR721	2017 1st Edition (Issue 76)
Kappa 700 SR	KSR700	2016 2nd Edition (Issue 75)
Kappa 700 SR	KSR701	2017 1st Edition (Issue 76)
Kappa 700 SR	KSR703	2017 1st Edition (Issue 76)
Kappa 700 SR	KSR706	2017 1st Edition (Issue 76)

Implantable Pulse Generators (IPGs) continued

Product Name	Model	Final Issue
Kappa 700 VDD	KVDD701	2012 2nd Edition (Issue 67)
Kappa 800 DR	KDR801	2013 1st Edition (Issue 68)
Kappa 800 DR	KDR803	2013 1st Edition (Issue 68)
Kappa 900 D	KD901	2017 1st Edition (Issue 76)
Kappa 900 D	KD903	2017 1st Edition (Issue 76)
Kappa 900 D	KD906	2017 1st Edition (Issue 76)
Kappa 900 DR	KDR901	2017 1st Edition (Issue 76)
Kappa 900 DR	KDR903	2017 1st Edition (Issue 76)
Kappa 900 DR	KDR906	2017 1st Edition (Issue 76)
Kappa 900 DR	KDR921	2017 1st Edition (Issue 76)
Kappa 900 SR	KSR901	2017 1st Edition (Issue 76)
Kappa 900 SR	KSR903	2017 1st Edition (Issue 76)
Kappa 900 SR	KSR906	2017 1st Edition (Issue 76)
Kappa 900 VDD	KVDD901	2017 1st Edition (Issue 76)
Legend II	8424	2012 1st Edition (Issue 66)
Legend II	8426	2012 1st Edition (Issue 66)
Legend II	8427	2012 1st Edition (Issue 66)
Minix	8340	2012 1st Edition (Issue 66)
Minix	8341	2012 1st Edition (Issue 66)
Minix	8341M	2012 1st Edition (Issue 66)
Minix	8342	2012 1st Edition (Issue 66)
Minix ST	8330	2012 1st Edition (Issue 66)
Minix ST	8331	2012 1st Edition (Issue 66)
Minix ST	8331M	2012 1st Edition (Issue 66)
Minuet	7107	2012 1st Edition (Issue 66)
Minuet	7108	2012 1st Edition (Issue 66)
Preva DR	7088	2012 1st Edition (Issue 66)
Preva DR	7089	2012 1st Edition (Issue 66)
Preva SR	8088	2012 1st Edition (Issue 66)
Preva SR	8089	2012 1st Edition (Issue 66)
Prevail S	8085	2012 1st Edition (Issue 66)
Prevail S	8086	2012 1st Edition (Issue 66)
Prodigy DR	7860	2012 1st Edition (Issue 66)
Prodigy DR	7861	2012 1st Edition (Issue 66)
Prodigy DR	7862	2012 1st Edition (Issue 66)
Prodigy SR	8158	2013 1st Edition (Issue 68)
Prodigy SR	8160	2013 1st Edition (Issue 68)
Prodigy SR	8161	2013 1st Edition (Issue 68)
Prodigy SR	8162	2013 1st Edition (Issue 68)
Sigma 100 S	SS103	2017 2nd Edition (Issue 77)
Sigma 100 S	SS106	2017 2nd Edition (Issue 77)
Sigma 200 D	SD203	2017 2nd Edition (Issue 77)

Implantable Pulse Generators (IPGs) continued

Product Name	Model	Final Issue
Sigma 200 DR	SDR203	2017 2nd Edition (Issue 77)
Sigma 200 S	SS203	2017 2nd Edition (Issue 77)
Sigma 200 SR	SSR203	2017 2nd Edition (Issue 77)
Sigma 300 D	SD303	2023 2nd Edition (Issue 89)
Sigma 300 DR	SDR303	2023 2nd Edition (Issue 89)
Sigma 300 DR	SDR306	2019 2nd Edition (Issue 81)
Sigma 300 S	SS303	2023 2nd Edition (Issue 89)
Sigma 300 SR	SSR303	2023 2nd Edition (Issue 89)
Sigma 300 SR	SSR306	2019 2nd Edition (Issue 81)
Sigma 300 VDD	SVDD303	2019 2nd Edition (Issue 81)
Thera-i DR	7960i	2012 1st Edition (Issue 66)
Thera-i DR	7961i	2012 1st Edition (Issue 66)
Thera-i DR	7962i	2012 1st Edition (Issue 66)
Thera-i SR	8960i	2012 1st Edition (Issue 66)
Thera-i SR	8961i	2012 1st Edition (Issue 66)
Thera-i SR	8962i	2012 1st Edition (Issue 66)
Thera-i VDD	8968i	2012 1st Edition (Issue 66)

LEADS

Pacing Leads

Product Name	Model	Final Issue
CapSure Sense	4073	2023 2nd Edition (Issue 89)
CapSure SP	4023	2012 2nd Edition (Issue 67)
CapSure SP	4024	2016 1st Edition (Issue 74)
CapSure SP	4523	2012 2nd Edition (Issue 67)
CapSure SP	4524	2016 1st Edition (Issue 74)
CapSure SP	5023	2012 2nd Edition (Issue 67)
CapSure SP	5023M	2012 2nd Edition (Issue 67)
CapSure SP	5024	2013 1st Edition (Issue 68)
CapSure SP	5024M	2013 1st Edition (Issue 68)
CapSure SP	5524	2013 1st Edition (Issue 68)
CapSure SP	5524M	2013 1st Edition (Issue 68)
CapSure Z	4033	2012 2nd Edition (Issue 67)
CapSure Z	4533	2012 2nd Edition (Issue 67)
CapSure Z	5033	2016 1st Edition (Issue 74)
CapSure Z	5034	2016 1st Edition (Issue 74)
CapSure Z	5534	2016 1st Edition (Issue 74)
CapSureFix	4067	2012 2nd Edition (Issue 67)
CapSureFix	4068	2016 1st Edition (Issue 74)

Pacing Leads continued

Product Name	Model	Final Issue
CapSureFix	4568	2017 2nd Edition (Issue 77)
CapSureFix	5068	2017 1st Edition (Issue 76)
CapSureFix	5568	2016 1st Edition (Issue 74)
CapSureFix	6940	2018 1st Edition (Issue 78)
Screw-In	4558M	2016 1st Edition (Issue 74)
SureFix	5072	2018 1st Edition (Issue 78)

Defibrillation Leads

Product Name	Model	Final Issue
Epicardial Patch	6921	2013 1st Edition (Issue 68)
Sprint	6932	2016 1st Edition (Issue 74)
Sprint	6942	2017 1st Edition (Issue 76)
Sprint	6943	2017 2nd Edition (Issue 77)
Sprint	6945	2017 2nd Edition (Issue 77)
Sub-Q	6999	2012 1st Edition (Issue 66)
Sub-Q Patch	6939	2012 1st Edition (Issue 66)
SVC/CS	6963	2013 1st Edition (Issue 68)
Transvene	6936	2013 1st Edition (Issue 68)
Transvene	6966	2013 1st Edition (Issue 68)
Transvene SVC	6937	2016 1st Edition (Issue 74)
Transvene SVC-CS	6933	2016 1st Edition (Issue 74)

Left Heart Pacing Leads

Product Name	Model	Final Issue
Attain CS	2188	2012 2nd Edition (Issue 67)

Epicardial/Myocardial Pacing Leads

Product Name	Model	Final Issue
Spectraflex	4951	2013 1st Edition (Issue 68)
Spectraflex	4951M	2013 1st Edition (Issue 68)

VDD Single Pass Pacing Leads

Product Name	Model	Final Issue
CapSure VDD	5032	2016 1st Edition (Issue 74)

Mailer Kits Available for Returning Product

Medtronic urges all physicians to return explanted products and to notify Medtronic when a product is no longer in use, regardless of reason for explant or removal from use. The procedures for returning products vary by geographic location.

Mailer kits with prepaid US postage are available for use within the United States to send CRT, ICD, IPG, and leads to Medtronic's CRM Returned Product Analysis Lab. These mailers are sized to accommodate the devices and leads from a single patient or clinical event and are designed to meet US postal regulations for mailing biohazard materials.

If the product being returned is located outside the United States, please contact your local Medtronic representative for instructions.

Medtronic also requests the return of devices from non-clinical sources, such as funeral homes, and will assume responsibility for storage and disposal of the product once received.

Mailer kits can be obtained by contacting the Returned Product Lab.

CRM Returned Product Analysis Laboratory
Phone: 1 (800) 328-2518, ext. 44800
Email: crdm.returnedproduct@medtronic.com



For questions related to returning explanted product from outside the United States, please contact your local Medtronic Representative.

Medtronic
710 Medtronic Parkway
Minneapolis, MN 55432-5604
USA
Tel: (763) 514-4000
Fax: (763) 514-4879

Toll-free: 1 (800) 328-2518
(24-hour technical support for physicians
and medical professionals)

medtronic.com

Medtronic